MILITARY SPECIFICATION

ADENOTOME, LaFORCE

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers an adenotome of the LaForce design for use in surgical procedures.

1.2 Sizes. The adenotomes shall be of the following sizes, as specified (see 6.2).

Size 2 Medium
Size 3 Large

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposals form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

GG-I-526	Instruments, Dental and Surgical; General Specifications.
MMM-A-250	Adhesive, Water Resistant, (for sealing fiberboard boxes).
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-T-76	Tape, Pressure-Sensitive Adhesive Paper (for carton sealing).

Military

MIL-P-116	Preservation, Methods of.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:12Z
Check the source to verify that this is the current version before use.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage

DRAWING

Defense Personnel Support Center (DPSC)

20310	Adenotome, LaForce

(Copies of specifications, standards and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 General requirements. The two sizes of adenotomes, of the LaForce design, shall be in accordance with the applicable requirements of GG-I-526, and as specified herein.

3.2 Material. The adenotome shall be fabricated from material of the class and type as listed in Table I.

3.3 Hardness. After proper stress relieving, the Rockwell hardness shall be as listed in Table I.

Table I.

Component	Class	Type	Hardness
Complete adenotome	3 or 4	410 or	C40-C45
		cold rolled 420	C40-C45
		or	B75-B85
		cold rolled	
		302, 303, or 304	B75-B85
As an alternate the following components shall be:			
Finger rings	8	Alloy No. 260 or 270	- - - -
Thumb screw and pins	4	416	B75-B85

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:12Z
Check the source to verify that this is the current version before use.

3.4 Style, design, and dimensions. Shall be an adenotome of the LaForce design. The style, design, and dimensions of the adenotome shall conform to Defense Personnel Support Center (DPSC) Drawing No. 20310.

3.4.1 Fixed member. Fixed member shall terminate anteriorly in the convex open end in which the blade slides. At the posterior end, the fixed member shall terminate in two finger rings, one above and the other below.

3.4.2 Movable member. Movable member shall terminate anteriorly in the flexible blade which slides in the grooves of the fixed member. The movable member shall rest upon and in close apposition to the fixed member. Posteriorly it shall terminate in the freely rotating thumb ring.

3.4.3 Blade. Two blades, of medium or large size in accordance with appropriate stock number, shall be furnished with each instrument. Blades shall be readily attached to or detached from the movable member by means of a plug placed under its posterior or rounded end.

3.5 Finish. Shall be mirror finish, or mirror and satin.

3.6 Identification marking. In addition to GG-I-526 requirements, adenotome of foreign manufacture shall also be marked when applicable.

3.7 Resistance to salt spray. Shall apply for class 8 metals.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:12Z
Check the source to verify that this is the current version before use.

4.1.2 Inspection. Shall be in accordance with GG-I-526.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the metal, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the FSN, the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels shall be as indicated in Table II. The unit of product for examination purposes shall be one adenotome.

Table II. Sampling for examination

Type of defect	Inspection level	AQL (percent defective)
Major	II	1.0
Minor	II	2.5

4.2.2 Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as indicated in Table III. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Table III. Sampling for tests

Component (unit of product)	Characteristic	Requirement	Test procedure	Inspection level	AQL (percent defective)
Adenotome	Composition analysis	3.2 and GG-I-526	GG-I-526	**	*
	Hardness	3.3 and GG-I-526	GG-I-526	S-2	*
	Dimensions	3.4	- - - -	S-2	*
	Copper sulfate test	GG-I-526	GG-I-526	S-2	*
	Boil test	GG-I-526	GG-I-526	S-2	*
	Salt spray test	GG-I-526	GG-I-526	S-2	*

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:12Z
Check the source to verify that this is the current version before use.

* Acceptance number shall be zero for all sample sizes.

** Sufficient amount from each lot of each metal used in the construction of the adenotome shall be selected for composition analysis.

4.3 Examination. The adenotome shall be examined to determine compliance with specification requirements. Examination shall be in accordance with MIL-STD-105 and the following classification of defects. Examination shall include, but shall not be limited to the defects listed in Table IV.

Table IV. Classification of defects

Categories	Defects
Critical	None defined.
Major	
101	Style and design not as specified.
102	Instrument not free from cracks, pits, dents, burrs, grind marks or overlays.
103	Thumb ring not freely rotating.
104	Cutting edge of blade not sharp.
105	Blade not easily replaceable.
Minor	
201	Finish not as specified.
202	Identification marking not present, correct, complete or legible.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One adenotome of the size specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in Table V shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:12Z
Check the source to verify that this is the current version before use.

Table V. Packaging Quantities

Size	Unit package	Intermediate package	Exterior container
2 or 3	1 unit	6 units	96 units

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Cleaning and drying. Each unit shall be thoroughly cleaned and dried in accordance with the applicable method as outlined in MIL-P-116.

5.1.3.2 Preservative. Not required.

5.1.3.3 Intimate wrap. Each unit shall be wrapped in neutral tissue or other suitable anticorrosive material.

5.1.3.4 Unit package. Each unit shall be packaged in accordance with MIL-P-116, method IA-15.

5.1.3.5 Intermediate package. Intermediate package shall be a box of appropriate size and design constructed in accordance with PPP-B-636, type CF, class weather-resistant. Boxes shall be closed by taping the top and bottom flaps with 2 inch wide tape conforming to PPP-T-76. Closing of bottom flaps with adhesive conforming to MMM-A-250 or metal stitches is permitted.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:12Z
Check the source to verify that this is the current version before use.

5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant, grade V3. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure of wood boxes shall be as specified in the appendix of the applicable box specification. Closure of fiberboard boxes shall conform to method II as specified in PPP-B-636.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Intended use. The adenotome is intended for use during surgical procedures such as excision of the adenoids and retaining same in basket.

6.2 Ordering data. Purchasers should exercise any desired options offered herein, and the procurement documents shoudl specify the following:

(a) Number, title, and date of this specification.
(b) Size (see 1.2).
(c) Applicable Federal Stock Number (FSN).
(d) Selection of applicable levels of packaging and packing required (see 5.1 and 5.2).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:12Z
Check the source to verify that this is the current version before use.

6.3 This specification is intended to cover the following items appearing in the Federal Supply Catalog:

Federal Stock Number	Item Identification
6515-300-1960	Adenotome, LaForce, Size 2
6515-300-1980	Adenotome, LaForce, Size 3

It does not include all types or sizes as indicated by the title nor all types which are commercially available.

6.4 Supersession data. This specification includes requirements of Purchase Description No. 10, dated 12 March 1968 for FSN 6515-300-1960 Adenotome, LaForce, Size 2 and FSN 6515-300-1980 Adenotome, LaForce, Size 3.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1544

Review activities:
Army - MD
Navy - MS
Air Force - 03

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:12Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document (s) or serve to amend contractual requirements.

SPECIFICATION
MIL-A-36725 FSN:

ORGANIZATION

CITY AND STATE | CONTRACT NUMBER

MATERIAL PROCURED UNDER A
☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO *(If "yes", in what way?)*

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)* | DATE.

DD FORM 1 JAN 66 1426 REPLACES EDITION OF 1 OCT 64, WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:12Z
Check the source to verify that this is the current version before use.