MIL-C-36726
11 December 1970

MILITARY SPECIFICATION

CLAMPS, PYLORUS, PAYR

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers pylorus clamps, of the Payr design, for use during surgical procedures of the human pylorus.

1.2 Classification. Clamps shall be of the following sizes as indicated below (see (6.2):

| Size | Overall Length (inches) |
|---|---|
| I | 8 |
| II | 11 |

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposals, form a part of the specification to the extent specified herein:

SPECIFICATIONS

Federal

| | |
|---|---|
| GG-I-526 | Instruments, Dental and Surgical; General Specifications. |
| MMM-A-250 | Adhesive, Water Resistant, (for Sealing Fiberboard Boxes). |
| PPP-B-585 | Boxes, Wood, Wirebound. |
| PPP-B-601 | Boxes, Wood, Cleated-Plywood. |
| PPP-B-621 | Boxes, Wood, Nailed and Lock-Corner. |
| PPP-B-636 | Box, Fiberboard. |
| PPP-T-76 | Tape, Pressure-Sensitive Adhesive Paper, (for Carton Sealing). |

Military

| | |
|---|---|
| MIL-P-116 | Preservation, Methods of. |

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:13Z
Check the source to verify that this is the current version before use.

STANDARDS

Military

| | |
|---|---|
| MIL-STD-105 | Sampling Procedures and Tables for Inspection by Attributes. |
| MIL-STD-129 | Marking for Shipment and Storage. |

DRAWING

Defense Personnel Support Center (DPSC)

| | |
|---|---|
| 20515 | Clamp, Pylorus, Payr |

(Copies of specifications, standards, and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 General requirements. The clamps shall be in accordance with the applicable requirements of GG-I-526, and as specified herein.

3.2 Material. The clamps shall be fabricated from class 4, type 410, 416 or 420 material. As an alternate, the screws or pins shall be fabricated from class 9 material having the following composition:

| | |
|---|---|
| Carbon: | 0.14 to 0.17 percent |
| Chromium: | 15.0 to 18.0 percent |
| Sulfur or Selenium: | 0.07 percent maximum |

The corrosion-resistance requirements for the pin shall conform to those specified for type 410 material.

3.3 Hardness. After proper stress relieving Rockwell hardness shall be C40 to C45 for class 4 material, and C15 to C25 for class 9 material.

3.4 Style, design and dimensions. Clamps specified herein, shall be of the Payr design, suitable for holding sections of the human pylorus for suturing. Handles of the clamps shall be rigidly constructed. Jaws of the clamps shall be uniformly serrated, and when the instrument is placed into the closed position, the serrations shall mesh properly for the purpose intended.

Size I and Size II Clamps. Shall be as indicated on DPSC Drawing No. 20515 for appropriate size, Federal Stock Number and Item identification.



Source: https://assist.dla.mil -- Downloaded: 202610-21T09:13Z
Check the source to verify that this is the currentversion before use.

3.5 Finish. Finish shall be mirror, or mirror and satin.

3.5.1 Passivation. The entire instrument shall be properly passivated.

3.6 Identification marking. In addition to GG-I-526 requirements, instruments of foreign manufacture (other than U.S.) shall also be marked "stainless".

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Shall be in accordance with GG-I-526.

4.1.1 Inspection. Shall be in accordance with GG-I-526.

4.1.2 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the corrosion-resistant steel, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the FSN, the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels shall be as indicated in Table I. The unit of product for examination purposes shall be one clamp.

Table I. Sampling Criteria for Examination

| Type of defect | Inspection level | AQL (percent defective) |
|---|---|---|
| Major | II | 1.0 |
| Minor | II | 2.5 |



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:13Z
Check the source to verify that this is the current version before use.

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, and as indicated in Table II. Unit of product for test purposes shall be one clamp.

Table II. Sampling Criteria for Tests

| Component | Characteristic | Requirement | Test Procedure | Insp. level | AQL |
|---|---|---|---|---|---|
| Clamp | Composition analysis | 3.2 and GG-I-526 | GG-I-526 | * | 1.0 |
| | Hardness | 3.3 and GG-I-526 | GG-I-526 | S-2 | 1.0 |
| | Dimensions | 3.4 | - - - - | S-2 | 2.5 |
| | Copper sulfate test | GG-I-526 | GG-I-526 | S-2 | 1.0 |
| | Boil test | GG-I-526 | GG-I-526 | S-2 | 1.0 |

* Four ounces of each lot of each metal used in the construction of the clamps shall be selected for chemical analysis.

4.2.3 Sample unit usage. Where feasible, the same sample unit shall be used for the determination of two or more test requirements.

4.3 Examination. The clamps shall be examined to determine compliance with all the requirements contained in this specification. Examination shall be conducted as specified in Tables I and III. Inspection, however, shall not be restricted to defects listed in Table III. Any clamp in the sample containing one or more defects shall be rejected and if the number of defective clamps in any sample exceeds the acceptance number for that sample, the lot represented by the sample shall be rejected.

Table III. Classification of Defects

| Categories | Defects |
|---|---|
| Critical | None defined. |
| Major | |
| 101 | Clamp not of the size specified for appropriate Federal Stock Number and Item Identification. |
| 102 | Clamp not free from sharp edges, rough areas, burrs, cracks, tool marks or pits. |
| 103 | Finish not as specified. |
| 104 | Jaw serrations do not mesh properly in closed position. |
| 105 | Jaw serrations not well cut, not well defined, or not uniform. |
| 106 | Handles not capable of easy closure and opening. |
| 107 | Pivoting points not free from excessive play. |
| Minor | |
| 201 | Identification marking not present, correct, complete or legible. |



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:13Z
Check the source to verify that this is the current version before use.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One clamp, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in Table IV shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table IV. Packaging Quantities

| Unit package | Intermediate package | Exterior container |
|---|---|---|
| 1 unit | 6 units | 72 units |

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Cleaning and drying. Each unit shall be thoroughly cleaned and dried in accordance with the applicable method as outlined in MIL-P-116.

5.1.3.2 Preservative. Not required.

5.1.3.3 Intimate wrap. Each unit shall be wrapped in neutral tissue or other suitable anticorrosive material.

5.1.3.4 Unit package. Each unit shall be packaged in accordance with MIL-P-11 method IA-15.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:13Z
Check the source to verify that this is the current version before use.

5.1.3.5 Intermediate package. Intermediate package shall be a box of appropriate size and design constructed in accordance with PPP-B-636, type CF, class weather-resistant. Boxes shall be closed by taping the top and bottom flaps with 2 inch wide tape conforming to PPP-T-76. Closing of bottom flaps with adhesive conforming to MMM-A-250 or metal stitches is permitted.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-b-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant, grade V3. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:13Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Intended use. The pylorus clamps described herein shall be suitable for holding sections of the human pylorus during surgical procedures.

6.2 Ordering data. Procurement documents should specify the following:

(a) Number, title and date of this specification.
(b) Size (see 1.2).
(c) Applicable Federal Stock Number (FSN).
(d) Selection of applicable level of packaging and packing.

6.3 This specification covers the following items listed in the Federal Supply Catalog:

| FSN | Item Identification |
|---|---|
| 6515-320-3800 | CLAMP, PYLORUS, Payr, 8 inch |
| 6515-320-3840 | CLAMP, PYLORUS, Payr, 11 inch |

It does not include all types and sizes as indicated by the title nor all types commercially available.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. - 6515-1545

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:13Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:13Z
Check the source to verify that this is the current version before use.

# SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corners, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document (s) or serve to amend contractual requirements .

SPECIFICATION

MIL-C-36726 FSN:

ORGANIZATION

CITY AND STATE | CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO *(If "yes", in what way?)*

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)* | DATE

DD FORM 1 JAN 66 1426 REPLACES EDITION OF 1 OCT 64, WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:13Z
Check the source to verify that this is the current version before use.

## CAUTION

## NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

---

*Fold*

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

---

*Fold*

DPSC FORM 2418
MAR 68

EDITION AUG 66, WILL BE USED UNTIL EXHAUSTED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:13Z
Check the source to verify that this is the current version before use.