MILITARY SPECIFICATION

RETRACTOR, EYELID, NONMAGNETIC, DESMARRES, 5-1/2 INCH

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a nonmagnetic eyelid retractor, of the Desmarres design, suitable for removing foreign bodies from the eye.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposals form a part of the specification to the extent specified herein:

SPECIFICATIONS

Federal

GG-I-526	Instrument, Dental and Surgical; General Specifications.
MMM-A-250	Adhesive, Water Resistant (for Sealing Fiberboard Boxes).
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-T-76	Tape, Pressure-Sensitive Adhesive Paper, (for Carton Sealing).

Military

MIL-P-116	Preservation, Methods of.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:08Z
Check the source to verify that this is the current version before use.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

DRAWING

Defense Personnel Support Center (DPSC)

20552	Retractor, Eyelid, *Nonmagnetic, Desmarres*, 5-1/2 inch

(Copies of specifications, standards and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 General requirements. The eyelid retractor shall be in accordance with the applicable requirements of GG-I-526, and as specified herein.

3.2 Material. Shall be fabricated from Class 3, Type 316 material. As an alternate, handle only, shall be fabricated from Class 8, Type 260, 270 or 360 material.

3.3 Hardness. Rockwell hardness of Class 3 material shall be B75 to B85.

3.4 Style, design and dimensions. Shall be a nonmagnetic, eyelid retractor of the Desmarres design. Shall be suitable for use with FSN 6515-299-8008 MAGNET, EYE, ELECTRIC in removing foreign bodies from the human eye. Retractor shall be in accordance with Defense Personnel Support Center (DPSC) Drawing 20552.

3.5 Finish. Finish shall be mirror finish, except for handle which may be satin finish.

3.5.1 Passivation. All components of the retractor shall be properly passivated prior to final assembly.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:08Z
Check the source to verify that this is the current version before use.

3.6 Nonmagnetic characteristics. The retractor shall show no magnetic response when tested as specified in 4.4.1.

3.7 Identification marking. In addition to GG-I-526 requirements, foreign instruments shall also be marked "stainless".

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Shall be in accordance with GG-I-526.

4.1.1 Inspection. Shall be in accordance with GG-I-526.

4.1.2 Certificates of quality. Certificates of quality supplied by the manufacturer of the corrosion resisting steel and brass may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the FSN, the item identification, the name of the material, the lot number, the lot size, the sample size, the date of testing, the test method, the individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels shall be as indicated in Table I. The unit of product for examination purposes shall be one retractor.

Table I. Sampling Criteria for Examination

Type of defect	Inspection level	AQL (percent defective)
Major	II	1.0
Minor	II	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, and as indicated in Table II. Unit of product for test purposes shall be one retractor.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:08Z
Check the source to verify that this is the current version before use.

Table II. Sampling Criteria for Tests

Component	Characteristic	Req.	Test procedure	Insp. level	AQL
Retractor	Composition analysis	3.2 and GG-I-526	GG-I-526	**	*
	Hardness	3.3 and GG-I-526	GG-I-526	S-2	*
	Dimensions	3.4	- - - -	S-1	2.5
	Boil test	GG-I-526	GG-I-526	S-2	*
	Nonmagnetic test	3.6	4.4.1	S-2	*
Handle (when Class 8 material)	Thickness of coating	GG-I-526	GG-I-526	S-2	*
	Salt-spray test	GG-I-526	GG-I-526	S-2	*

* Acceptance number shall be zero for all sample sizes.

** Four ounces of each lot of each metal used in construction of the retractors shall be selected for chemical analysis.

4.2.3 Sample unit usage. Where feasible, the same sample unit shall be used for the determination of two or more test requirements.

4.3 Examination. The retractors shall be examined to determine compliance with all the requirements contained in this specification. Examination shall be conducted as specified in Tables I and III. Inspection, however, shall not be restricted to defects listed in Table III. Any retractors in the sample containing one or more defects shall be rejected and if the number of defective retractors in any sample exceeds the acceptance number for the sample, the lot represented by the sample shall be rejected.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:08Z
Check the source to verify that this is the current version before use.

Table III. Classification of Defects

Categories	Defects
Critical	None defined.
Major	
101	Retractor not free from sharp edges, rough areas, burrs, cracks, tool marks and pits.
102	Finish not as specified.
103	Junction of blade shank to handle not secure.
104	Plating of handles (when fabricated from brass) not uniform.
Minor	
201	Identification marking not present, correct, complete or legible.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Nonmagnetic test. The eyelid retractors shall be placed at rest, laying flat, upon a smooth nonferrous surface. An Alnico horseshoe-shaped magnet, weighing between 0.131 and 0.145 pound shall then be used to slowly approach, and make contact with, the surfaces of the retractor. The magnet shall be fabricated from Alnico No. 5 alloy, and shall be proven capable of supporting 43 cubic inches of steel (3-1/2 inches x 3-1/2 inches x 3-1/2 inches) weighing 12 $\pm$ 1/2 pounds.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A or C as specified (see 6.2).

5.1.1 Unit of issue. One retractor, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in Table IV shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:08Z
Check the source to verify that this is the current version before use.

Table IV. Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	6 units	288 units

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Cleaning and drying. Each unit shall be thoroughly cleaned and dried in accordance with the applicable method as outlined in MIL-P-116.

5.1.3.2 Preservative. Not required.

5.1.3.3 Intimate wrap. Each unit shall be wrapped in neutral tissue or other suitable anticorrosive material.

5.1.3.4 Unit package. Each unit shall be packaged in accordance with MIL-P-116, method IA-15.

5.1.3.5 Intermediate package. Intermediate package shall be a box of appropriate size and design constructed in accordance with PPP-B-636, type CF, class weather-resistant. Boxes shall be closed by taping the top and bottom flaps with 2 inch wide tape conforming to PPP-T-76. Closing of bottom flaps with adhesive conforming to MMM-A-250 or metal stitches is permitted.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:08Z
Check the source to verify that this is the current version before use.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant, grade V3. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Intended use. The eyelid retractors are to be used in removing foreign bodies from the eye.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:08Z
Check the source to verify that this is the current version before use.

6.2 Ordering data. Procurement documents should specify the following:

(a) Title, number and date of this specification.
(b) Applicable levels of packaging and packing (see 5.1 and 5.2).
(c) Federal stock number.

6.3 This specification is intended to cover Federal Stock Number 6515-360-7900 RETRACTOR, EYELID, Nonmagnetic, Desmarres, 5-1/2 inch, appearing in the Federal Supply Catalog.

6.4 Suspersession data. This specification includes requirements of Purchase Description No. 5 dated 12 May 1967.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1492

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:08Z
Check the source to verify that this is the current version before use.