MILITARY SPECIFICATION

RONGEUR, CURVED, BANE, 7 INCH

This amendment forms a part of Military Specification MIL-R-36694, dated 24 June 1970 and is approved for use by all Departments and Agencies of the Department of Defense.

Page 2:

3.2 Line 2, after "percent" add "minimum" and delete remainder of sentence.

3.4 End of sentence, delete period and add: "except, in end view, delete '.015 max. clearance' and substitute: '.020 max. clearance'; overall length, delete '7-1/8 ± 1/8' and substitute: '7-1/8 + 1/8, - 1/4'; tongue of spring, delete '7/64 ± .005' and substitute: '.088 ± .026'; thickness of spring, delete '.051 ± .003' and substitute: '.045 to .054'."

Pages 2 and 3:

4.1, 4.1.1 and 4.1.2, Delete entirely and substitute:

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:09Z
Check the source to verify that this is the current version before use.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the stainless steel, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

Page 5:

5.1.3.2 Preservative. Delete and substitute:

"5.1.3.2 Preservative. Preservative shall not be utilized."

5.1.3.4 Unit package. Line 2, delete "IA - 8" and substitute: "IC - 1".

Page 6:

5.2 Packing. Delete and substitute:

"5.2 Packing.

"5.2.1 Level A.

"5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant. Closure and strapping shall be as specified in the appendix of the applicable box specification. Special requirements as specified in PPP-B-636 shall apply for fiberboard boxes.

"5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraph of appendix thereto. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed in accordance with 30.4 of PPP-B-636.

"NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

"5.2.2 Level B.

"5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification. Closure of fiberboard boxes shall conform to method II as specified in PPP-B-636. In addition, fiberboard boxes shall conform to the special requirements as specified in PPP-B-636.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:09Z
Check the source to verify that this is the current version before use.

"5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order."

Custodians:
Army- Md
Navy- Ms
Air Force- 03

Review activities:
Army- Md
Navy- Ms
Air Force- 03

Preparing activity:
Defense Supply Agency- DH

Project No. 6515- 1852

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:09Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:09Z
Check the source to verify that this is the current version before use.