MIL-S-36703
Amendment-2
25 February 1974
Superseding
Amendment-1
3 August 1973

MILITARY SPECIFICATION

SAW, BONE CUTTING, AUTOPSY, STRYKER, 110 Volt, AC-DC

This Amendment forms a part of Military Specification MIL-S-36703, dated 21 April 1971, and is approved for use by all Departments and Agencies of the Department of Defense.

PAGE 4:

3.3 Style, design and dimensions. Delete period at end of paragraph and substitute ", except that the overall width between brush holders shall be 2-7/8 inches reference."

3.3:1 Motor and switch. In fifth line, delete "machined metal plugs", and substitute: "durable, corrosion-resistant plugs".

3.3.7 Blade. Delete "FSN 6515-299-8718", and substitute: "NSN 6515-00-299-8718".

PAGE 6:

3.3.10 Autopsy bone saw performance. At end of paragraph add: "The 100 hours minimum continuous operation test may be waived upon request by the contractor, provided that the contractor has previously furnished items in accordance with requirements of this specification, and that there has been no major design change of motor and drive mechanism."

PAGE 7:

3.4 Service data. Delete "FSN 6515-299-8717" and substitute "NSN 6515-00-299-8717".

PAGE 13:

5.1.3.1 Unit package. Line 4, delete "IA-15" and substitute "IC-2".

Following paragraph 5.2.1.1 insert the following:

"NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage or redistribution.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:10Z
Check the source to verify that this is the current version before use.

"5.2.1.2 Case liner. Each Level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Fiberboard box shall be waterproof in accordance with paragraph 30.4 of PPP-B-636."

5.2.2.1 Exterior container. At end of paragraph add: "Closure of fiberboard boxes shall conform to method II as specified in PPP-B-636. In addition, fiberboard boxes shall conform to the special requirements as specified in Table I of PPP-B-636."

PAGE 14:

6.3 Delete "Federal Stock No. 6515-299-8717", and substitute "National Stock No. 6515-00-299-8717".

6.5 Delete entirely.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MD
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1803

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:10Z
Check the source to verify that this is the current version before use.