MILITARY SPECIFICATION

SAW, BONE CUTTING, AUTOPSY, STRYKER, 110 VOLT, AC-DC

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers an electric motor driven saw, with high speed oscillating drive, Stryker design, suitable for cutting human bone in autopsy techniques.

2. APPLICABLE DOCUMENTS

2.1 The following documents, of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

GG-I-526	Instruments, Dental and Surgical; General Specifications.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.

Military

MIL-P-116	Preservations, Methods of.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:10Z
Check the source to verify that this is the current version before use.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

DRAWING

Defense Personnel Support Center (DPSC)

24245	SAW, Bone Cutting, Autopsy, Stryker, 110 Volt, A.C.-D.C.

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply.

UNDERWRITERS' LABORATORIES, INC. (UL)

Underwriters' Laboratories Standards

Underwriters' Laboratories Standards for Safety

UL 45 Electric Tools

(Application for copies should be addressed to the Underwriters' Laboratories Inc., Walt Whitman Road Melville, L. I., N. Y. 11746; 207 East Ohio Street, Chicago, Illinois 60611; and 1655 Scott Boulevard, Santa Clara, California 95050.)

AMERICAN NATIONAL STANDARDS INSTITUTE (A.N.S.I.)

A.N.S.I. C 73.11 Outlet Receptacles, Attachment Plug Caps and Appliance Plugs.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:10Z
Check the source to verify that this is the current version before use.

(Application for copies should be addressed to the American National Standards Institute, 1430 Broadway, N. Y., N. Y. 10018.)

3. REQUIREMENTS

3.1 Fire and Casualty hazards.

3.1.1 The contractor shall submit to the contracting agency proof, prior to delivery, that the equipment supplied under this specification conforms to the applicable requirements of Underwriters' Laboratories, Inc. Standard UL 45. Such proof shall consist of a laboratory report or listing by the Underwriters' Laboratories, Inc. The label or listing of the Underwriters' Laboratories shall be acceptable as evidence that the equipment conforms to the above. A copy of each laboratory report shall be furnished to the contracting officer.

3.1.2 In lieu of the label or listing by the Underwriters' Laboratories, Inc., the contractor may submit a laboratory report by a nationally recognized testing agency adequately equipped and competent to perform such services, as shall be evidenced by the attachment of its seal, stamp or label to the article, or the issuance of a certificate. The contractor shall submit the laboratory report which shall include test results and methods used to determine compliance, to the contracting officer for approval. The Government reserves the right to require a facilities report showing the qualifications of the testing laboratory and to conduct a facilities survey to determine its acceptability.

3.1.3 Compliance with the above requirements regarding fire and casualty hazards, does not absolve the contractor from complete compliance with the requirements of this specification in order to secure the acceptance of his commodity.

3.2 Material. Material shall be of the highest quality throughout and suitable for the purpose intended. The arbor proper, of the saw, including arbor nut and collar, shall be fabricated from class 3, type 302, 303 or 304 material in accordance with GG-I-526 and shall have a Rockwell hardness of B 75 minimum. The arbor pin shall be fabricated from class 4, type 416 material in accordance with GG-I-526 and shall have a Rockwell hardness of C 38 to C 42 after proper stress relieving. The arbor nut wrench shall be fabricated from corrosion-resistant steel. All material fabricated from corrosion-resistant steel shall be passivated.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:10Z
Check the source to verify that this is the current version before use.

3.3 Style, design and dimensions. Shall be an electric motor driven autopsy saw, of the Stryker design, suitable for cutting human bone in autopsy procedures without excessively damaging soft tissue. Saw shall be of the nonrotating, high frequency, oscillating design, complete with an oscillating head, two blades, one blade arbor assembly, one Allen wrench, one arbor nut wrench, and spare parts as specified herein. Style, design and dimensions shall be in accordance with DPSC Drawing 24245.

3.3.1 Motor and switch. Motor shall operate on either 110 volt direct current, or 110 volt, single phase, 60 cycles per second, alternating current. Motor shall be fan cooled, permanently lubricated, universal type, rated at 0.1 horsepower at 10,000 revolutions per minute (RPM) continuous. Brush holders shall be protected by machined metal plugs threaded into and flush with saw housing. Saw shall have "off" and "on" switch (properly identified) mounted on power handle. Simple and effective means shall be provided for removing and replacing motor brushes. Motor and switch shall comply with requirements of Underwriters' Laboratories, Inc. Standard UL 45 for continuous rated universal motors for electric tools.

3.3.2 Power cable and cord. The saw shall be equipped with a 3-wire, heavy duty, electrical power cord, 15 feet $\pm$ 1/2 foot long, securely attached to the motor end of the power handle, and an attached 3-prong plug to the end of the cable. Plug shall be in accordance with A.N.S.I. Standard C 73.11. The power cable shall be provided with a spring type component, or other suitable device, that shall cover at least 2 inches of the cable, at point of attachment to the power handle to assure adequate strain relief of the cable.

3.3.3 Blade arbor shaft assembly. The centerline of the blade arbor shaft assembly shall be offset from the centerline of the motor shaft, as shown on drawing, to provide blade in a safe and convenient position for saw operation. The arbor shaft assembly shall be supported in sleeve bearings on each side of open link mechanism that transmits rotating motion of motor into oscillatory motion of the blade. Blade arbor assembly shall include a slotted collar, nut and pin, designed to provide a proper and secure fit for the blades specified in 3.3.7.

3.3.4 Center of balance. With the power cable of the saw supported or detached from the saw, the center of balance of the saw assembly shall be 6-1/2 $\pm$ 1/2 inches from the front side of the arbor shaft assembly (measured parallel to the longitudinal axis of the power handle) as shown on drawing.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:10Z
Check the source to verify that this is the current version before use.

3.3.5 Drive mechanism . The rotary motion of the motor shall be transmitted into an oscillatory motion effecting strokes of 1/8 ± 1/64 inch excursion at one inch radius, at the blade end, by means of a cam and open link mechanism. The cam shall be equipped with a roller bearing. The open link mechanism shall be mechanically fastened to the arbor blade assembly to suitably transmit required oscillatory rate specified herein. Gear transmission design shall not be acceptable.

3.3.6 Power handle (housing). The entire working mechanism shall be enclosed by a compact, rigid, and thoroughly protected metallic corrosion-resistant housing. The housing, except for the ventilating openings, shall totally enclose all stationary and rotating electrical parts. Four inlet air ports shall be provided at the brush end of the motor housing. Air shall be exhausted through eight air ports suitably sized at the front end of the motor housing. Finish of the housing shall be bright, chromium plated. Contours of mating housing parts shall match, with no sharp corners or edges. All external fastening devices shall be recessed into the housing. Housing components shall be accurately fitted with adequate oil seals and gaskets to prevent oil leakage. The distal end of the power handle shall be fitted with a shaft extension (chuck) to permit secure fit of the arbor. The chuck shall be provided with an Allen head setscrew for securing the blade arbor within the chuck.

3.3.7 Blade. Each autopsy saw shall be furnished with two large-sectional blades in accordance with Defense Medical Purchase Description for FSN 6515-299-8718 BLADE, BONE SAW, ELECTRIC, AUTOPSY, Stryker.

3.3.8 Wrenches. The autopsy saw shall be supplied with one Allen wrench securely attached to the power cable specified in 3.3.2, for use in changing blades in the arbor of the saw. An arbor nut wrench shall be furnished to facilitate change of blades on the blade arbor. The arbor nut wrench shall be in accordance with drawing for design and dimensions.

3.3.9 Fastening devices. All screws, pins, bolts and similar fastening devices shall be installed with means of preventing loss of tightness.

3.3.10 Autopsy bone saw performance. With input voltage set at 110 Volts, the unit shall be capable of meeting the following performance characteristics:

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:10Z
Check the source to verify that this is the current version before use.

Characteristic	Requirement
No load cutting strokes (oscillations) per minute (Free running)	17,000 minimum
* Full load oscillations under full load current of 1.2 amperes	13,000 minimum
No load current (Free running)	0.7 ampere maximum
* Temperature rise of motor housing during 5 minute operating period, under full load current of 1.2 amperes (when operated from a cold start at 75° to 76° Fahrenheit)	40 degrees Fahrenheit maximum over ambient temperature
* Continuous operation without breakdown or failure, under full load current of 1.2 amperes without loss to cutting ability	100 hours minimum

* Full load current shall be defined as occurring when the output shaft is loaded with a mass of 7.0 ± 1.0 ounces (reference steel mass dimensions of 2 inches outer diameter and a height of 1/2 inch). The full load current of 1.2 amperes shall be read on a suitable ammeter at the loaded condition.

3.3.11 Carrying case. The autopsy saw shall be provided with a durable aluminum carrying case suitably fitted with cushioning material which shall secure all components during transit. In addition, the case shall be compartmentalized to secure the following listed blades with the attached arbors. (These blades, although not components of the saw as covered by this specification may be desired by certain user activities, see 6.4).

Blade	Quantity
Spinal column blade (Stryker Corporation No. 1105) suitable for removing the roof (laminae) of the spinal cord.	1
Mastoid saw blade (Stryker Corporation No. 1106) suitable for removing an inch square block of the mastoid to its full depth.	1
Small sectioned blade (Stryker Corporation No. 1101) suitable for use on small bones or for small animal work.	1

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:10Z
Check the source to verify that this is the current version before use.

3.3.11.1 Carrying case identification. A metal plate shall be secured to the top surface of the carrying case which shall be permanently stamped or etched with the letters "U.S.", the Federal stock number, and the contractor's name or registered trademark.

3.3.12 Spare parts. In addition to the components listed above, the autopsy saw shall be supplied with the following parts which shall be secured within the carrying case:

Spare Part	Quantity
Arbor (See 3.3.3).	1
Slotted Collar (See 3.3.3).	1
Nut (See 3.3.3)	1
Allen wrench (See 3.3.8)	2
Allen head setscrew (See 3.3.6)	4

The arbor, slotted collar, and nut shall be supplied fully assembled. The four setscrews shall be fitted to a plate secured to an inside surface of the carrying case.

3.4 Service data. The contractor shall enclose, with each unit, two copies of each of one or more booklets which contain complete instructions for installation, operation, maintenance, and a list of component parts. Wiring diagrams shall be included for electrical items or components. Each part of the unit other than common fasteners, etc., shall be identified in the service data and parts lists by the manufacturer's part number. As soon as practical after award of the contract or purchase order, the contractor shall furnish to the contracting officer, via the cognizant Government inspector, two preliminary copies of his proposed service data, for review and comment by the Defense Personnal Support Center, Directorate of Medical Materiel. The cover of each service manual shall indicate the following:

FSN 6515-299-8717
SAW, BONE CUTTING, AUTOPSY, STRYKER, 110 Volts, AC-DC
Contractor's Trademark, Model Number, Name and Address

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:10Z
Check the source to verify that this is the current version before use.

The preliminary service data shall be submitted sufficiently early to permit adequate review by the Defense Personnel Support Center (based on a maximum period of 30 days for such review) and to allow incorporation by the contractor of required revisions and corrections in the final service data without delaying delivery under the contract or purchase order. In addition, the contractor shall mail 15 copies of the approved service data to the contracting officer for distribution. Should the supplier have furnished acceptable manuals for the identical item within two years preceding the date of contract, and propose to furnish manuals identical with those previously accepted, then the requirements for review of drafts and submission of 15 additional copies of the approved service data may be waived upon request.

3.5 Identification marking. Each autopsy bone cutting saw shall be identified with a nameplate located on the power handle (motor housing) bearing the name or registered trademark of the manufacturer, manufacturer's address, type or model number, serial number, voltage, full load current and cycles per second to be used for its operation. The name plate shall further bear the following warnings: "NOT APPROVED FOR HAZARDOUS LOCATIONS", and "DO NOT AUTOCLAVE OR IMMERSE IN LIQUID." The above specified warnings shall be in clear, large and legible characters.

3.6 Workmanship. Workmanship shall be first class throughout. Autopsy saws shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:10Z
Check the source to verify that this is the current version before use.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties).

4.1.3 Certificates of quality. Certificates of quality supplied by the manufacturer of the materials used in the fabrication of the autopsy bone cutting saw, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the FSN, the item identification, the name of the component/ material, the lot number, the lot size, the sample size, the date of testing, the test method, the individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and as indicated in Table I. Unit of product for examination purposes shall be one autopsy bone cutting saw.

Table I. Sampling Criteria for Examination

Type defect	Inspection Level*	AQL (percent defective)
Major	S-4	1.0
Minor	S-4	6.5

*If acceptance number is not exceeded, lot shall be acceptable. If rejection number is equalled or exceeded, then the inspection at level II may be performed to determine the final rejection or acceptance.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:10Z
Check the source to verify that this is the current version before use.

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, and as indicated in Table II. Unit of product for test purposes shall be one autopsy bone cutting saw.

Table II. Sampling Criteria for Tests

Component	Characteristic	Requirement	Test procedure	Inspection level	AQL
Autopsy saw assembly	Motor test (compliance with UL Standard 45)	3.1.1	4.4.2	100%	0
	Dimensions	3.3	-----	S-1	1.5
	Performance test	3.3.10	4.4.1	S-3	1.0
Arbor	*Material, hardness & boil test	3.2	GG-I-526	S-2	1.0
Slotted collar	*Material, hardness & boil test	3.2	GG-I-526	S-2	1.0
Arbor nut	*Material, hardness & boil test	3.2	GG-I-526	S-2	1.0
Pin	*Material, hardness, boil test and copper sulfate test	3.2	GG-I-526	S-2	1.0
Blades (2 required)	SEE FSN 6515-899-8718 FOR ALL REQUIREMENTS				

*Four ounces of each lot of each metal used shall be selected for chemical analysis.

4.3 Examination. The autopsy bone cutting saws shall be examined to determine compliance with specification requirements. Examination shall be conducted as specified in Tables I and III. Examinations, however, shall not be restricted to defects listed in Table III. Any saw, including all components and spare parts specified, containing one or more defects shall be rejected, and if the number of defective items in any sample exceeds the acceptance number for that sample, the lot represented by the sample shall be rejected.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:10Z
Check the source to verify that this is the current version before use.

Table III. Classification of Defects

Categories	Defects
Critical	None defined.
Major	
101	Evidence of Underwriters' Laboratories (or any acceptable independent testing agency) approval not obtained as indicated in 3.1.
102	Blades not furnished as specified in 3.3.7.
103	Spare parts not furnished as specified.
104	Switch not as specified.
105	Power cable and plug not as specified.
106	Power cable not reinforced at junction at motor housing, or not type specified.
107	Housing ventilation air ports not as specified.
108	Finish of housing not smooth and uniform.
109	Saws not free from burrs, sharp edges, pits or cracks.
110	Fastening devices not recessed into housing.
111	Evidence of lubricant leakage.
112	Carrying case not furnished.
113	Allen wrench and arbor nut wrench not furnished.
114	Service data missing.
115	Identification markings and precaution warnings not as specified.
116	Gears used to transmit oscillatory motion.
117	Means to replace motor brushes clumsy or severe enough to cause cracking of motor housing.
118	Motor not permanently lubricated.
119	All stationary and rotary electrical parts not totally enclosed in motor housing.
Minor	
201	Carrying case not of material specified.
202	Finish of housing not bright.

4.4 Tests. Tests shall be conducted as necessary to determine compliance with specification requirements.

4.4.1 Performance test. Sample bone saws shall be tested for the requirements specified herein and as listed on identification plate (full load current shall be as defined in paragraph 3.3.10).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:10Z
Check the source to verify that this is the current version before use.

a. Amperes at full load and no load on 115 volts, 60 cycles per second, alternating current.

b. Oscillations per minute at no load and full load.

c. Horsepower output.

d. Breakdown time (in hours) when saw is operated under full load current.

e. Temperature rise of motor housing, under full load current, during 5 minutes operating period.

4.4.2 Motor test. Each sample autopsy saw shall be tested to determine compliance with Underwriters' Laboratories, Inc. Standard UL 45 requirements for continuous rated universal motors for electric tools for:

a. Continuity of grounding connections.
b. Dielectric strength.
c. Insulation resistance.
d. Current input.
e. Temperature rise.
f. Current leakage.

4.4.3 General performance tests. Before bone saws have been tested as specified in 4.4.1 and 4.4.2, the bone saw motors shall be checked for switch operation, excessive noise and vibration.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One saw, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in Table IV shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table IV. Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	Not required	6 units

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:10Z
Check the source to verify that this is the current version before use.

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit package . Components shall be positioned in carrying case and cushioned in such a manner as to prevent shifting of or damage to the contents. The entire unit shall be further packaged in accordance with MIL-P-116, method IA-15.

5.1.4 Level C . Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A .

5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant, grade V3. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:10Z
Check the source to verify that this is the current version before use.

5.3 Marking.

5.3.1 Interior packages. Each unit package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Intended use . This bone cutting saw is to be used for autopsy purposes in cutting through human bones or for laboratory use involving small animals.

6.2 Ordering data. Procurement documents should specify the following:

(a) Title, number, and date of this specification.
(b) Selection of applicable levels of packaging and packing (See 5.1 and 5.2).
(c) Federsl stock number.

6.3 This specification covers the following item appearing in the Federal Supply Catalog as indicated below:

Federal Stock No.	Item Identification
6515-299-8717	SAW, BONE CUTTING, AUTOPSY, Stryker, 110 Volt, AC-DC

6.4 The reference to Stryker Corporation for blades (See 3.11) not furnished with the above stock number is presented as a possible source and is intended as a guide.

6.5 Supersession data. This specification includes requirements of Purchase Description No. 10, dated 12 April 1966.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1236

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026l0-21T09:10Z
Check the source to verify that this is the currentversion before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS. This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document (s) or serve to amend contractual requirements.

SPECIFICATION
MIL-S-36703

ORGANIZATION

CITY AND STATE

CONTRACT NUMBER

MATERIAL PROCURED UNDER A
☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING.

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?
   ☐ YES ☐ NO *(If "yes", in what way?)*

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)*

DATE

DD FORM 1 JAN 66 1426 REPLACES EDITION OF 1 OCT 64, WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:10Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

Fold

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

Fold

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:10Z
Check the source to verify that this is the current version before use.