MIL-S-36715A
27 September 1978
SUPERSEDING
MIL-S-36715
5 November 1970

MILITARY SPECIFICATION

SCISSORS, GENERAL SURGICAL, NELSON, 9 INCH

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers general surgical scissors of the Nelson design, suitable for use in special surgical procedures.

1.2 Classification. Scissors shall be of the following types, as specified (see 6.2):

Type I. - Curved
Type II. - Straight

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents of the issue in effect on date of invitation for bid or request for proposal form a part of this specification to the extent specified herein.

SPECIFICATIONS

FEDERAL

GG-I-526	Instruments, Dental and Surgical; General Specifications.
ZZ-R-690	Rubber Dam, Dental.
PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-636	Boxes, Shipping, Fiberboard.
PPP-B-676	Boxes, Setup.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 20261O-21T09:11Z
Check the source to verify that this is the current version before use.

MILITARY

MIL-P-116 Preservation-Packaging, Methods of.

STANDARDS

MILITARY

MIL-STD-105 Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 Marking for Shipment and Storage.

DRAWINGS

DEFENSE PERSONNEL SUPPORT CENTER (DPSC)

23185 - Scissor, General Surgical, Straight, Nelson 9 Inch
23186 - Scissor, General Surgical, Curved, Nelson 9 Inch

(Copies of specifications, standards, drawings and publications required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bid or request for proposal shall apply.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM)

ASTM E 18 Rockwell Hardness and Rockwell Superficial Hardness of Metallic Materials, Tests for.
ASTM E 92 Vickers Hardness of Metallic Materials, Test for.
ASTM E 140 Hardness Conversion Tables for Metals (Relationship Between Brinell Hardness, Vickers Hardness, Rockwell Hardness, Rockwell Superficial Hardness, and Knoop Hardness).

(Application for copies should be addressed to the American Society for Testing and Materials, 1916 Race Street, Philadelphia, PA 19103.)

(Technical society and technical association specifications and standards are generally available for reference from libraries. They are also distributed among technical groups and using Federal agencies.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:11Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 General requirements. The scissors shall be in accordance with the applicable requirements of GG-I-526, and as specified herein.

3.2 Material. The scissors shall be fabricated from class 4, type 420 or 420F material except that the carbon content shall be 0.30 to 0.40%, chromium content shall be 12.0 to 15.0%, molybdenum 0.60% maximum (optional), and vanadium 0.15% maximum (optional).

3.3 Hardness. After proper stress relieving, the scissors shall have a Rockwell hardness of C50 to C55 when tested as specified in 4.4.2. The requirements of GG-1-526 shall apply.

3.4 Style, design and dimensions. Shall be a Nelson lung dissecting scissors for use in special surgical procedures such as lung resections and general vascular surgery.

3.4.1 Type I scissors. Type I scissors shall be curved and of the Nelson design. Both points shall be blunt. Shall be in accordance with DPSC Drawing 23186.

3.4.2 Type II scissors. Type II scissors shall be straight and of the Nelson design. Both points shall be blunt. Shall be in accordance with DPSC Drawing 23185.

3.5 Finish. Finish shall be as specified on applicable DPSC drawings for appropriate type. The inside of the blades shall have a crocus finish as specified in GG-I-526.

3.5.1 Passivation. Scissors shall be properly passivated.

3.6. Cutting edges. Scissors shall cut a continuous length of at least 1-3/8 inches of the test material, when tested as specified in 4.4.1.

3.7 Identification marking. In addition to GG-I-526 requirements, instruments of foreign manufacture shall also be marked "stainless".

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:11Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the corrosion-resisting steel may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the National Stock Number (NSN), the item identification, the name of the material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and table I. The unit of product for examination purposes shall be one scissors.

TABLE I. Sampling for examination.

Type defect	Inspection level	AQL (percent defective)
Major	II	1.0
Minor	II	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and table II. The unit of product for test purposes shall be one scissors. Where feasible, the same sample unit shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 202640-21T09:11Z
Check the source to verify that this is the current version before use.

TABLE II. Sampling for tests.

Component	Characteristic	Requirement	Test procedure	Inspection level	AQL (percent defective)
Scissors	Hardness	3.3 and GG-I-526	4.4.2	S-2	2.5
	Dimensions	3.4	-----	S-2	2.5
	Scissors performance test	3.6	4.4.1	S-4	1.0
	Copper sulfate test	GG-I-526	GG-I-526	S-2	2.5
	Boil test	GG-I-526	GG-I-526	S-2	2.5

4.2.3 Raw materials. A sufficient quantity of each lot of metal used in the construction of the scissors shall be selected for chemical analysis.

4.3 Examination. The scissors shall be examined to determine compliance with all the requirements contained in this specification. Examination shall be conducted as specified in tables I and III. Examination, however, shall not be restricted to defects listed in table III. Any scissors in the sample containing one or more defects shall be rejected and if the number of defective scissors in any sample exceeds the acceptance number for that sample, the lot represented by the sample shall be rejected.

TABLE III. Classification of defects.

Categories	Defects
Major	
101	Scissors not of the proper type specified.
102	Points of scissors sharp.
103	Scissors not of finish specified.
104	Scissors not free from cracks, pits, burrs, rough areas, or tool marks.
105	Scissors not free from excessive play at fastening point.
Minor	
201	Identification markings not present, not complete, not correct or not legible.

4.3.2 Packaging inspection. The inspection of the preservation-packaging marking shall be in accordance with the group A and B quality conformance inspection requirements of MIL-P-116. The inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:11Z
Check the source to verify that this is the current version before use.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Scissors performance test. The test material shall be rubber dam meeting the requirements for Size 1 of ZZ-R-690. Three separate, consecutive tests shall be performed by using each scissors as described and specified in GG-I-526.

4.4.2 Hardness test. Hardness shall be determined in accordance with ASTM E18 or E92. Conversion of Vickers hardness values to Rockwell values shall be in accordance with ASTM E140.

5. PACKAGING

5.1 Preservation-packaging. Preservation-packaging shall be level A or C, as specified (see 6.2).

5.1.1 Unit of issue. One pair of scissors constitutes a unit of issue.

5.1.2 Level A.

5.1.2.1 Cleaning. Scissors shall be cleaned as specified in MIL-P-116, process C-1.

5.1.2.2 Drying. Scissors shall be dried as specified in MIL-P-116.

5.1.2.3 Preservative application. Preservatives shall not be used.

5.1.2.4 Unit packaging. Each pair of scissors shall be wrapped in neutral tissue or other suitable material and packaged in accordance with MIL-P-116, method IC-1.

5.1.2.5 Intermediate package. Twelve unit packages containing scissors of one type only shall be packaged in a box conforming to PPP-B-566, PPP-B-636, class domestic or PPP-B-676. Closure shall be as specified in the appendix of the applicable box specification.

5.1.3 Level C. Scissors shall be clean, dry and individually packaged in a manner that will afford adequate protection against corrosion, deterioration and physical damage during shipment from the supply source to the first receiving activity.

5.2 Packing. Packing shall be level A, B, or C, as specified (see 6.2).

5.2.1 Level A. Eight intermediate packages containing scissors of one type only shall be packed in a shipping container designed for a type 2 load and conforming to PPP-B-636, class weather-resistant. Closure, strapping, and waterproofing shall be as specified in the appendix of the box specification.

Source: https://assist.dla.mil -- Downloaded: 20260-21T09:11Z
Check the source to verify that this is the current version before use.

5.2.2 Level B. Eight intermediate packages containing scissors of one type only shall be packed in a shipping container designed for a type 2 load and conforming to PPP-B-636, class domestic. Closure shall be as specified in the appendix of the box specification.

5.2.3 Level C. The packaged scissors shall be packed in shipping containers in a manner that will afford adequate protection against damage during shipment from the supply source to the first receiving activity. These packs shall conform to the applicable carrier's rules and regulations.

5.3 Marking. Each unit package, intermediate package and exterior container shall be marked as specified in MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents shall specify the following:

a. Title, number and date of this specification.
b. Type required (see 1.2).
c. Applicable NSN.
d. Selection of applicable levels of packaging and packing required (see 5.1 and 5.2).

6.2 This specification covers the following items appearing in the Federal Supply Catalog:

NSN	Item Identification
6515-00-923-5000	SCISSORS, GENERAL SURGICAL, Curved, Nelson, 9 inch
6515-00-923-5004	SCISSORS, GENERAL SURGICAL, Straight, Nelson, 9 inch

6.3 Changes from previous issue. Asterisks are not used in this revision to identify changes with respect to the previous issue, due to the extensiveness of the changes.

Custodians:
Army - MD
Navy - MS
Air Force-03

Review activity:
DOD-MB

Preparing activity:
DLA-DM

Project No. 6515-2139

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:11Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:11Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS: This form is provided to solicit beneficial comments which may improve this document and enhance its use. DoD contractors, government activities, manufacturers, vendors, or other prospective users of the document are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity. A response will be provided to the submitter, when name and address is provided, within 30 days indicating that the 1426 was received and when any appropriate action on it will be completed.

NOTE: This form shall not be used to submit requests for waivers, deviations or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

DOCUMENT IDENTIFIER *(Number)* AND TITLE

MIL-S-36715A SCISSORS, GENERAL SURGICAL, NELSON, 9 INCH

NAME OF ORGANIZATION AND ADDRESS OF SUBMITTER

☐ VENDOR ☐ USER ☐ MANUFACTURER

1. ☐ HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE? ☐ IS ANY PART OF IT TOO RIGID, RESTRICTIVE, LOOSE OR AMBIGUOUS? PLEASE EXPLAIN BELOW.

A. GIVE PARAGRAPH NUMBER AND WORDING

B. RECOMMENDED WORDING CHANGE

C. REASON FOR RECOMMENDED CHANGE(S)

2. REMARKS

SUBMITTED BY *(Printed or typed name and address — Optional)*

TELEPHONE NO.

DATE

DD FORM 1426 1 OCT 76

EDITION OF 1 JAN 72 WILL BE USED UNTIL EXHAUSTED.

Source: https://assist.dla.mil -- Downloaded: 202640-21T09:11Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

POSTAGE AND FEES PAID



DEFENSE PERSONNEL SUPPORT CENTER
DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 20260-21T09:11Z
Check the source to verify that this is the current version before use.