MIL-T-36685
Amendment No. 1
2 May 1973

MILITARY SPECIFICATION

TUBE, CONNECTING, SYRINGE AND NEEDLE

This amendment forms a part of Military Specification MIL-T-36685, dated 13 August 1970, and is approved for use by all Departments and Agencies of the Department of Defense.

Page 5:

3.6 Security of assembly. Delete period at end of sentence and add the following: "at an inspection level of S-3 and AQL of 1.5."

Page 7:

4.2.2 Sampling for tests. In Table II for Security of assembly, delete "S-2" and "*" and substitute "S-3" and "1.5" respectively.

Page 9:

4.4.3 Security of assembly. Line 4, delete "4" and "15" and substitute "3" and "5".

4.4.6 Sterility test. Last line, delete "steorothermophiles" and substitute "stearothermophilus".

Page 11:

5.2.1.1 Exterior container. At end of paragraph add: "Fiberboard boxes shall conform to the special requirements as specified in Table II of PPP-B-636."

5.2.2.1 Exterior container. At end of paragraph add: "Closure of fiberboard boxes shall conform to method II as specified in PPP-B-636. In addition, fiberboard boxes shall conform to the special requirements as specified in Table I of PPP-B-636."

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1724

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:01Z
Check the source to verify that this is the current version before use.