MIL-T-36692B
29 September 1977
SUPERSEDING
MIL-T-36692A
18 March 1971

MILITARY SPECIFICATION

TUBE, DUODENAL, SURGICAL, RADIOPAQUE, CANTOR, 16 FR.

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a sterile, disposable, plastic surgical duodenal tube of Cantor design suitable for intestinal decompression of adult humans preceding, during and following surgery of the abdomen.

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents of the issue in effect on date of invitation for bids or request for proposals form a part of this specification to the extent specified herein.

SPECIFICATIONS

FEDERAL

MMM-A-250	-	Adhesive, Water-Resistant (For sealing Fiberboard Boxes).
PPP-B-636	-	Boxes, Shipping Fiberboard.
PPP-T-76	-	Tape, Pressure Sensitive Adhesive Paper, (For carton sealing).

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Headquarters, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing on the end of this document or by letter.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:08Z
Check the source to verify that this is the current version before use.

MILITARY

MIL-S-36586 - Sterilization Strip Test, Bacterial Spore.

STANDARDS

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.

DRAWING

Defense Personnel Support Center (DPSC)

20664 - TUBE, CANTOR

(Copies of specifications, standards, drawings, and publications, required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply.

UNITED STATES PHARMACOPEIAL CONVENTION, INC.

Pharmacopeia of the United States (USP)

(Application for copies should be addressed to the Mack Publishing Company, Easton, PA 18042.)

(Technical society and technical specifications and standards are generally available for reference from libraires. They are also distributed among technical groups and using Federal agencies.

3. REQUIREMENTS

3.1 Material.

3.1.1 Plastic tubing. Plastic tubing shall be medical grade polyvinyl chloride, sterile, non-toxic (safe), radiopaque, pyrogen-free and comply with the animal implantation test when tested as specified in section 4. The tubing shall contain less than 5 parts per million (ppm) of lead when tested as specified in 4.5.1. The tubing shall comply with the USP requirements for class IV plastics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:08Z
Check the source to verify that this is the current version before use.

3.1.2 Connector. The connector shall be fabricated of materials suitable for the intended purpose and shall be sterile, non-toxic and pyrogen-free when tested as specified in section 4. There shall be a rigid plastic tubular section inserted into the end of the duodenal tube to expand it and a short section of rubber tube over the exterior to serve as a female connector. The connector, plastic portion only, shall contain less than 5 (ppm) of lead when tested as specified in 4.5.1.

3.1.3 Rubber bag. The rubber bag shall be a laminate of natural gum rubber latex and neoprene rubber suitable in all respects for the intended purpose and sterile when tested as specified in 4.5.2.

3.2 Style, design, and dimensions. Shall be a surgical duodenal tube of Cantor design suitable for intestinal decompression of adult humans preceding, during and following surgery of the abdomen. The tube shall be 10 feet long and 16 Fr. in diameter. There shall be 8 eyes located in the distal end of the tube. The tube shall be marked to show approximate depths of insertion as follows: Stomach - 17 inches, Pylorus - 24 inches, and Duodenum - 30 inches, and 4, 5, 6, 7, 8, and 9 feet. The tube only shall be coated with a medical grade silicone oil. The bag shall be neoprene rubber on the exterior laminated to natural gum latex rubber on the interior. The bag shall be sealed and affixed to the distal end of the tube. Access to the bag shall be by hypodermic needle and syringe. The bag shall accommodate 10 cc of mercury. The surgical duodenal tube shall be in accordance with DPSC Drawing No. 20664.

3.3 Sterility. The duodenal tube shall be sterile when tested with a biological indicator as specified in 4.5.2.

3.3.1 Certificate of sterility. A certificate of sterility, stating that samples from each sterilizer lot have been tested and found to be sterile, shall be submitted to the procuring activity. The certificate of sterility shall include:

a. Item identification.
b. National stock number (NSN).
c. Contract or purchase order number.
d. Sterility lot or control number.
e. Date of sterilization.
f. Quantity of product in sterilizer lot.
g. Type of sterilization (gas, radiation, heat, steam, etc.,).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:08Z
Check the source to verify that this is the current version before use.

h. Form of biological indicator (ie., inoculated paper spore strip or carrier, inoculated product, or inoculated simulated product).
i. Test organisms.
j. Test medium, incubation temperature, incubation time, and dates of tests.
k. Number of test units.
l. Statement that samples comply with sterility tests.
m. Signature of head of laboratory or a designated qualified alternate.

When inspection is made at destination, the certificate of sterility shall be submitted by the contractor to the Directorate of Medical Materiel, ATTN: DPSC-ATQ, Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, PA 19101, and a copy forwarded with each shipment to the consignee. When inspection is made at source, certificate of sterility shall be furnished to the cognizant Government Quality Assurance Representative for submission to the procuring activity.

* 3.3.2 Sterility lot. A sterility lot is that single quantity of product subjected to the same manufacturing operation and sterilized simultaneously in the same sterilizer chamber. Each unit package shall be clearly marked with a sterility lot (control) number and date of sterilization as specified in 5.3.1. Each sterility lot shall be traceable to the pertinent manufacturing lot(s).

3.4 Freedom from pyrogens. Tubes shall be free from pyrogens when tested as specified in 4.5.3, except that the rubber bag need not be subjected to this test.

3.5 Safety (non-toxicity). Tubes shall be safe when tested as specified in 4.5.4. The rubber bag need not be subjected to this test.

3.6 Animal implantation. The plastic tubing only of the surgical duodenal tube shall show no evidence of toxicity or adverse tissue reaction upon gross and microscopic examination when tested as specified in 4.5.5.

3.7 Certification for freedom from pyrogens, safety, and animal implantation tests. A certificate(s) listing all lot or control numbers shall state that representative sample(s) from each lot or control have been tested and found to be safe, free from pyrogens, and have passed the animal implantation test. This certificate(s) shall include the specific tests performed and the test results. The record of the animal implantation test and its results shall be certified by the Pathologist who performed or supervised performance of the animal implantation test. For destination inspection only, one copy of certificate(s) shall be submitted by the contractor directly to the Directorate of Medical Materiel, ATTN: DPSC-ATQ, Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, PA 19101, and one copy shall be forwarded with each shipment to the consignee.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:08Z
Check the source to verify that this is the current version before use.

3.8 Security of assembly. The bag and connector shall withstand a three pound weight for 15 minutes when tested as specified in 4.5.6.

3.9 X-Ray Opacity. The plastic tubing shall throw a definite, uniform shadow when tested as specified in 4.5.7. The degree of opacity shall be such that the shadow of a small foreign body in front or rear of the tube shall be visible for diagnosis.

3.10 Bag suitability. The bag shall not exceed 5 inches in circumference when tested as specified in 4.5.8.

3.11 Instruction sheet. Each tube shall be supplied with an insert providing instructions for the preparation, insertion, and removal of the tube.

3.12 Workmanship. Workmanship shall be first class throughout. The surgical duodenal tube shall be free from defects which detract from its appearance or impair its serviceability. The surgical duodenal tube shall show smooth outer surfaces and be free from burrs. It shall be free from chips, soils, or other foreign matter.

4. QUALITY ASSURANCE PROVISIONS

* 4.1 Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties of the item).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:08Z
Check the source to verify that this is the current version before use.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as indicated in table I. The unit of product for examination purposes shall be one surgical duodenal tube.

TABLE I. Sampling criteria for examination.

Type of defect	Inspection level	AQL (percent defective)
Major	I	1.0
Minor	I	2.5

* 4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as indicated in table II. The unit of product for test purposes shall be one surgical duodenal tube. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

TABLE II. Sampling criteria for tests.

Component	Characteristic	Requirement	Test procedure	Inspection level	AQL (percent defective)
Plastic tubing	Lead	3.1.1	4.5.1 & 4.5.1.1	S-1	1.5
Plastic portion of connector	Lead	3.1.2	4.5.1 & 4.5.1.1	S-1	1.5
Unit of product	Dimensions	3.2	--	S-2	2.5
Unit of product	Sterility	3.3	4.5.2	S-2	1.0
Plastic tubing	Freedom from pyrogens	3.4	4.5.3	S-2	1.0
Plastic tubing	Safety	3.5	4.5.4	S-2	1.0

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:08Z
Check the source to verify that this is the current version before use.

TABLE II. Sampling criteria for tests. - Continued

Component	Characteristic	Requirement	Test procedure	Inspection level	AQL (percent defective)
Plastic tubing	Animal implantation	3.6	4.5.5	S-2	1.0
Unit of product	Security of assembly	3.8	4.5.6	S-3	0.4
Plastic tubing	X-ray opacity	3.9	4.5.7	S-2	1.0
Bag	Bag suitability	3.10	4.5.8	S-2	1.0

4.3 Examination. The surgical duodenal tube shall be examined to determine compliance with specification requirements. Examination shall be conducted as indicated in tables I and III. Examination, however, shall not be restricted to the defects listed in table III.

TABLE III. Classification of defects

Categories	Defects
Major	
101	Defective connector.
102	Absence of required perforations.
103	Exterior of tubing not coated with silicone oil.
104	Markings incomplete, missing, or illegible.
105	Permanent kinks in tubing restricting flow.
Minor	
201	Presence of foreign material or loose particles.
202	Instruction sheet not present.

4.4 Packaging inspection. The inspection of the unit, intermediate package, packing and marking for shipment and storage shall be in accordance with quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:08Z
Check the source to verify that this is the current version before use.

4.5 Tests. Tests shall be conducted to determine compliance with the specification requirements.

4.5.1 Test for lead. Ten grams of sample shall be taken, ignited to char, and then further ignited in a muffle furnace at 700° Centigrade (C) to a white ash. The ash shall be boiled in 15 ml. of 10 percent nitric acid, filtered into a volumetric flask and made up to 100 ml. An aliquot portion, equivalent to one gram of sample, shall be taken and placed in a separator. The dithizone test for lead, as given in the Pharmacopeia of the United States (USP) shall be applied using a control of 5 ppm or 12 ppm of lead solution as given in the USP, test. If the sample solution is darker than the limit, smaller aliquots shall be used until the color of the control is matched by the sample solution and the concentration of lead in ppm is determined.

4.5.1.1 Alternate test for lead. Accurately weigh a 5-gram sample into a 1-oz. Vycor crucible. Reduce to ash in a muffle furnace at 650° C for eight hours. Let cool to room temperature and add 15 ml of 1:1 hydrochloric acid. Bring to a boil on a hot plate, allow to boil for 2 to 3 minutes, remove, let cool, and filter through Whatman #2 filter paper into a 50-ml volumetric flask. Rinse the residue on the filter with deionized water and dilute to 50 ml. Prepare standard lead solutions of 1, 2 and 5 ppm lead. Determine the absorbances of these solutions at the 283 millimicron wave length using a Hollow Cathode Pb lamp and a 10x scale expansion on a spectrophotometer, such as the Perkin-Elmer Model 303 Atomic Absorption Spectrophotometer. Determine the absorbance of the sample solutions using the same conditions as with the standards.

Calculation:

$$\frac{A \times F \times \text{dilution}}{W} = \text{ppm lead}$$

A = Absorbance
F = Factor = $\frac{(\text{Concentration standard in ppm Pb})}{\text{absorbance}}$ 1/
W = Sample weight in grams

1/ The three standards are averaged to give ppm/unit A.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:08Z
Check the source to verify that this is the current version before use.

* 4.5.2 Sterility test. The duodenal tube shall be tested for sterility in accordance with the method specified in the USP for non-parenteral devices using inoculated product, inoculated simulated product, or a paper spore strip (carrier). Product containing a biological indicator shall be marked or labeled so that it clearly differentiates from product intended for distribution. The incubation medium and minimum time in each case shall be soybean casein digest for 7 days for inoculated product, inoculated simulated product, or paper spore strip (carrier). The incubation temperature shall be as indicated paragraph 4.5.2.1. The minimum number of test units, regardless of lot size, shall be 10, except for inoculated carrier which shall be 20 for Type I Sterilization and 30 for Type II Sterilization including inoculated carrier and product test units. The inspection level shall be as indicated in table II. Sampling criteria for tests.

* 4.5.2.1 Inoculated product or simulated product. For each type of sterilization, the organism used shall be as listed below:

Type of Sterilization	Organism	Incubation Temperature
Steam	B. stearothermophilus	55° - 65°C
Dry Heat	B. stearothermophilus, or	55° - 65°C
	B. subtilis var. niger (or globigii)	30° - 35°C
Gas	B. subtilis var. niger (or globigii)	30° - 35°C
Radiation	B. pumilus	30° - 35°C

The number of viable cells shall be sufficient to provide assurance that the product has been sterilized.

* 4.5.2.2 Inoculated carrier. Inoculated carriers (paper spore strips) if used shall be in accordance with MIL-S-36586. Other paper carriers shall be inoculated to be of equivalent resistance. The method of incubation and number of test units shall be as indicated in 4.5.2 Sterility test and 4.5.2.1. Inoculated product or simulated product. The method of sterilization will control the one organism in the spore strip to be tested for as shown in 4.5.2.1 NOTE: When an inoculated carrier is chosen, product samples must also be tested as required by the USP.

* 4.5.2.3 Alternate method. Suppliers conducting sterility tests or designating lots in a manner other than that indicated herein shall submit proof that the alternate method provides equal assurance concerning the sterility of the product. However, all sterility tests shall be conducted using biological indicators. Requests for approval of any alternate method of testing for sterility shall be made to the Contracting Officer, and approval obtained before employment of the alternate method on a government contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:08Z
Check the source to verify that this is the current version before use.

* 4.5.2.4 Product testing. Testing for sterility of the product after receipt by the government shall be without a biological indicator and in accordance with the USP.

4.5.3 Freedom from pyrogens. Surgical duodenal tubes shall be tested for freedom from pyrogens in accordance with USP for transfusion and infusion assemblies.

4.5.4 Safety (non-toxicity). The surgical duodenal tubes, including the plastic component of the connector, shall be tested in accordance with the USP for transfusion and infusion assemblies.

4.5.5 Animal implantation test. Samples of the plastic tube only shall be tested in accordance with the method given in the USP for Implantation Test, except for the following changes: Last paragraph, second line, after "72 hours" add "and not to exceed 144 hours." Last paragraph, delete third and fourth sentences in entirety and substitute: "The tissues surrounding the implants shall be examined, both grossly and microscopically, by a pathologist certified by the American Board of Pathology or The American College of Veterinary Pathologists."

4.5.6 Security of assembly. The tubing shall be severed, one part consisting of ten inches of tubing plus the bag. The security of assembly of the bag to the tubing shall then be tested by gradually applying a three-pound weight to the severed end of the tube while the bag is held in a clamping device so that the tubing hangs vertically downward. The weight shall be applied to the bag for 15 minutes.

4.5.7 X-ray opacity. Each sample of a surgical duodenal tube shall be radio-graphed according to the following conditions:

85 KVP, 10 milliampere-seconds; 25 inch distance, 24 centimeters of paraffin USP superimposed over the duodenal tube and a standard make of cassette with screens, films and processing solutions. The foreign body used for test purposes shall be small fragment of sheet lead 1/2 millimeter in thickness.

4.5.8 Bag suitability. The bag shall be injected at its center with 10 cc of air from hypodermic syringe equipped with a 21-gage needle.

4.5.9 Class IV Plastics. The methods of tests for Class IV plastics, shall comply where applicable with the methods as indicated above. For tests other than those specified herein, the USP methods should be followed.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:08Z
Check the source to verify that this is the current version before use.

* 5. PACKAGING

5.1 Packaging. Packaging shall be level A, or C, as specified (see 6.2).

5.1.1 Unit of issue. One package containing ten individual packaged tubes, as specified, constitutes one unit of issue.

5.1.2 Level A.

5.1.2.1 Individual container. Each tube shall be neatly coiled in such a manner to prevent kinking. The coiled tube shall be packaged in a sealed clear or transparent, peel-open type bag capable of maintaining sterility of contents unless damaged or opened. The peel-open type bag shall be designed with tabs to facilitate opening.

5.1.2.2 Unit package. Ten individually packaged tubes shall be packaged in a box conforming to PPP-B-636, type CF, class weather-resistant. The box shall be closed by taping the tabs and bottom flaps with 2 inch wide tape conforming to PPP-T-76. Closing of bottom flaps with adhesive conforming to MMM-A-250 or metal stitches is permitted.

5.1.3 Level C. Tubes shall be clean, dry and individually packaged sterile. Ten individually packaged tubes shall be further packaged in a manner that will afford adequate protection against deterioration and physical damage during shipment from the supply source to the first receiving activity.

5.2 Packing. Packing shall be level A, B or C, as specified (see 6.2).

5.2.1 Level A. Four unit packages shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-636, class weather-resistant. Closure, strapping and water-proofing shall be as specified in the appendix of the box specification.

5.2.2 Level B. Four unit packages shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-636, class domestic. Closure shall be as specified in the appendix of the box specification.

5.2.3 Level C. The packaged tubes shall be packed in shipping containers that will afford the degree of protection against damage during direct shipment from the supply source to the first receiving activity. These packs shall conform to the applicable carriers rules and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:08Z
Check the source to verify that this is the current version before use.

5.3 Marking.

5.3.1 Individual package. Each bag, in addition to commercial markings, shall be marked with lot (control) number, the word "STERILE" and the legend "STERILITY GUARANTEED UNLESS PACKAGE IS DAMAGED OR OPENED". In addition, an instruction sheet shall be provided as specified in 3.11. The lot (control) number and the sterility legend shall be shown by use of a pressure-sensitive label. The lot number shall be preceded by the words "Lot No." The hot-stamping of the lot number in the heat seal area is not permitted. As an alternate, the sterility legend may be permitted on the instruction sheet which shall be visible through the bag.

5.3.2 Unit package. Each unit package shall be marked in accordance with MIL-STD-129. Contract or purchase order number and name of contractor and the sterility lot (control) number shall be shown. In addition, the method of sterilization (G-Gas, R-Radiation, C-Cold, S-Steam, D-Dry Heat Sterilization) shall be shown following the level of packaging and the date of packaging. Example "A-6/76-G." Final sterilization shall be on or after indicated packaging date.

5.3.3 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129. Marking shall include the sterility lot (control) number and the method of sterilization marking as specified in 5.3.2.

6. NOTES

6.1 Intended use. The surgical duodenal tube (Cantor) is a mercury-weighted, radiopaque, single lumen, polyvinyl plastic tube with a neoprene-latex bag intended for passage through a naris, the nasopharynx and pylorus into the small intestine for decompression. Its movement down the alimentary tract is facilitated by the free-flowing nature of mercury and the peristaltic action on the bolus formed by the mercury in the bag.

6.2 Ordering data. Procurement documents should specify the following:

(a) Number, title, and date of this specification.
(b) Applicable levels of packaging and packing (see 5.1 and 5.2).
(c) National stock number.

6.3 Federal Supply Catalog. This specification is applicable to the following Federal Supply Catalog item:

National Stock Number	Item Identification
6515-00-065-9599	TUBE, DUODENAL, SURGICAL, Radiopaque, Cantor, 16 Fr.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:08Z
Check the source to verify that this is the current version before use.

6.4 Changes from previous issue. The margins of this specification are marked with an asterisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and Contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the previous issue.

Custodians:
Army - MD
Navy - MS
Air Force - 02

Review activity:
DOD-MB

Preparing activity:
Defense Logistics Agency - DM

Project No. 6515-2042

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:08Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:08Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS: This form is provided to solicit beneficial comments which may improve this document and enhance its use. DoD contractors, government activities, manufacturers, vendors, or other prospective users of the document are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity. A response will be provided to the submitter, when name and address is provided, within 30 days indicating that the 1426 was received and when any appropriate action on it will be completed.
NOTE: This form shall not be used to submit requests for waivers, deviations or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

DOCUMENT IDENTIFIER *(Number)* AND TITLE

MIL-T-36692B Tube, Duodenal, Surgical Radiopaque, Cantor 16 FR.

NAME OF ORGANIZATION AND ADDRESS OF SUBMITTER

☐ VENDOR ☐ USER ☐ MANUFACTURER

1. ☐ HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE? ☐ IS ANY PART OF IT TOO RIGID, RESTRICTIVE, LOOSE OR AMBIGUOUS? PLEASE EXPLAIN BELOW.

A. GIVE PARAGRAPH NUMBER AND WORDING

B. RECOMMENDED WORDING CHANGE

C. REASON FOR RECOMMENDED CHANGE(S)

2. REMARKS

SUBMITTED BY *(Printed or typed name and address — Optional)*	TELEPHONE NO.
	DATE

DD FORM 1 OCT 76 1426 EDITION OF 1 JAN 72 WILL BE USED UNTIL EXHAUSTED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:08Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

POSTAGE AND FEES PAID

U.S. MAIL

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:08Z
Check the source to verify that this is the current version before use.