
FED. SUP CLASS
6640



MATERIAL: Rack shall be made of high density polyethylene plastic.

DESIGN: The rack shall have a sloping face with a drain trough and tube. It shall have a total of 30 pegs, 16 pegs 6.35 mm diameter x 76 mm long, 2 pegs 6.35 mm diameter x 25.4 mm long, 8 pegs 12.7 mm diameter x 114 mm long, 4 pegs 12.7 mm diameter x 76 mm long. The overall dimensions are 560 mm wide x 640 mm high x 190 mm deep. The rack is for hanging at the rear of a laboratory bench, so that the drain tube will drain into a sink.

NOTE: 1. Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.

USER ACTIVITIES: ARMY - MD, DM
NAVY - MS

REVIEWER ACTIVITIES: ARMY - ME
NAVY - OS
AIR FORCE - 03

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

APPROVED 26 JUL 1982 REVISED

P.A. EA / Other Cust MS 99	INTERNATIONAL INTEREST	TITLE: RACK, DRYING, LABORATORY	MILITARY STANDARD MS36394
Procurement Specifications: NONE	SUPERSEDES:		SHEET 1 OF 1

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE.

PROJECT NUMBER 6640-0797
STANDARD DATE

Source: https://assist.dla.mil -- Downloaded: 2025-07-17 15:57:52
Check the source to verify that this is the current version before use.