FED. SUP CLASS
6640



PART NUMBER	TUBE SIZE STUBS GAUGE	OD
MS36480-1	18	1.25 mm
MS36480-2	20	0.89 mm
MS36480-3	27	0.41 mm

MATERIAL: The connector shall be made of stainless steel hypodermic needle size tubing. The handle shall be made of polypropylene.

DESIGN: The handle shall be [illegible] the hypodermic tubing and shall [illegible] at loop for tied down and hang[illegible]

NOTES: 1. [illegible] standard by [illegible]ensions are [illegible]ment, provided [illegible]mance and are [illegible]uirements

CANCELLED
CANCELLED AFTER 29 AUG 1990
NO SUPERSEDING STANDARD

Ⓐ CANCELLATION

APPROVED 12 FEB 1982 REVISED Ⓐ 29 AUG 1990

USER ACTIVITIES:

REVIEWER ACTIVITIES: ARMY - MD
NAVY - OS
AIR FORCE - 01

ANSC N/A
DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for competitive use shall be made from this document

P A. EA	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
Other Cust MS 99		CONNECTOR, MICRO	MS36480
Procurement Specifications None	SUPERSEDES:		SHEET 1 OF 1

DD FORM 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE Project Number 6640-1636-01
CANCELLATION

Source: https://assist.dla.mil -- Downloaded: 202640-23T17:41Z
Check the source to verify that this is the current version before use.