
FED. SUP CLASS
6640



PART NUMBER	NOMINAL DIMENSIONS (IN.)				
	A WIDTH OF JAW	B DEPTH OF JAW	C OVERALL LENGTH	D MAX JAW GRIPPING CAPACITY	E WIRE DIAMETER
MS36009-1	1.25	0.75	5.0	1.50	0.090
MS36009-2	1.50	1.0	6.0	1.75	0.090

MATERIAL: Clamps shall be of spring brass wire.

FINISH: Nickel or chrome plated.

DESIGN: Clamps with finger grips shall be made of one piece construction formed to the approximate shape as shown. The tip shall close when the clamp is relaxed, and when depressed shall open to accommodate capacity as specified.

NOTES:
1. Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.
2. For design features purposes, this standard takes precedence over procurement document referenced herein.
3. Referenced documents shall be of the issue in effect on date of invitations for bid.
4. FSCM: 64484.

Ⓑ ENTIRE STANDARD REVISED

P.A.	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
EA		CLAMP, TEST TUBE, STODDARD TYPE W/FINGER GRIPS	MS36009
Other Cust MS 99			
Procurement Specifications MIL-C-3648	SUPERSEDES:		SHEET 1 OF 1

DD, FORM, MAY 73, 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE PROJECT NUMBER 6640-1310

APPROVED 23 Dec 1959 REVISED Ⓐ 15 Dec 1967 Ⓑ 4 Dec 1985

USER ACTIVITIES: ARMY - ME, NAVY - MS, AIR FORCE - 03

REVIEWER ACTIVITIES: ARMY - MD, NAVY - AS

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited. This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:32Z
Check the source to verify that this is the current version before use.