FED. SUP CLASS
6640



DIMENSIONAL DATA				
MORTAR			PESTLE	
PART NUMBER	TOP OD mm ± 5%	CAPACITY ml ± 10%	PART NUMBER	LENGTH mm ± 5%
MS36106-1	210 (8.2")	946 (32 oz)	MS36106-2	244 (9.6")
MS36106-3	79 (3.1")	60 (2 oz)	MS36106-4	102 (4")
MS36106-5	89 (3.5")	120 (4 oz)	MS36106-6	127 (5")
MS36106-7	108 (4.2")	240 (8 oz)	MS36106-8	152 (6")
MS36106-9	133 (5.2")	473 (16 oz)	MS36106-10	190 (7.5")

MATERIAL: Mortar and pestle shall be of soda lime glass in accordance with Federal Specification DD-G-541, Type II.

DESIGN: Mortar shall be bowl-shaped with pour-out. Base shall be flat with a diameter large enough to provide a stable resting surface.

Pestle shall be designed to afford hand comfort in grinding. The grinding end shall be rounded to facilitate grinding.

Mortar and pestle shall be in accordance with Federal Specification NNN-M-560, Type II, Class A.

FINISH: All surfaces shall be smooth.

NOTES:
1. For design feature purposes, this standard takes precedence over procurement documents referenced herein.
2. Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements as specified herein.
3. Referenced documents shall be of the issue in effect on date of invitations for bid.

(B) ENTIRE STANDARD REVISED

APPROVED 22 September 1960 REVISED (A) 31 May 1967 (B) 1 MAR 1982

USER ACTIVITIES: ARMY - CE, MI

REVIEWER ACTIVITIES: ARMY - ME, MD
NAVY - YD
AIR FORCE - 03

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

P.A. EA	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
Other Cust MS 99		MORTAR AND PESTLE, LABORATORY glass	MS36106
Procurement Specifications NNN-M-560	SUPERSEDES:		SHEET 1 OF 1

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

PROJECT NUMBER 6640-0616
REVISION DATE

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:34Z
Check the source to verify that this is the current version before use.