A-A-54032
24 November 1989

COMMERCIAL ITEM DESCRIPTION

REFRIGERATOR, MORTUARY

Mobile, Two Body Capacity

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a mobile, mortuary refrigerator capable of operating from stationary electrical power source or a portable generator.

Salient characteristics:

The unit shall be a mobile mortuary refrigerator capable of storing two bodies at a temperature range of at least 33°F to 37°F when ambient air temperature is 95°F plus or minus 5°F. The refrigerator shall consist essentially of the cabinet, carriage assembly, and refrigeration system. The refrigerator shall be of sectional design and of a size to permit passage of sections through a 32 inch by 80 inch doorway. The structural framing shall be designed to provide support for the fully loaded refrigerator. The joints of all sections of the refrigerator shall be provided with gaskets to prevent the infiltration of air and moisture. The space between the inner liner and outer shell of the refrigerator, including the doors, shall be completely filled with insulation. The insulation shall be industrial-type fiberglass insulation board, expanded polystyrene, or foamed-in-place urethane. The inner liners and outer shells of the refrigerator, including the doors, shall be fabricated of aluminum or stainless steel. The exterior shall be fabricated of aluminum or stainless steel. The aluminum shall have a stucco embossed finish or a bright, smooth finish.

The door shall be furnished with a replaceable cushion type gasket. The entire unit shall be able to be easily cleaned and disinfected. Shall be provided with four casters at least 5 inches in diameter. Two shall be rigid and two shall be of the swivel type with wheel brake.

AMSC N/A FSC 4110

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Each door shall be supported by two hinges and secured, when in a closed position, by a latch and strike assembly with sufficient force to affect a gasket seal. The latch and strike assembly shall be of the adjustable type and shall have locking facilities capable of taking a 1-3/4 inch padlock.

Each door shall have a card holder mounted to the outside surface and shall be fabricated of the same material as the outer skin of the door.

The cadaver compartment shall be equipped with a mortuary rack consisting of rack frame, carriage assembly, and one removable tray. The tray and carriage shall support a 500 pound load without sagging or racking. The refrigerator shall have rollers large enough to allow movement of unit by one person over various surfaces. Bumpers shall be provided on the carriage to absorb any shock of telescoping the carriage assembly. Body trays shall be fabricated of one-piece stainless steel and shall have an overall minimum dimension of 22 inches wide by 83 inches long by 2-3/4 inches deep. The trays shall be provided with handles or hand holes at each end for lifting. Body trays shall be mounted on rollers that allow for easy exposure or transfer of cadaver.

The refrigeration system shall consist of an electric motor-driven, 1/2 HP air-cooled condensing unit, evaporator and controls. The refrigeration system shall be designed to operate from a stationary electrical power source or a portable generator capable of providing 110 volts, 50 Hz/60 Hz, single phase.

The refrigerating equipment shall be enclosed in the top section of the refrigerator with adequate ventilating grills or louvers. Provisions shall be made for uniform air distribution within the refrigerated space.

The overall dimensions of the mortuary refrigerator shall not exceed 100 inches in length by 38 inches in width and 70 inches in height.

The mortuary refrigerator compressor shall comply with Underwriters Laboratories, Inc. Standard No. 471.

Workmanship. The mortuary refrigerator shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One mortuary refrigerator, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one mortuary refrigerator as specified in this description.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests (dimensional) shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-3. The unit of product for test purposes shall be one mortuary refrigerator.

Test. Each mortuary refrigerator shall be operated in accordance with the manufacturer's instructions to determine compliance with requirements herein.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

4110-01-266-3757

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

PHS
FDA-MPQAS

Project No. 4110-0426

Location:
CID/REFRIGERATOR/H1B-33

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC- RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-54032	REFRIGERATOR, MORTUARY, Mobile, Two Body Capacity

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*:

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.