A-A-54038
30 March 1990

COMMERCIAL ITEM DESCRIPTION

SUCTION APPARATUS, OROPHARYNGEAL, TRACHEAL

Portable, Battery Operated, (Aeromedical)

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a fully integrated, battery operated, portable oropharyngeal, tracheal suction apparatus, with EMI/RFI suppression, self contained in a carrying case.

Salient characteristics:

The suction apparatus shall be a portable, self-contained, oropharyngeal/tracheal suction apparatus with EMI and RFI suppressed circuitry. It shall be ready for immediate use, as received, without the use of tools. It shall include the following items:

a. Rechargeable power pack with controls and rechargeable batteries.

b. Battery charger.

c. Motor and pump.

d. EMI/RFI suppression circuitry.

e. A two speed selector control.

f. Vehicle electric cord, minimum 6 feet long.

g. Vacuum bottle, 600 ml minimum.

h. Suction tubing, minimum 45 inches long.

i. Operating instructions and other documentation.

j. Carrying case.

The suction apparatus shall be suitable for use in aspirating secretions, blood, or vomitis in emergency treatment of unconscious or injured personnel in desert, tropic, arctic, or aeromedical environments.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

The apparatus shall include controls to select power on/off and low/high speed; indicating lamps for power on and battery charger on; circuit breaker/fuse to protect motor from overload; adjustable vacuum valve to regulate maximum vacuum; and gauge to indicate actual vacuum level.

The apparatus shall not include an automatic shut-off. The design of the motor-pump shall be such that overflow of patient debris shall not clog or jam it during use and it shall be readily cleanable.

The apparatus shall operate from its internal batteries (power pack), an external 12 VDC 120 amp-hour heavy-duty vehicle battery, and any 5 amp DC service with voltage within range of 12 to 15 volts.

The dual power source battery charging circuit shall be capable of accepting input power from a vehicle generating system rated 120 amperes for 12 volt system, vehicle 12 volt battery and an AC system of 115 volt, 50/60/400 Hz.

The collection container and tubing shall be sterilizable by ethylene oxide gas or disinfected by cold sterilizing solution. Shall be easily removable without tools for emptying and cleaning.

The complete suction apparatus shall be housed in a rigid or semi-rigid, flame retardant, mildew resistant, plastic case with handle(s), hinge and closure latches. The top of the case shall be permanently marked as follows: EMI/RFI Filtered for Aeromedical Use.

It shall provide user/maintenance manuals (2) which will meet as a minimum ASTM F 960 requirements.

The complete apparatus in its carrying case, less the above mentioned manuals, shall not weigh more than 20 pounds. The dimensions of the carrying case shall not exceed 20 inches by 15 inches by 12 inches high.

The suction apparatus shall meet the performance requirements listed in this specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Performance requirements:

	Characteristic	Performance
a.	Minimum suction capacity.	500 mm Hg
b.	Operating temperature range.	from (-)4°F to (+)115°F
c.	Recharge time	24 hours, AC/DC sources
d.	Free airflow at patient end of suction tube	30 LPM minimum on high speed, and 13.5 minimum on low speed.
e.	Run time (four tests: two requirements, one per extreme of ambient temperature requirement) for both high and low speed. Free airflow at patient end of suction tube	60 minutes minimum at low speed while moving a minimum of 13.5 LPM (both temperature extremes) 35 minutes minimum at maximum speed while moving a minimum of 30 LPM.
f.	Collection container minimum capacity	600 ml minimum

The apparatus shall meet the requirements of ASTM F-960 excluding the sections dealing with filters and pleural/closed wound drainage vessels. It shall also meet the leakage current requirements of UL and CSA.

Workmanship. The suction apparatus shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One suction apparatus along with two sets of operating and maintenance manuals, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for Inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with a AQL of 2.5 (percent defective) and an inspection level of S-1. The unit of product for examination purposes shall be one suction apparatus, as specified in this description.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-2 The unit of product for test purposes shall be one suction apparatus.

First article. Two first article samples are required; the first dedicated solely for EMI/RFI testing, the second for all other testing required by this specification.

Tests. Tests shall be performed to assure compliance with performance requirements herein and in accordance with test requirements in ASTM F960. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE. The following National Stock Number is covered by this document:

6515-01-284-8704

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4653

Location:
CID/SUCTION/APPAR/ORPHO/TRAC/B-1D-35

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-54038	SUCTION APPARATUS, OROPHARYNGEAL, TRACHEAL

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION (Mark one)
	☐ VENDOR
	☐ USER
3b. ADDRESS (Street, City, State, ZIP Code)	☐ MANUFACTURER
	☐ OTHER (Specify)

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER (Last, First, MI) – Optional	7b. WORK TELEPHONE NUMBER (Include Area Code) – Optional
7c. MAILING ADDRESS (Street, City, State, ZIP Code) – Optional	8. DATE OF SUBMISSION (YYMMDD)

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4956 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.