A-A-54042
19 January 1990

COMMERCIAL ITEM DESCRIPTION

PACEMAKER, CARDIAC

External Type, Battery Powered

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers an external, battery-powered, cardiac pacemaker.

Salient characteristics:

The cardiac pacemaker shall be an external, battery-powered, cardiac pacer designed to provide temporary R-wave inhibited or asynchronous pacing for emergency, prophylactic, and pre- or post-operative patient management.

The cardiac pacemaker shall be contained in a plastic housing approximately 1.5 inches in depth, 3 inches in width, and 6 inches in height maximum.

The cardiac pacemaker shall have functional parameters as detailed herein.

- Pulse rate shall be 50-150 PPM, plus or minus 5 percent.
- Pulse width shall be 1.8 mS, plus or minus 0.2 mS.
- Output shall be 2-20 mA, plus or minus 15 percent.
- Refractory period shall be 330 mS, plus or minus 30 mS.
- Sensitivity shall be 1.0-10 mV, plus or minus 30 percent.

The cardiac pacemaker shall have touch or rotary controls, VVI and V00 modes, push or twist-to release lead connectors for patient cable pins or standard leadwires (0.030-0.085 inches (diameter), and a hinged or sliding clear plastic cover over all controls.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

The cardiac pacemaker shall have a durable plastic carrying case approximately 13-1/2 inches long, 10 inches wide, and 4 inches deep.

It shall be capable of withstanding sterilization with ethylene oxide gas at a temperature not to exceed 120°F.

The manufacturer shall have notified the FDA, and shall have received a 510 (K) number.

Metal parts shall be fabricated of a suitable corrosion-resisting steel, properly hardened and passivated.

If powered by an alkaline battery the cardiac pacemaker shall be capable of 30 days of continuous operation.

The cardiac pacemaker shall be protected against defibrillator countershock to 400 joules.

A low battery condition shall be indicated via loss of LED function, LED flashing, or via meter.

The hard carrying case shall include the pacer and its accessories as described herein:

- The cardiac pacemaker shall have alkaline or lithium batteries.
- Two straps shall be provided - one body and one arm, with hook and loop closure.
- A patient cable shall be included.
- A physician's manual shall be included.

The unit shall be fabricated from materials which are inherently resistant to corrosion or which are treated to render them corrosion-resistant. Where painting or plating is used, appropriate cleaning and priming shall be performed prior to finishing. Unless otherwise specified, finish shall be equivalent or superior to that of comparable commercial equipment, and shall be free of gaps, scratches, chips and other imperfections.

Workmanship. The cardiac pacemaker shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One cardiac pacemaker, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one cardiac pacemaker, as specified in this description.

Sampling for tests. Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-1. The unit of product for test purposes shall be one cardiac pacemaker.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics. Tests shall include, but shall not be limited to the following:

Performance test. The cardiac pacemaker shall be completely assembled in its operating configuration and tested to determine compliance with this specification. Testing shall include Acceptance Testing.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Acceptance testing. The cardiac pacemaker must pass the tests as follows:

- Pulse rate range. Measure the pulse rate range. It shall be at least 50 to 150 PPM, plus or minus 5 percent.

- Pulse width. Measure the pulse width. It shall be 1.8 mS, plus or minus 0.2 mS.

- Output range. Measure the output range. It shall be at least 2 to 20 mA, plus or minus 15 percent.

- Sensitivity range. Measure the sensitivity range. It shall be at least 1.0 to 10 mV, plus or minus 30 percent.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6515-01-169-0282

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4195

Location: CID/PACEMAKER/H17-35

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

Fold along this line

DEFENSE LOGISTICS AGENCY

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4956 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-54042	PACEMAKER, CARDIAC

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR ☐ USER ☐ MANUFACTURER ☐ OTHER *(Specify)*
b. ADDRESS *(Street, City, State, ZIP Code)*	

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 1426 82 MAR PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.