A-A-54043
30 Nov 1989

COMMERCIAL ITEM DESCRIPTION

OTOSCOPE AND OPHTHALMOSCOPE SET, BATTERY TYPE

The General Services Administration has authorized the use of this Commercial Item Description to replace Federal Specification GG-O-770D which is being cancelled.

This commercial item description covers a combination otoscope and ophthalmoscope set suitable for use in the examination of and surgery within the ear and in the examination of the fundus of the eye. A power handle shall be used common to the two heads.

Salient characteristics:

General. The set shall consist of the following components:

Otoscope head.

Ophthalmoscope head.

Battery handle without batteries.

Three (3) aural specula; large, medium and small.

One (1) nasal speculum.

Two (2) replacement lamps for otoscope head.

Two (2) replacement lamps for ophthalmoscope head.

Carrying case.

Material. The components shall be made of the following material(s) as specified:

Battery Handle - copper alloy housing and plastic electrical controls.

Otoscope Head - metal housing, plastic specula holder, plastic resin or glass magnifier, stainless steel pivot screw.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Ophthalmoscope Head - hard plastic or metal housing, metal bayonet connector, optical grade plastic resin or glass lenses and filters. Rubber brow rest.

Specula - nylon or polypropylene.

Carrying Case - hard or soft outside, plastic form fitting insert, corrosion-resisting metal hinge.

Style and design. The battery handle shall be able to hold two "D" size batteries, either in carbon-zinc or alkaline configuration including Eveready E 95 type without shorting. The metal connector to the head shall be spring-loaded, bayonet-type having an indexing lock mechanism. The battery handle shall be provided with a rheostat and a pushbutton "on" interlocking switch.

The ophthalmoscope head shall have a mirror reflecting system to provide not less than ten (10) lumens per square foot axial illumination at a point 7-3/4 inches from the upper edge of the mirror, when the lamp is operated at not more than 2.70 volts. It shall not be necessary to manipulate any part of the head to obtain optimum illumination. The axis of the filament image shall be fixed and essentially parallel to the upper edge of the mirror. The illuminated spot produced by the ophthalmoscope shall be uniform and free from distortions. It shall consist of at least 23 lenses plus one zero aperture, mounted upon a single dial. The minimum lens range shall be from minus 35 diopters to plus 20 diopters. As an alternate, range of minus 20 to plus 40 diopters is acceptable. Each lens, when it is in the selected lens aperture, shall be identified with its numerical power, through a magnified window with internal light. The aperture and filter system shall contain a minimum of five different positions: pinhole, normal aperture, clear slit, red-free filter and either a grid or color conversion filter. Each aperture and filter shall be capable of being used with any lens selected.

The ophthalmoscope lamp housing shall be designed to insure proper positioning of the lamp in respect to the optical system.

The ophthalmoscope lamp shall be a single-contact type, close-coiled filament, inert gas filled, prefocused and precentered. The lamp base shall be equipped with a radially protruding pin locator for correct position in the lamp housing.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

The specula shall be designed to permit snug fit to the otoscope head to allow easy attachment to or removal from the otoscope. The aural (3, 4 and 5 mm dia.) and nasal (9 mm dia.) specula shall be the reuseable type. They shall not become distorted, crack nor discolorated when subjected to boiling and alcohol immersion. The distal ends of the specula shall have rounded edges and free of sharpness. The interior surfaces of the specula shall have a non-reflecting finish to prevent glare to the operator.

The magnifier shall be attached to the otoscope head by means of an journal stud made of austenitic stainless steel or chrome plated brass. The magnifier shall be spring-loaded to hold its adjustment and shall be capable of being moved out of the line of vision to permit oblique observation, observation without magnification and insertion of surgical instruments. The magnifying lens shall furnish a magnified view of at least two diameters.

The carrying case shall be fabricated from either hard or soft material with hinge on one side. Suitable means of closure by zipper or latch shall be provided. There shall be a form fitting molded insert in the bottom of the case to accommodate all the components specified in this set.

Finish. The otoscope head and the battery handle shall be chrome plated to mirror finish. Battery handle on the grasping part shall be knurled. The ophthalmoscope body shall be non-gloss or semi-gloss except for the connector, which shall have metallic finish.

Performance. The axial illumination at a point located at the distal end of the large aural speculum for the otoscope shall be not less than 210 lumens per square foot when lamp is operated at not more than 2.70 volts. The axial illumination provided by the ophthalmoscope head, at a point 7-3/4 inches from the upper edge of the mirror, shall be not less than 10 lumens per square foot when the lamp is operated at no more than 2.70 volts. The ophthalmoscope lamp shall be rated 2.5 volts at a maximum of 0.66 ampere and minimum life of 9 hours.

Identification marking. The ophthalmoscope head, otoscope head, and the battery handle shall be permanently marked with the name or registered trademark of the original manufacturer and country of origin. Markings on the carrying case shall be the name or registered trademark, as a minimum.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Workmanship. The otoscope and ophthalmoscope set shall be free from defects which detract from their appearance or impair its serviceability.

Unit. Set (SE). One otoscope and ophthalmoscope set with carrying case, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with AQL of 1.0 percent defective and an inspection level of S-2. The unit of product for examination purposes shall be one set as specified in this description.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirement of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use the covered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-00-550-7199	OTOSCOPE AND OPHTHALMOSCOPE SET, BATTERY TYPE

Ordering data. (Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

ARMY - MD
NAVY - MS
AIR FORCE - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC: 46001

Project No. 6515-4488

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-54043	OTOSCOPE AND OPHTHALMOSCOPE SET, BATTERY TYPE

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*:

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 WASHINGTON D. C.
POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC- RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.