A-A-54062
7 February 1990

COMMERCIAL ITEM DESCRIPTION

FORCEPS, ANEURYSM CLIP APPLYING,
BAYONET, 8-1/4 INCH (Yasargil)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers bayonet-style forceps designed to apply Yasargil aneurysm clips which are either mini size (Type I forceps) or standard size (Type II forceps).

Salient characteristics:

General. Both types I and II forceps shall be compatible with the clips intended for temporary and permanent occlusion (See Note 2). Both are 8-1/4 inches long and have a box-lock construction with a back-action spring on each handle and a spring loaded catch.

Material. The forceps shall be Class 4 martensitic stainless steel and properly heat treated for the steel type used in accordance with ASTM Designation F899 (See Note 3). The jaws shall be annealed cobalt, chrome, molybdenum, nickel alloy in accordance with International Standards Organization (ISO) 5832 or ASTM Designation F562 (See Note 4).

Style, design and dimensions. The forceps shall be in accordance with Defense Personnel Support Center (DPSC) Drawing No. 24474 for style, design and dimensions.

Finish. The finish of the forceps shall be satin, matte, black oxide, or equivalent. The matte finish shall exhibit a uniform nondirectional surface texture equivalent to that produced by lightly blasting the prepared surface with fine abrasives under pressure.

Passivation. Satin or matte finished portions of the forceps shall be properly passivated.

AMSC N/A FSC 6515
DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Performance. When closed and locked, the forceps shall hold the appropriate-sized clip with its tips together. When opened, the forceps shall release the clip.

Identification marking. Each clip-applying forceps shall be legibly marked with the contractor's name or trademark, part number, and size and type of clip it is intended to apply. In addition, the country of manufacture shall also be stated.

Workmanship. The forceps shall be free from defects which detract from its appearance or impair its serviceability.

Visual and dimensional examination. Forceps shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements.

Unit. Each (EA). One forceps, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for Inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for examination purposes shall be one aneurysm clip-applying forceps. Sampling shall be Inspection level II with an acceptable quality level (AQL) of 1.0.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for test purposes shall be one aneurysm clip applying forceps. Sampling shall be Inspection Level S-2 with an AQL of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE 1: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION	TYPE
6515-01-165-1158	Forceps, Aneurysm Clip Applying, Standard Size, Bayonet, 8-1/4 Inch, Yasargil	II
6515-01-242-2228	Forceps, Aneurysm Clip Applying, Mini Size, Bayonet, 8-1/4 Inch, Yasargil	I

NOTE 2: The mini size clip NSN's compatible with the Type I forceps, are:

6515-01-234-8137
6515-01-288-4414
6515-01-288-4415
6515-01-288-3492
6515-01-288-8138 for permanent occlusion, and

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

6515-01-234-8140
6515-01-234-8139
6515-01-234-8129 for temporary occlusion

6515-01-234-8130
6515-01-234-8131
6515-01-234-8133
6515-01-234-8132
6515-01-288-5440
6515-01-234-8136
6515-01-234-8135 for permanent occlusion, and

6515-01-288-3493
6515-01-288-3494
6515-01-288-3495
6515-01-288-5312 for temporary occlusion

NOTE 3: For the forceps components fabricated from Class 4 steel, an example of a steel type to be used is 410x. When properly heat treated, its finished Rockwell Hardness will be C43 to C48, and it will pass the corrosion testing stated in ASTM Designation F1089 without the need for verification testing.

NOTE 4: Because the jaw material is inherently corrosion resistant, no such testing is needed. The annealed material has a minimum hardness of Rockwell C43 or Vickers H400 which does not require verification testing.

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4715

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 Alexandria, VA

POSTAGE WILL BE PAID BY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTS
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A 54062	FORCEPS, ANEURYSM CLIP APPLYING, BAYONET, 8-1/4 INCH (Yasargil)

3a. NAME OF SUBMITTING ORGANIZATION

b. ADDRESS *(Street, City, State, ZIP Code)*

4. TYPE OF ORGANIZATION *(Mark one)*

- ☐ VENDOR
- ☐ USER
- ☐ MANUFACTURER
- ☐ OTHER *(Specify)*: _____

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

Source: https://assist.dla.mil — Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.