A-A-54070
15 October 1990

COMMERCIAL ITEM DESCRIPTION

MIXING TIPS, DENTAL IMPRESSION, SYRINGE, Disposable, 48s

The General Services Administration has authorized the use of this Commercial Item Description.

This commercial item description covers disposable dental impression syringe mixing tips suitable for use in mixing and extruding vinyl polysiloxane impression material.

Salient characteristics:

The mixing tips shall be capable of automatic mixing and extruding syringeable vinyl polysiloxane material directly into a dental tray or intraoral tips.

Material. The tips shall be fabricated of a natural colored polypropylene.

Style, design and dimensions. The tips shall be in accordance with Figure 1, for style and design and be suitable for use with IMPRESSION KIT, DENTAL, Vinyl Polysiloxane, Hydrophilic, Automatic Mixing. (See notes). The overall length of the tip shall be 2.98 ± 0.010 inches. The outside diameter shall be .28 ± 0.010 inch except for the extreme ends, the nozzle head and delivery tip. The tip shall consist of two parts, a 12-stage mixer and a nozzle. The 12-stage mixer shall be placed into the nozzle so that the flat blade is fully seated into the two grooves in the nozzle head. The tip shall attach to the syringe by means of a twist lock with two retainers. The mixing tip shall maintain, throughout the dental procedure, a proper fit to the syringe. The tips shall be disposable. The tips shall exhibit non-sticking qualities and shall not discolor or contaminate the impression material. The mixing tip shall weigh less than 2 grams. A maximum of 2 ml of impression material shall be left in the mixing tip as waste at the end of use.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Workmanship. The mixing tip shall be clean and free of nicks, cracks, specks, deformation, warpage, blockage, or any other defect which detracts from its appearance or impair its serviceability.

Unit. Package (PG). One package containing forty-eight (48) mixing tips, as specified, constitutes one unit.

Unit package. The mixing tips shall be packaged in a durable plastic bag.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Certificates of quality. Certificates of quality for raw material, supplied by the manufacturer of the material, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection, examination and test characteristics. The unit of product for inspection, examination and test purposes shall be one mixing tip. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, except for dimensions which shall be conducted at inspection level S-1, and sampling for tests (when applicable shall be conducted at inspection level S-2).

Examination. Examination shall include, but shall not be limited to the following:

Defects

Mixing tips not smooth.
Mixing tips not natural color.
Mixing tips do not have or the two retainers are deformed.
Mixing tip (1) weighs more than 2 grams.
Mixer not 12-stage.

Tests.

Chemical analysis. In the absence of certificates of quality, the chemical analysis of the plastic shall be determined in accordance with method 111 or 112 of FED-STD-151. In the event of a conflict, method 111 shall govern.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pounds units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and of regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE 1: The following National Stock Number is covered by this document:

National Stock Number	Item Identification
6520-01-217-4342	MIXING TIPS, DENTAL IMPRESSION, SYRINGE, Disposable, 48s

NOTE 2: The following National Stock Number is referenced by this document:

National Stock Number	Item Identification
6520-01-217-4339	IMPRESSION KIT, DENTAL

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

CIVIL AGENCY COORDINATING ACTIVITIES:
VA-OSS
PHS
FDA-MPQAS

58924

PREPARING ACTIVITY:

DoD-MB

AGENT:

DLA-DM

PROJECT NO. 6520-2337
LOCATION: NCR - DISK 8

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-54070	MIXING TIPS, DENTAL IMPRESSION, SYRINGE, Disposable, 48s

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*: _____

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC- RST
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.