A-A-54071
23 February 1990

COMMERCIAL ITEM DESCRIPTION

GAG, MOUTH, DENTAL, ADULT

The General Services Administration has authorized the use of this Commercial Item Description.

This commercial item description covers a dental adult Molt mouth gag for use in holding or increasing the mouth opening during dental procedures.

Salient characteristics:

Material. The mouth gag shall be fabricated from class 4 martensitic corrosion resistant steel in accordance with ASTM F899. The blade cushions shall be fabricated from natural amber rubber.

Hardness. The Rockwell hardness of the instrument shall be 46 to 50 C, when tested in accordance with ASTM E18 or E92 and E140.

Style, design and dimensions. Shall be an adult mouth gag of the Molt design. The gag shall be capable of successive openings at each ratchet catch without unlocking. The operation of the mouth gag shall be easy, without drag or binding. The style, design and dimensions of the mouth gag shall be in accordance with Defense Personnel Support Center (DPSC) Drawing No. 24564. Finger ring inside surfaces shall be well rounded and polished. The gag shall be supplied with one set of removable blade cushions. The cushions shall cover the operative ends of the mouth gag.

Finish. The finish of the Molt mouth gag shall be satin, or matte. The matte finish shall exhibit a uniform nondirectional surface texture equivalent to that produced by lightly blasting the prepared surface with fine abrasives under pressure.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Passivation. The mouth gag shall be properly passivated.

Corrosion resistance. The mouth gag shall show no signs of corrosion or plating of copper when tested in accordance with ASTM F1089.

Performance.

Operation of ratchet catch mechanism. When tested as specified herein, the ratchet catch shall engage and disengage the ratchet teeth smoothly. There shall be no binding or hesitation.

Engagement of ratchet teeth by ratchet catch mechanism. When tested as specified herein, the mouth gag shall open and not unlock at each place checked by closing. There shall be no hesitation or binding. The ratchet teeth shall fully engage at each position.

Operation of retractor without locking. Action shall be smooth and easy when tested as specified herein. The gag shall not bind in any position.

Identification markings. Mouth gags shall be permanently marked with the contractors name or registered trademark and the word "stainless". Mouth gags of foreign manufacture shall bear the name of the country of manufacture.

Workmanship. The mouth gag shall be free from defects which detract from its appearance or impair its serviceability.

Unit package. The mouth gag shall be packaged in a container that prevents damage during shipment or storage.

Unit. Each (EA). One gag, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Certificates of quality. Certificates of quality for raw material, supplied by the manufacturer of the steel, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and an Inspection level of S-2 with an AQL of 1.0 (percent defective). The unit of product for test purposes shall be one mouth gag.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and an Inspection level of II with an AQL of 2.5 (percent defective except for dimensions, which shall be conducted at inspection level S-1 with an AQL of 1.0 (percent defective). Unit of product for examination purposes shall be one mouth gag. Examination shall include but, shall not be limited to the following defects listed below.

Defects.

Instrument not free from cracks, pits, burs, sharp edges, scale, rough areas or corrosion.
Finish of the instrument not as specified.
Style and design not as specified.
Identification markings not present, not complete, not correct, or not legible.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Tests.

Chemical analysis. In the absence of certificates of quality for the raw material, sufficient material from each lot of metal used in the construction of the gag shall be analyzed in accordance with methods 111 or 112 of Federal Test Method Standard No. 151.

Performance tests.

Operation of ratchet catch mechanism. Without operating the mouth gag, fully disengage the ratchet catch mechanism from the ratchet teeth so the spring is displaced. Repeat 5 times.

Engagement of ratchet teeth by ratchet catch mechanism. Starting with the prongs fully closed, open the mouth gag completely. Stop at 4 or 5 places and apply force onto the blades and attempt to close the mouth gag. Note the degree of engagement between ratchet teeth and ratchet catch mechanism.

Operation of mouth gag without locking. Disengage catch mechanism from the ratchet teeth. Starting with the mouth gag arms completely open, close the mouth gag arms until the jaws are together (closed). Return to the open position and repeat three times. Note ease of action.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerance specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

National Stock Number	Item Identification
6520-01-260-1228	GAG, MOUTH, DENTAL, ADULT

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

A-A-54071

CIVIL AGENCY COORDINATING ACTIVITIES:
VA-OSS
PHS
FDA-MPQAS

PROJECT NO. 6520-2046

LOCATION: NCR DISK # 5

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-RSTE
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-54071	Gag, Mouth, Dental, Adult

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last First MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City State ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil/ Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.