A-A-54072

22 January 1990

# COMMERCIAL ITEM DESCRIPTION

## LIGHT, SLIT, OPHTHALMIC (WITH TONOMETER)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a slit light consisting of a binocular microscope, a slit illumination system and accessories required to perform detailed examination of the tissue and media of the living human eye. The mechanical design shall provide for rapid, smooth operation of the binocular microscope and the slit illumination system, either separately or simultaneously, during the examination.

Salient characteristics:

The slit lamp shall consist of a binocular microscope, a slit illumination system and the following accessories: one fixation device, one Goldman applanation tonometer, two objective lenses: 1X and 1.6X, 100 chin rest tissues, and one instrument table top.

Binocular Microscope.

The binocular microscope shall be of the horizontal, stereoscopic, straight tube type. Shall maintain a parfocal image when changing powers. Shall have a magnification changing system with at least three viewing magnifications of 10X, 16X and 25X. These magnifications may be achieved through 2 eyepieces of 10X and 16X or an appropriate combination of eyepieces to achieve the proper viewing magnifications.

The binocular microscope and slit illumination system shall operate separately or simultaneously during an examination.

The Eyepiece focus-adjustment shall have a minimum range of ± 5 Diopters.

The interpupillary distance adjustment shall be continuously variable and include the range from 55 mm to 75 mm.

AMSC N/A FSC 6540

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:28Z
Check the source to verify that this is the current version before use.

Slit Illumination System.

The slit width shall be continuously variable from 0.0 mm to 8.0 mm, minimum.

The slit length adjustment shall be continuously variable and include the range from 1.0 mm to 8.0 mm. The slit angle shall be 90 degrees from vertical to horizontal.

Shall have aperture diaphrams with diameters of 8.0 mm, 5.0 mm, 3.0 mm, 2.0 mm, 1.0 mm and 0.2 mm.

The unit shall have at least 4 built-in filters: cobalt blue, red-free, neutral density and heat dissipation.

Shall have a vertically adjustable chin-rest with a minimum adjustable range of 80 mm.

The optics shall be able to clearly resolve the corneal endothelum at 10X power.

Accessories.

The unit shall be supplied with the following accessories.

One fixation device.
100 chin-rest tissues.
One Goldman applanation tonometer.
One instrument table top.
Two objective lenses: 1X and 1.6X.

Spare parts.

Spare illumination bulbs shall be supplied with each unit as follows:

Two fixation bulbs, two main illumination bulbs, and three of each fuse type used in the unit.

Finish.

The unit shall be fabricated from materials which are inherently resistant to corrosion or which are treated to render them corrosion-resistant. When painting or plating is used, appropriate cleaning and priming shall be performed prior to finishing. Unless otherwise specified, finish shall be equivalent or superior to that of comparable commercial equipment, and shall be free of gaps, scratches, chips, and other imperfections.



Source: https://assist.dla.mil -- Downloaded: 202640-26T09:28Z
Check the source to verify that this is the current version before use.

Power requirements.

The slit lamp shall operate a 110 VAC + 10%, 60 HERTZ power supply and use a standard hospital grade plug.

Workmanship. The slit lamp and accessories shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Each (EA). One slit lamp and accessories, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

First article inspection. First article inspection shall be performed on the slit lamp and accessories when a first article sample is required. This inspection shall include examination and tests.

Quality conformance inspection. Quality conformance inspection shall include the examination and tests specified herein.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Certificates of quality. Certificates of quality, supplied by the manufacturer of the materials, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the item identification, the name of the component/material, the lot number and the lot size.

Sampling for tests. Sampling for test shall be conducted in accordance with commercial practice. The unit of product for test purposes shall be one slit lamp.

A. First article. Where a first article sample is required, all tests shall be performed on that unit.

B. Acceptance testing. When a first article sample has not been required, one unit shall be subjected to all tests specified herein. If a first article sample has successfully passed all tests, performance tests shall not be required.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for one or more tests.

Performance tests. The slit lamp shall be tested for compliance with performance requirements as follows:

Operational test. The slit lamp shall be connected to a suitable power source, in accordance with manufacturer's instructions, and shall be operated to demonstrate proper functioning of the mechanical, electrical and optical components.

The binocular microscope shall be operated to demonstrate that it will clearly resolve the corneal endothelum at the 10X magnification power and will maintain a parfocal image when changing powers (stay in focus after changing powers without additional adjustment).

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Regulation Subpart 23.4, to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6540-01-032-4518

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

AGENT:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6540-1221

Location: Disk: CID/LIGHT/3C/3a17



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.