A-A-54075
31 January 1990

COMMERCIAL ITEM DESCRIPTION

VITRECTOMY KIT, OPHTHALMOLOGY

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers an individually packaged, sealed, sterile and disposable vitrectomy kit intended for single use in the field for performing vitrectomy procedures.

Salient characteristics.

The kit shall consist of a cutting head assembly with attached tubing for irrigation and aspiration, and a suction syringe used for removal of the vitreous. Shall be provided with an I.V. spike with tubing attached that will connect to the irrigation tubing. The kit shall be provided with at least two (2) latex sheaths and a plastic funnel for sterile delivery of the reusable power unit. The reusable power unit shall not be included with the kit itself. The kit shall be adaptable and include all necessary tubing adapters for connection to a footswitch-controlled irrigation and aspiration unit.

The vitrectomy kit shall have a set up time of no more than one minute as demonstrated by an experienced user. Each kit shall be individually packaged, sterilized and sealed in a rigid but light weight container or tray which may be easily opened without any special instruction or equipment. Each kit shall have marked the sterilization date and lot number on the front of the package.

The entire kit shall not weigh more than one (1) pound and shall include one set of assembly instructions.

All Luer fittings mentioned in this specification shall conform to the performance requirements as specified in ANSI MD70.1.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

CUTTING HEAD ASSEMBLY.
The cutting head assembly shall use a guillotine cutting head located within the aspiration probe and designed for a single use. Shall have attached plastic aspiration tubing that will connect to a 30 cc, plastic, translucent suction syringe by use of a low pressure Luer lock fitting. The female end of the Luer lock shall be attached to the tubing and the male end shall be attached to the syringe. The aspiration probe shall be made of 19 gauge stainless steel and shall have an aspiration port not greater than 0.45 mm and a nominal probe length of 25 mm.

IRRIGATION TIP ASSEMBLY AND I.V. SPIKE.
The irrigation probe shall be constructed of 16 gauge stainless steel tubing. Shall be designed to fit over the aspiration probe and lock in place. The probe shall have attached plastic tubing. A plastic, translucent, vented I.V. spike with attached tubing and roller clamp shall be supplied in each kit. The irrigation probe tubing and the I.V. spike tubing shall connect by means of a semi-rigid Luer slip fitting. The female end of the Luer fitting shall be attached to the irrigation tubing and the male end to the I.V. spike tubing.

Workmanship. The vitrectomy kit shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing five individually sealed vitrectomy kits, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one vitrectomy kit.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/ contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE. The following National Stock Number is covered by this document:

6515-01-218-4427

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4673

Location: CID/VITRECTOMY/S3C/17

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.