A-A-54076
29 January 1990

COMMERCIAL ITEM DESCRIPTION

HYDROMETER, GRADUATED SCALE API GRAVITY

The General Services Administration has authorized the use of this commercial item description in preference to Military Specification MIL-H-36266A (DSA-DM) which is cancelled.

This commercial item description covers American Petroleum Institute (API) gravity glass hydrometers.

Salient characteristics.

Shall be API gravity hydrometers for petroleum products and other liquids of similar surface tensions (i.e. 33 dynes/centimeter or less).

Shall comply with all requirements in ASTM Designation E100, as applicable to ASTM hydrometer numbers for API gravity hydrometers (see "Ordering Data").

Workmanship. The hydrometers shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Each (EA). One hydrometer, of the ASTM number specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

AMSC N/A FSC 6630

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspections and tests. Sampling for inspections and tests shall be conducted in accordance with commercial practice.

Inspections and tests. Inspections and tests shall be conducted in accordance with ASTM Designations E100 and E126 to determine compliance with specification requirements.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/ contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Other publications. The following documents form a part of this specification to the extent that is specified herein. Unless otherwise indicated, the effect on date of invitation for bids or request for proposals shall apply.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM)

ASTM E100 STANDARD SPECIFICATION FOR ASTM HYDROMETERS

ASTM E126 STANDARD METHOD FOR INSPECTION AND VERIFICATION OF HYDROMETERS

(Application for copies should be addressed to American Society for Testing and Materials, 1916 Race Street, Philadelphia, PA 19103).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Number is covered by this document:

National Stock Number	Item Identification	ASTM No.
6630-00-244-9980	HYDROMETER, GRADUATED SCALE 69 to 81° A.P.I. Range	8H-62
6630-00-244-9981	HYDROMETER, GRADUATED SCALE 79 to 91° A.P.I. Range	9H-62
6630-00-244-9982	HYDROMETER, GRADUATED SCALE 49 to 61° A.P.I. Range	6H-62
6630-00-244-9983	HYDROMETER, GRADUATED SCALE 59 to 71° A.P.I. Range	7H-62
6630-00-244-9984	HYDROMETER, GRADUATED SCALE 29 to 41° A.P.I. Range	4H-62
6630-00-244-9985	HYDROMETER, GRADUATED SCALE 39 to 51° A.P.I. Range	5H-62
6630-00-244-9986	HYDROMETER, GRADUATED SCALE 9 to 21° A.P.I. Range	2H-62
6630-00-244-9987	HYDROMETER, GRADUATED SCALE 19 to 31° A.P.I. Range	3H-62
6630-00-252-2252	HYDROMETER, GRADUATED SCALE 65 to 71° A.P.I. Range	34H-62
6630-00-252-2253	HYDROMETER, GRADUATED SCALE 70 to 76° A.P.I. Range	35H-62
6630-00-252-2254	HYDROMETER, GRADUATED SCALE 55 to 61° A.P.I. Range	32H-62
6630-00-252-2255	HYDROMETER, GRADUATED SCALE 60 to 66° A.P.I. Range	33H-62
6630-00-252-2256	HYDROMETER, GRADUATED SCALE 45 to 51° A.P.I. Range	30H-62
6630-00-252-2257	HYDROMETER, GRADUATED SCALE 50 to 56° A.P.I. Range	31H-62
6630-00-252-2258	HYDROMETER, GRADUATED SCALE 25 to 31° A.P.I. Range	26H-62

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

National Stock Number	Item Identification	ASTM No.
6630-00-252-2259	HYDROMETER, GRADUATED SCALE 30 to 36° A.P.I. Range	27H-62
6630-00-252-2260	HYDROMETER, GRADUATED SCALE 15 to 21° A.P.I. Range	24H-62
6630-00-252-2261	HYDROMETER, GRADUATED SCALE 20 to 26° A.P.I. Range	25H-62
6630-00-252-2262	HYDROMETER, GRADUATED SCALE 5 to 11° A.P.I. Range	22H-62
6630-00-252-2263	HYDROMETER, GRADUATED SCALE 10 to 16° A.P.I. Range	23H-62
6630-00-252-6389	HYDROMETER, GRADUATED SCALE 35 to 41° A.P.I. Range	28H-62
6630-00-252-6390	HYDROMETER, GRADUATED SCALE 40 to 46° A.P.I. Range	29H-62
6630-00-286-4871	HYDROMETER, GRADUATED SCALE Minus 1 to Plus 11° A.P.I. Range	1H-62

Ordering data. (ASTM No(s) required intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6630-0493

Location: CID/HYDRO/GRAD/S31/17

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.