DD-B-00582(DSA-DM)
30 November 1964

INTERIM FEDERAL SPECIFICATION

BOTTLE, EYE IRRIGATING, ZIEGLER

This interim Federal specification was developed by the Defense Medical Supply Center, 3rd Avenue and 29th Street, Brooklyn, New York 11232, based upon currently available technical information. It is recommended that Federal agencies use it in procurement and forward recommendations for changes to the preparing activity at the address shown above.

1. SCOPE

1.1 This specification covers one type (Ziegler) of glass irrigating bottle, suitable for flushing the conjunctival sac of the eye.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of the specification to the extent specified herein:

Federal Specifications:

PPP-B-566	Boxex, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Boxes, Set-Up, Paperboard.

Federal Standard:

Fed. Test Method Std. No. 601 Rubber: Sampling and Testing.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:43Z
Check the source to verify that this is the current version before use.

(Activities outside the Federal Government may obtain copies of Federal Specifications and standards as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U. S. Government Printing Office, Washington 25, D. C.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Atlanta, Chicago, Kansas City, Mo., Dallas, Denver, San Francisco, Los Angeles, Seattle, and Washington, D. C.

(Federal Government activities may obtain copies of Federal Specifications and Standards and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

Military Specification:

MIL-L-10547 Liners. Case and Sheet. Overwrap; Water-Vaporproof or Waterproof, Flexible.

Military Standards:

MIL-STD-105 Sampling Procedures and Tables for Inspection by Attributes.

MIL-STD-129 Marking for Shipment and Storage.

Defense Medical Supply Center Drawing:

20528 Bottle, Eye Irrigating, Ziegler, 8 oz.

(Copies of specifications, standards, and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material. The material used shall be as indicated herein and of the highest quality, suitable in all respects for the purpose intended.

3.1.1 Glass. The irrigating bottle shall be fabricated of soda lime glass.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:43Z
Check the source to verify that this is the current version before use.

3.1.2 Rubber. The bulb shall be fabricated of a suitable rubber compound (see 3.3).

3.2 Dimensions, style, and design. The glass irrigating bottle shall be of the Ziegler type, and shall have an 8 ounce capacity, with ground glass cover, rubber bulb, and glass irrigator tip. The bottle shall be especially designed for flushing of the conjunctival sac of the eye, or for other irrigating purposes. Dimensions, style, and design shall be in accordance with the drawing.

3.3 Accelerated aging. The bulb, when subjected to the accelerated aging test specified in 4.4.1, shall not become brittle, sticky, or tacky.

3.4 Workmanship. Workmanship shall be first class throughout. Bottles shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, auditory, etc., investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling. Sampling for examination shall be conducted in accordance with MIL-STD-105.

4.3 Examination. The bottles shall be examined to determine compliance with all requirements contained in this specification.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:43Z
Check the source to verify that this is the current version before use.

4.4.1 Accelerated aging test. Bulbs shall be subjected to the accelerated aging test described in method 7221 of No. 601.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One (1) bottle, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in table I shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in procurement document.

Table I. Packaging quantities

Unit package	Intermediate package	Exterior container
1 unit	12 units	36 units

5.1.2.1 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit package. Each unit shall be packaged in a carton of appropriate size which shall be designed to insure against shifting of, or damage to, the contents. Carton shall be adequately secured to prevent accidental opening.

5.1.3.2 Intermediate package. Intermediate package shall be a carton of appropriate size constructed in accordance with PPP-B-566 or PPP-B-676, except that commercial colors shall be acceptable, or PPP-B-636. Closure shall be adequate to prevent spillage of contents under normal handling.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 3 use; PPP-B-601, table II, using type III, class 1 plywood; PPP-B-621, table III; or PPP-B-636, type CF, class Weather-resistant.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:43Z
Check the source to verify that this is the current version before use.

4.2 Examination. Each production unit shall be examined for compliance with the requirements of section 3 of this specification. This element of inspection shall cover all visual examinations and dimensional measurements. Noncompliance with any specified requirement or presence of one or more defects preventing or lessening maximum efficiency shall constitute cause for rejection.

4.3 Radiography. After completion of the component parts of the switch, including repair of surface defects, heat treatment or stress-relieving, but prior to assembly, the switch shall be subject to radiographic inspection. A total of twelve radiographs shall be made for each switch assembly. The locations to be radiographed shall be selected to provide information on the critical areas of each of the components, and shall be submitted for approval prior to radiography. Radiographic testing shall conform to ASTM Specification E142. Radiographs shall be compared to class 5 for discontinuity types A, B, and C, of ASTM Specification E71. Defects classified "Unacceptable" shall be repaired in accordance with paragraph 3.4.5, or the casting shall be rejected. After repair of "Unacceptable" defects, additional radiographs shall be made of the components as directed, to insure the structural integrity and service life of the component. If the radiographs reveal defects which are classified as "Borderline", additional radiographs shall be made of the component as directed. When the extent or location of the "Borderline" defects is judged by the contracting officer to affect the structural integrity or service life of the casting, the defects shall be repaired in accordance with 3.4.5, or the casting shall be rejected.

5. PREPARATION FOR DELIVERY

5.1 Preservation, packaging and packing. The complete switches shall be preserved, packaged, and packed in a manner which will insure arrival at destination in satisfactory condition and which will be acceptable to the carrier at lowest rates. Packing shall comply with OCC Uniform Freight Classification Rules or ATA National Motor Freight Classification Rules.

5.2 Marking. The switches shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Intended use. These crane track switches are the standard design of the Bureau of Yards and Docks for use on turnouts for portal crane trackage and are in accordance with Bureau criteria for portal crane track alinement. They are for use in new construction and, where conditions permit in rehabilitation of existing trackage.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:43Z
Check the source to verify that this is the current version before use.

6.2 Ordering data. Procurement documents should specify the following information:

(a) Title, number, and date of this specification.
(b) Type of switch (see 1.2).

Custodian:

Navy - YD

Preparing activity:

Navy - YD

Project No. 2250-N0001Y

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:43Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

FORM APPROVED BUDGET BUREAU NO. 119-R004

INSTRUCTIONS

This sheet is to be filled out by personnel either Government or contractor, involved in the use of the specification in procurement of products for ultimate use of the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity.

SPECIFICATION: INTERIM FEDERAL SPECIFICATION BOTTLE, EYE IRRIGATING, ZIEGLER DD-B-00582(DSA-DM) dated, 30 November 1964.

ORGANIZATION

CITY AND STATE

CONTRACT NO.

QUANTITY OF ITEMS PROCURED

DOLLAR AMOUNT
$

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO IF "YES" IN WHAT WAY?

4. REMARKS (ATTACH ANY PERTINENT DATA WHICH MAY BE OF USE IN IMPROVING THIS SPECIFICATION. IF THERE ARE ADDITIONAL PAPERS, ATTACH TO FORM AND PLACE BOTH IN AN ENVELOPE ADDRESSED TO PREPARING ACTIVITY)

SUBMITTED BY (PRINTED OR TYPED NAME AND ACTIVITY)

DATE

DD FORM 1426

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:43Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

Fold Under

DEFENSE SUPPLY AGENCY
DEFENSE MEDICAL SUPPLY CENTER
3RD AVENUE AND 29TH STREET
BROOKLYN, N.Y., 11232

OFFICIAL BUSINESS

POSTAGE FEES PAID
DEFENSE SUPPLY AGENCY

Defense Supply Agency
Defense Medical Supply Center
ATTN: Directorate of Technical Operations (Code T-MS)
3rd Avenue and 29th Street
Brooklyn, N.Y., 11232

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:43Z
Check the source to verify that this is the current version before use.