DD-B-597A
May 25, 1976
SUPERSEDING
DD-B-597
September 22, 1959

# FEDERAL SPECIFICATION

## BOTTLE, STOPPER, NARROW MOUTH, ROUND

This specification was approved by the Commissioner, Federal Supply Service, General Services Administration, for the use of all Federal agencies.

## 1. SCOPE

1.1 This specification covers 5-gallon capacity, carboy type, narrow mouth bottles without stoppers.

## 2. APPLICABLE DOCUMENTS

2.1 The following documents, of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein:

Federal Specifications:

DD-G-541 - Glass (Laboratory).
PPP-B-636 - Boxes, Shipping, Fiberboard.

(Activities outside the Federal Government may obtain copies of Federal Specifications, Standards, and Handbooks as outlined under the General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20402.

(Single copies of this specification and other Federal Specification required by activities outside the Federal Government for bidding purposes are available without charge from Business Service Centers at the General Services Administration Regional Offices in Boston, New York, Washington, DC, Atlanta, Chicago, Kansas City, MO, Fort Worth, Denver, San Francisco, Los Angeles, and Seattle, WA.

(Federal Government activities may obtain copies of Federal Specifications, Standards, and Handbooks and the Index of Federal Specifications and Standards from established distribution points in their agencies.)



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:43Z
Check the source to verify that this is the current version before use.

Military Specifications:

MIL-P-116 - Preservation - packaging, Methods of.

Military Standards:

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.

(Copies of Military Specifications, Standards and Drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM)

ASTM C-148 Polariscopic Examination of Glass Containers.

(Application for copies should be addressed to the American Society for Testing and Materials, 1916 Race Street, Philadelphia, PA 19103.)

3. REQUIREMENTS

3.1 Material. Bottle shall be fabricated of glass. Glass shall be in accordance with DD-G-541, type I, class A.

3.2 Construction. Design and construction shall be in accordance with figure 1. Neck shall be tool-finished and shall accept a number 12 rubber stopper.

3.3 Strain. Ring strain and longitudinal strain shall not exceed temper no. 1 when tested as specified in 4.4.1.

3.4 Identification marking. Each bottle shall be permanently marked with the name or registered trademark of the manufacturer.



Source: https://assist.dla.mil -- Downloaded: 202610-26T11:43Z
Check the source to verify that this is the current version before use.





Figure 1 - Bottle, Stopper.

DIMENSIONS ARE IN INCHES



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:43Z
Check the source to verify that this is the current version before use.

3.5 Workmanship. Workmanship shall be first class throughout. Bottles shall be free of defects listed in table III and any other defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Certificate of quality. Certificate of quality, supplied by the manufacturer of the glass, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the item identification, the name of the material, the lot number, the date of testing, the test method, individual test results and the specification requirements.

4.1.3 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and table I. Unit of product shall be one bottle.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:43Z
Check the source to verify that this is the current version before use.

TABLE I. Sampling for examination

| | Inspection level | AQL (percent defective) |
|---|---|---|
| For visual examination | | |
| Major defects | II | 2.5 |
| Minor defects | II | 4.0 |
| For dimensional examination | S-1 | 4.0 |

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and table II. Unit of product shall be one bottle and acceptance number shall be zero.

TABLE II. Sampling for tests

| Characteristic | Requirement | Test procedure | Inspection level |
|---|---|---|---|
| Strain | 3.3 | 4.4.1 | S-2 |

4.3 Examination. The bottles shall be examined to determine compliance with all requirements contained in this specification. Examination shall be in accordance with, but shall not be limited to, the defects listed in table III.

TABLE III. Classification of defects

| Category | Defect |
|---|---|
| Major | |
| 101 | Bottle not stable on flat surfaces. |
| 102 | Bottle not free of cracks and chips. |
| 103 | Bottle not free of sharp and rough edges. |
| Minor | |
| 201 | Bottle not free of clusters of bubbles and stones. |
| 202 | Bottle not clean and dry. |
| 203 | Bottle not free of etched and crazed areas. |
| 204 | Bottle not free of deformation. |
| 205 | Identification marking not complete, correct, legible and permanent. |

4.3.1 Dimensional examination. Bottle shall be examined for defects in dimensions. Any dimension not within the tolerance specified herein shall be classified as a defect.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:43Z
Check the source to verify that this is the current version before use.

4.3.2 Examination of preparation for delivery. An examination shall be made to determine compliance with the packaging, packing and marking requirements of section 5. Defects shall be scored as specified in table IV. Sampling shall be in accordance with MIL-STD-105. The sample unit shall be one container fully prepared for delivery. The lot shall be the number of containers offered for inspection at one time. The inspection level shall be S-2 with AQL of 4.0 expressed in terms of percent defective.

TABLE IV Examination of preparation of delivery

| Examine | Defect |
|---|---|
| Contents | Not as specified. |
| Container | Not as specified. |
| Marking | Omitted; incorrect; illegible; improper size, location, sequence or method of application. |
| Material | Component missing or damaged. |
| Workmanship | Bulging or distortion of containers. |

4.3.2.1 Examination of closure, waterproofing and banding of container. When the shipping container is required to comply with PPP-B-636, the examination for defects in closure, waterproofing or banding shall be in accordance with the appendix of the box specification.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Strain. Strain shall be determined in accordance with ASTM C148, Method A or B.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A, B or C, as specified (see 6.2).

5.1.1 Level A.

5.1.1.1 Unit packaging. Each bottle shall be packaged in a box conforming to PPP-B-636, class weather-resistant, variety DW. The box shall be designed with suitable interior fittings to prevent damage to the contents. Closure, strapping and waterproofing shall be as specified in the appendix of the box specification.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:43Z
Check the source to verify that this is the current version before use.

5.1.2 Level B. Each bottle shall be packaged as specified in 5.1.1, except that the box shall conform to PPP-B-636, class domestic, variety DW. Waterproofing and strapping shall not be required.

5.1.3 Level C. Bottles shall be individually packaged in shipping containers that will afford adequate protection against damage during direct shipment from the supply source to the first receiving activity. These packs shall conform to the applicable carrier's rules and regulations.

5.2 Packing. Bottles packaged as specified in 5.1 shall require no further packing.

5.3 Marking. Each unit package shall be marked with exterior container markings as specified in MIL-STD-129. Marking shall include fragile labels and the legend: "GLASS--HANDLE WITH CARE."

6. NOTES

6.1 Intended use. The bottle covered in this specification is intended to contain distilled water for use in the production of intravenous solution and for use in analytical chemistry.

6.2 Ordering data. Procurement documents should specify the following:

a. Title, number and date of this specification.
b. Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.3 Type I plain round 2-1/2 gallon capacity bottle has been deleted from this revision. Carboy bottle covered by this specification was type II in previous edition.

6.4 This specification does not cover all types, classes, grades or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the type normally procured to meet Federal requirements.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:43Z
Check the source to verify that this is the current version before use.

MILITARY CUSTODIANS:

Army - MD
Navy - MS
Air Force - 03

Civil Agency Coordinating Activities:

GSA - FSS
HEW - NIH
USDA - AFS
VA - DMS

Preparing activity:

DSA - DM

Project No. 6640-0964

Orders for this publication are to be placed with General Services Administration, acting as an agent for the Superintendent of Documents. See section 2 of this specification to obtain extra copies and other documents referenced herein. Price 35 cents each.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:43Z
Check the source to verify that this is the current version before use.

# STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

*INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/ vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

DD-B-597A - Bottle, Stopper, Narrow Mouth, Round

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426

REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

S/N 0102-014-1802

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:43Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, PA 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

U.S. MAIL

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:43Z
Check the source to verify that this is the current version before use.