DD-M-00200b (DSA-DM)
May 13, 1970
Interim Revision of
Fed. Spec. DD-M-200a
August 6, 1958

INTERIM FEDERAL SPECIFICATION

MEDICINE GLASS

This Interim Federal Specification was developed by the Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, Pennsylvania 19101, based upon currently available technical information. It is recommended that Federal agencies use it in procurement and forward recommendations for changes to the preparing activity at the address shown above. The General Services Administration has authorized the use of this Interim Federal Specification as a valid exception to Federal Specification DD-M-200a.

1. SCOPE AND CLASSIFICATION

1.1 Scope. This specification covers two types of medicine glasses suitable for use in hospitals and dental offices.

1.2 Classification. Medicine glasses shall be of the following types as specified (see 6.1):

Type I. Medicine Glass, (Hospital).
Type II. Dish, Medicament (Dental).

2. APPLICABLE DOCUMENTS

2.1 Specifications and standards. The following documents, of the issue in effect on date of invitations for bid or requests for proposal, form a part of this specification to the extent specified herein:

FSC 6520, 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:42Z
Check the source to verify that this is the current version before use.

SPECIFICATIONS

Federal

PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.

STANDARDS

Federal

FED-STD-123	Marking for Domestic Shipment (Civilian Agencies).

(Activities outside the Federal Government may obtain copies of Federal Specifications and Standards as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U.S. Government Printing Office, Washington, D.C., 20402.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston; New York; Atlanta; Chicago; Kansas City, Mo.; Fort Worth; Denver; San Francisco; Auburn, Washington; and Washington, D.C.

(Federal Government activities may obtain copies of Federal Specifications and Standards and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

SPECIFICATION

Military

MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:42Z
Check the source to verify that this is the current version before use.

(Copies of Military specifications, standards, and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material. Types I and II medicine glasses shall be fabricated of colorless, transparent flint glass of a quality that will retain its luster and transparency.

3.2 Style, design, and dimensions.

3.2.1 Type I Medicine Glass, (Hospital). Glasses shall be of the style, design, and dimensions specified in figure 1. Minimum graduations for each glass shall be as shown on figure 1. Additional graduations in ounces and cubic centimeters shall be acceptable. The scale, numbers, and all other markings shall be permanently etched, engraved, or molded on the outside of the glass, and shall be readily observed regardless of the color or character of the liquid in the glass.

3.2.2 Type II Dish, Medicament (Dental). Glasses shall be of the type known in the dental profession as "Deppen" medicine glasses. The glasses shall be hexagonal, octagonal, or decagonal in shape, and shall have a concavity in the upper and lower surfaces. Hexagonal glasses, when furnished, shall be of the style, design, and dimensions specified in figure 1. Octagonal and decagonal glasses, when furnished, shall be of the general design and dimensions specified in figure 1.

3.3 Heat treatment. Medicine glasses shall be thoroughly annealed and shall meet the requirements of 3.3.1 and 3.3.2.

3.3.1 Boiling resistance. The medicine glasses shall show no evidence of corrosion, chipping, scumming, or cracking when tested as specified in 4.4.1.

3.3.2 Thermal shock resistance. Medicine glasses shall not crack when tested as specified in 4.4.2.

3.4 Identification marking. Each glass shall be marked with the name or registered trademark of the manufacturer. Markings shall be permanently etched, engraved, or molded on the outside of the glasses, and shall be readily observed regardless of the color or character of the liquid in the glass. Marking on type II glasses is optional.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:42Z
Check the source to verify that this is the current version before use.

FIG. 1

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:42Z
Check the source to verify that this is the current version before use.

3.5 Workmanship. Workmanship shall be first class throughout. The medicine glasses shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in this specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties).

4.2 Quality conformance inspection. The examinations and tests required to assure conformance of those items or lots of items to be offered are classified as specified in tables I and III (see 4.2.1 and 4.3) and table II (see 4.2.2).

4.2.1 Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels (AQL's) shall be as indicated in table I. Unit of product for examination purposes shall be one medicine glass.

Table I. Acceptable Quality Levels

Type defect	Inspection level	AQL (Percent defective)
Major	II	1.0
Minor A	II	2.5
Minor B	S-1	4.0

4.2.2 Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and table II, except that the acceptance number shall be zero for all sample sizes. Unit of product for test purposes shall be one medicine glass.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:42Z
Check the source to verify that this is the current version before use.

Table II. Sampling Criteria

Characteristic	Requirement	Test procedure	Inspection level
Boiling resistance	3.3.1	4.4.1	S-2
Thermal shock resistance	3.3.2	4.4.2	S-2

4.3 Examination. The medicine glasses shall be examined for defects including, but not limited to, those listed in table III.

Table III. Classification of Defects

Categories	Defects
Major:	
101	Type of medicine glass not as specified.
102	Style and design of medicine glass not as shown in applicable figure.
103	Graduation or identification markings not on outside of glasses.
104	Graduation markings not permanent or omitted altogether.
105	Glass not free from cracks, chips, or sharp edges.
Minor A:	
201	Graduation markings not minimum specified.
202	Graduation markings not clear and legible.
203	Glass not free of striae, seeds, blisters, or other imperfections.
204	Glasses not symmetrically formed.
205	Identification marking not as specified.
206	Packaging and packing not as specified.
207	Marking on packaging and pack not as specified.
Minor B:	
251	Dimensions not as specified.

4.4 Tests.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:42Z
Check the source to verify that this is the current version before use.

4.4.1 Boiling test. The sample medicine glasses shall be placed in a closed vessel having a small steam vent, completely immersed in distilled water, and boiled continuously for 6 hours for type I and 4 hours for type II. Boiling distilled water shall be added, when necessary, to compensate for evaporation losses. The medicine glasses shall be suspended in the water so that they do not come in contact with the containing vessel.

4.4.2 Thermal shock test. The sample medicine glasses shall be immersed in tap water at 18.5 ± 2.5 degrees Centigrade (65 ± 5 degrees Fahrenheit), allowed to remain 10 minutes, then suddenly transferred into boiling water. This operation shall be repeated five times for type I and four times for type II glasses. Samples for this test shall not be any of those subjected to the boiling test specified in 4.4.1.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A, or C, as specified (see 6.1).

5.1.1 Unit of issue. One dish, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in table IV shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table IV. Packaging Quantities

Type	Unit package	Intermediate package	Exterior container
I	Not required	12 units	288 units
II	Not required	12 units	576 units

5.1.2.1 Variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Intermediate package. Intermediate package shall be a box of appropriate size, fabricated of no less than 200 pound test fiberboard and constructed in accordance with PPP-B-636, type CF, class domestic, style optional. Box shall be provided with full height partitions fabricated of the same strength material as that used in the body of the box, so arranged that each glass is securely held in an individual cell.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:42Z
Check the source to verify that this is the current version before use.

5.1.4 Level C. Units shall be packaged in standard commercial containers, of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B, or C, as specified (see 6.1).

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and shall be constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant, grade V3. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.1.2 Case liner. Each Level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and shall be constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers, of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Civil agencies. Marking for shipment shall be in accordance with FED-STD-123.

5.3.2 Military activities.

5.3.2.1 Intermediate package. Each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown. In addition, marking shall include the following legend:

"GLASS - HANDLE WITH CARE"

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:42Z
Check the source to verify that this is the current version before use.

5.3.2.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129. Marking shall include the following legend:

"GLASS - HANDLE WITH CARE"

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Title, number, and date of this specification.

(b) Type required (see 1.2).

(c) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 Specification changes. Because of the extensiveness of changes to this document, the margins have not been annotated with asterisks to indicate where changes from the previous issue were made.

6.3 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet Federal Government requirements.

6.4 This specification covers the following items shown in the Federal Supply Catalog:

FSNS	ITEMS
6520-550-2000	DISH, MEDICAMENT, DENTAL
6530-780-7000	MEDICINE GLASS

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project Number 6520-1234

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

94

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:42Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document (s) or serve to amend contractual requirements.

SPECIFICATION
DD-M-00200b(DSA-DM) FSN:

ORGANIZATION

CITY AND STATE | CONTRACT NUMBER

MATERIAL PROCURED UNDER A
☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING.

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?
   ☐ YES ☐ NO (If "yes", in what way?)

4. REMARKS (Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)

SUBMITTED BY (Printed or typed name and activity - Optional) | DATE

DD FORM 1 JAN 66 1426 REPLACES EDITION OF 1 OCT 56, WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:42Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

Fold

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

Fold

DPSC FORM 2418
MAR 68

EDITION AUG 66, WILL BE USED UNTIL EXHAUSTED

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:42Z
Check the source to verify that this is the current version before use.