MILITARY SPECIFICATION

BULB, TEMPERATURE, THERMORESISTOR, -70° TO +150°C, TYPE G-1

This amendment forms a part of Military Specification MIL-B-7370B, dated 10 September 1969, and is approved for use by all Departments and Agencies of the Department of Defense.

Page 1

2.1: Under STANDARDS, Military add:

"MIL-STD-105 Sampling Procedures And Tables For Inpsection By Attributes"

Page 8

4.5.2: Delete and substitute:

"4.5.2 Sampling tests. Sampling shall be performed in accordance with MIL-STD-105, Inpsection Level II, AQL 2.5 percent."

4.5.3: Delete and substitute:

"4.5.3 Special tests. Unless otherwise stated (6.4), three bulbs selected at random from the first 15 bulbs on each contract or order shall be subjected to all of the following tests. Failure of any test by any bulb shall be cause for rejection. Any failure shall be reported to the procuring activity, and no bulbs shall be delivered or accepted until the procuring activity is satisfied that all bulbs meet all requirements.

Page 9

4.5.4: Delete in its entirety.

Page 11

6.4: Delete and substitute:

"6.4 Ordering data. Procurement documents should specify:

a. Levels of packaging and packing required (see Section 5)

b. Conditions under which the special test may be waived (see 4.5.3).

These tests may be waived to those manufacturers who have previously complied with these requirements on a government contract within the past 12 months. Certification of compliance will be required, noting date and contract number."

Custodians:
Air Force - 82
Navy - AS

Preparing Activity:
Air Force - 82

Project No. 6685-0519

Review Activities:
Army - EL, AV
Navy - AS

FSC 6685

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T08:26Z
Check the source to verify that this is the current version before use.