A-A-50725
30 July 1988

COMMERCIAL ITEM DESCRIPTION

TEST SET, TELEPHONE SIGNAL

The General Services Administration has authorized the use of this Commercial Item Description.

Scope. This Commercial Item Description (CID), describes a signal test set. This CID is meant as a minimum requirement for the signal test set in which only those manufacturers that meet or surpass the following requirements are qualified per this CID.

Classification. Equipment covered by this CID may be commercially available equipment modified to the extent necessary to meet the following description. The equipment shall be type III, class 5, style F and color R in accordance with MIL-T-28800.

APPLICABLE DOCUMENTS

Specifications. The following specification forms a part of this CID to the extent specified herein. Unless otherwise specified, the issues of these documents shall be those listed in the issue of the Department of Defense Index of Specifications and Standards (DODISS) and supplement thereto, cited in the solicitation.

SPECIFICATION

MILITARY

MIL-T-28800	Test Equipment for use with Electrical and Electronic Equipment, General Specification for.

(Copies of specifications required by contractors in connection with specific acquisition functions should be obtained from the contracting activity or as directed by the contracting officer.)

Other publications. The following document forms a part of this CID to the extent specified herein. Unless otherwise specified, the issue of the documents which are DOD adopted shall be those listed in the issue of the DODISS specified in the solicitation. Unless otherwise specified, the issues of documents not listed in the DODISS shall be the issue of the nongovernment documents which is current on the date of the solicitation.

AMERICAN NATIONAL STANDARDS INSTITUTE

ANSI-C62.41-1980

AMSC N/A FSC 6625

DISTRIBUTION STATEMENT.A Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:46Z
Check the source to verify that this is the current version before use.

(Application for copies of ANSI publications should be addressed to the American National Standards Institute, 1430 Broadway, New York, NY 10018.)

Order of precedence. In the event of a conflict between the text of this CID and the references cited herein (except for associated detail specifications, specification sheets or MS standards) the text of this CID shall take precedence. Nothing in this CID, however, shall supersede applicable laws and regulations unless a specific exemption has been obtained.

Salient characteristics. the equipment shall be capable of operation and be bid sample tested within the accuracies, limits, and specifications herein.

1. Safety. The equipment shall be UL (Underwriters Laboratories) listed and approved and/or shall comply with the type III safety requirements of MIL-T-28800.

a. Leakage current protection. Protection shall be provided from leakage current in excess of 3.5 milliamperes (ma) rms (AC and DC) from any accessible conductive parts of equipment (including control shafts with knobs broken or removed and recessed calibration or adjustment controls) to either pole of the power source for any position of the power switch(es).

b. Radio frequency interference (RFI) capacitors. Any capacitor used in RFI filters shall be rated for at least four(4) times the operating voltage.

c. Type III ventilation openings.

(1) Ventilation openings shall be designed to prevent accidental contact with live electrical circuits and mechanical moving parts.

(2) Top ventilation openings shall incorporate small mesh screening or similar design to prevent the expulsion of hot electronic particles due to failure of electronic devices such as capacitors.

2. Parts, materials, and processes. Except for marking, the signal test set shall comply with the type III parts, materials, and process requirements of MIL-T-28800. Nonstandard parts approval requests shall be in acordance with MIL-T-28800. Front panel marking shall be in accordance with MIL-T-28800.

a. Restricted material. The equipment shall comply with the restricted material requirements of MIL-T-28800, including the requirement for mercury and radioactive materials, and shall contain no combination of materials which cause deterioration of any material contained in the equipment due to the effects of outgassing.

3. Design and construction. Unless otherwise specified herein, the equipment shall comply with the type III design and construction requirements of MIL-T-28800. The equipment shall be of solid-state construction. Encapsulation and embedment (potting) of subassemblies may be used only for high voltage circuits where construction is susceptible to leakage or corona, indicators and associated controls shall be in close proximity to each other, all potentiometer controls shall have at least a 270 degree rotation range (4.71 radians).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:46Z
Check the source to verify that this is the current version before use.

a. Signal input and output connectors. Unless otherwise specified, all signal input and output connectors shall be type 310.

b. Internal cooling. The equipment shall be capable of operating continuously for twenty four(24) hours at 55°C. If forced air is used, it shall be as specified in MIL-T-28800.

(1) Forced air cooling motor. DC brushless motors shall be utilized forthe forced air cooling motor in the equipment.

(2) Air filter. Equipment utilizing a cooling scheme which exhausts air from the equipment through the cooling fan shall not require an externally accessible air filter.

(3) Thermal switch. A thermal switch shall be provided which will interrupt circuit power to the signal test set when the internal temperature rises to the degree of possible damage to the electronic components. A minimum of 10°C hysteresis drop from this tripping temperature must occur before circuit power is again applied.

4. Electrical power sources. The equipment shall operate from nominal commercial, military, and shipboard power sources of -42 to -54 VDC.

a. Current drain. Maximum current drain of all components combined shall not exceed 500 milliamperes.

b. Fuses. Fuse(s) shall be provided, on the exterior rear panel for replacement without entering the enclosure. The fuse(s) shall automatically disconnect the equipment from the power source in the event of excessive current requirements.

c. Surge and transient voltage protection. The signal test set shall contain surge and transient voltage protection internal of the enclosure. The transient voltage protection circuit shall be so designed as to prevent electronic device damage within the system per type III solic state tolerance bands of MIL-T-28800, the surge voltage protection circuit shall conform to ANSI/IEEE Std. C62.41-1980, 2.5KV, category A. The surge protection circuit shall drain excess power due to surges directly to the ground conductor for personnel protection. Normal operation of the signal test set may be innibited and false operational signals may be generated under input voltage transients greater than ±20 percent from within the nominal type III steady-state tolerance bands of MIL-T-28800, damage (other than fuse failure) to the signal test set shall not occur during voltage transient conditions greater than ±20 percent within the maximum limits (30 percent voltage transients specified in MIL-T-28800. Normal operation of the signal test set shall be automatically resumed when the transient has passed and any blown fuses are replaced. Blown fuses shall be externally accessible. Testing shall be in accordance with MIL-T-28800.

5. Dimensions and weight. The size and weight of individual components shall be consistent with current commercial capabilities, refer to paragraphs 5.a and 5.b of this CID.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:46Z
Check the source to verify that this is the current version before use.

a. Equipment dimensions. The nominal dimensions per component for the signal test set should be approximately 9.5 inches height x 15.5 inches wide x 11.0 inches deep.

b. Equipment weight. The weight including all accessories (excluding manual) per component, shall not exceed 20 lbs.

6. Enclosure requirements. Unless otherwise specified herein, the equipment shall comply with the style I enclosure requirements of MIL-T-28800.

a. Rackmount. Rackmount capability shall be available as an option for the equipment specified herein.

b. Reusable pouch or container. An attached accessory pouch may be used for storage of all supplied accessories in lieu of front-cover storage.

7. Marking and identification.

a. Marking. The visible surface adjacent to panel facilities such as variable controls, indicators, jacks, connectors and fuse holder(s), shall be marked with a suitable word, phrase, or abbreviation thereof, indicating the purpose or use of the part.

(1) Panel marking. The front panel marking shall comply with the type III parts, materials and processes panel marking requirements of MIL-T28800, however, marking identification processes other than the type III requirements specified in MIL-T-28800 may be used, provided the marking and identification processes are capable of passing the front panel marking test when tested in accordance with this CID.

8. Environmental requirements.

a. Temperature requirements. Unless otherwise specified herein, the equipment shall meet its specified performance characteristics and other requirements when subjected to the class 5 temperature requirements specified in MIL-T-28800 and as listed below.

(1) Non-operating temperature. The equipment shall meet its performance and accuracy requirements in an non-operating environment of -20°C to +85°C.

(2) Operating temperature. The equipment shall meet its performance and accuracy requirements in an operating environment of 0°C to 65°C.

b. Humidity requirements. The equipment shall meet its required performance and accuracy requirements under humidity requirements for class 5 equipment as specified in MIL-T-28800.

(1) The equipment shall meet its performance and accuracy requirement in an operating and storage environment of 90% humidity.

c. Altitude requirements. The equipment shall meet its required performance and accuracy requirements under the altitude requirements for class 5 equipment as specified in MIL-T-28800.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:46Z
Check the source to verify that this is the current version before use.

d. Vibration requirements. The equipment shall meet the class 5 vibration requirements as specified in MIL-T-28800.

e. Mechanical shock requirements.

(1) Shock, functional. The equipment shall meet the class 5 shock, pulse requirement as specified in MIL-T-28800.

(2) Transit drop. The equipment shall meet transit drop requirement as specified in MIL-T-28800.

(3) Bench handling. The equipment shall meet the requirements as specified in MIL-T-28800.

(4) Shock, high impact. The equipment shall meet the requirements as specified in MIL-T-28800.

f. Electromagnetic interference (EMI). The equipment shall comply with MIL-T-28800 for class 5 equipment.

g. Magnetic environment. The equipment shall comply with the magnetic environment requirements of MIL-T-28800.

9. Reliability. Reliability verification testing, shall be in accordance with the type III, class 5 environmental conditions of MIL-T-28800.

a. Mean time between failures (MTBF). The equipment mean time between failures (MTBF) shall be 2,000 hours. The lower test MTBF shall be 2,000 and the upper test shall be MTBF 4,000.

b. Pre-conditioning requirement. Each deliverable equipment shall be capable of being subjected to the environmental screening effects of burn in and thermal cycling, with thermal cycling being accomplished in the sample testing phase.

(1) Burn-in. Burn-in shall expose the equipment to extended periods of elevated temperatures. (40°C for a minimum of 100 hours). Each equipment which fails during burn-in shall be repaired and fully retested, until failure free. Failures which occur during the burn-in test shall be noted and reported but shall not count toward the establishment of MTBF. Prior to and following burn-in, a satisfactory operation test (SOT) shall be performed. A satisfactory operation check (SOC) shall be performed conducted at selected times during the burn-in test.

(2) Thermal cycling. Thermal cycling shall expose the equipment to hot and cold cycles for class 5 equipment per MIL-T-28800. Each equipment which fails during thermal cycling shall be repaired and full cycling test shall be noted and reported but shall not count toward the cycling, a satisfactory operation check (SOC) shall be conducted during the last two upper and lower temperature limits.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:46Z
Check the source to verify that this is the current version before use.

10. Calibration and maintenance adjustments. The design of the system shall provide for readily accessible calibration adjustments and maintenance adjustments. These adjustments shall be provided by variable value components which are adjustable by the use of simple means. The calibration by substitution of selected components or parts in unacceptable unless specifically approved. The calibration adjustments, wherever possible, shall be accessible without removal of the instrument case or modules.

a. Maintainability. The equipment shall meet the type III maintainability requirements specified in MIL-T-28800.

11. Interchangeability. The equipment shall meet the type III interchangeability requirements specified in MIL-AT-28800.

12. Digital interface. The equipment shall meet the digital interface requirements per MIL-T-28800.

13. Performance characteristics.

a. The signal test set shall be capable of testing the performance of loop signaling and E & M signaling circuits in exchange and toll of telephone plants.

b. The signal test set shall provide controlled pulse generation with an adjustable pulse speed range of 3 to 20 pps and an adjustable pulse width range of to 275 msec.

c. The signal test set shall provide measurement of pulse speed accurate to ±0.5 pps and pulse ratio accurate to 1% for E & M and loop solid-state and 3% for loop selector and trunk type A relays.

d. The signal test set shall provide the following methods of measuring loop signaling:

(1) Selector type A relay performance,

(2) Trunk type A relay performance,

(3) Solid-state detector performance.

e. The signal test set shall provide switch-selectable E & M signaling options.

f. The signal test set shall provide supervisory lamps for busy circuit monitoring.

14. Calibration interval. After calibration, the equipment shall meet each performance requirement within the tolerances specified for a period of at least 9 months, based on an operating time of 1500 hours.

Preservation, packaging, labeling, and marking. The preservation, packaging, packing, labeling, and marking shall be a specified in the contract or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:46Z
Check the source to verify that this is the current version before use.

Certification. The contractor shall certify that the product offered meets the salient characteristics of this description and that the product conforms to the producer's own drawings, specifications, standards and quality assurance practices and is the same product offered for sale in the commercial marketplace. Small business that cannot offer products with established commercial market acceptability, may compete by submision of bid samples. (The Government reserves the rights to require proof of such conformance before first delivery and thereafter as may be otherwise provided for under the provisions of the contract.)

Known acceptable products. The following product(s) are known to have previously met the requirement of this CID and need not resubmit bid samples or certification. All other manufacturers must meet the requirements of this CID and submit to bid testing and certification as required.

Manufacturer	Model	NSN
GTE, Lenkurt	900-56686-00	6625-01-071-5725

Custodian:
Air Force - 99

Preparing Activity
Air Force - 82

Review Activities

(Project 6625-0699)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:46Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements All users of military standardization documents are invited to provide suggestions This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed In block 6, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems Enter in block 6 any remarks not related to a specific paragraph of the document If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements

☆ U.S Government Printing Office 1982—360 979/5804

(Fold along this line)

(Fold along this line)

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 73236 WASH DC

POSTAGE WILL BE PAID BY HQ U S. AIR FORCE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:46Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-	Test Set Telephone Signal

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR ☐ USER ☐ MANUFACTURER ☐ OTHER *(Specify)* _____
b ADDRESS *(Street, City, State, ZIP Code)*	

5 PROBLEM AREAS

a. Paragraph Number and Wording

b Recommended Wording

c Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *(Last, First, MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:46Z
Check the source to verify that this is the current version before use.