METRIC

A-A-51509B
12 February 1993
SUPERSEDING
A-A-51509A
11 April 1990 and
A-A-53230A
8 June 1989

COMMERCIAL ITEM DESCRIPTION

TUBE, TRACHEAL,
Radiopaque (Murphy)

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers disposable Murphy-type tracheal tubes commonly used in human patients in the practice of anesthesiology and in respiratory care.

Salient characteristics:

Style and design. Except as specified below, shall be in accordance with American National Standard for anesthetic equipment-tracheal tubes, ANSI Z79.14 or ASTM F1242-89. The tracheal tubes shall be the Murphy-type and shall be of the following types and sizes:

Type I with cuff

2.5 mm	6.5 mm
3.0 mm	7.0 mm
3.5 mm	7.5 mm
4.0 mm	8.0 mm
4.5 mm	8.5 mm
5.0 mm	9.0 mm
6.0 mm	10.0 mm

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

Type II without cuff

2.5 mm
3.5 mm
4.5 mm
5.5 mm

The tracheal tubes shall be supplied sterile and shall be disposable. The Luer fitting on the inflation valve shall be in accordance with ANSI/HIMA MD70.1, Standard for Medical Materiel Luer Taper Fittings or ISO 594. Each tracheal tube shall be provided with a 15 millimeter (mm) tracheal tube straight connector.

Material. The tracheal tube, tracheal tube cuff, the inflation valve, and the pilot balloon shall be fabricated of a polyvinyl chloride plastic material or other suitable plastic medical acceptable material meeting requirements of ANSIZ79.14 or ASTM F1242-89. Tube material shall conform to the United States Pharmacopeia (USP) requirements for Class IV plastics. The poppet shall be fabricated from neoprene or other suitable material meeting requirements of ANSIZ79.14 or ASTM F1242-89. If a valve spring is used, it shall be fabricated of AISI 300 Series corrosion-resistant steel.

The connector shall be fabricated of polypropylene or other suitable plastic and shall comply with the requirements of Z79.2 American National Standard for tracheal tube connectors and adaptersor ASTM 1242-89. The tracheal tubes shall have a radiopaque stripe extending the entire length of the tube.

Leakage. The cuff and inflation system shall show no evidence of leakage when it has been inflated to an air pressure of 22 mm Hg for not less than 30 minutes.

Lead content. The tracheal tube and cuff shall contain not more than 5 parts per million of lead.

Sterility. The tracheal tube and connector shall be sterile in accordance with the USP or AAMI guidelines. Sterility assurance level shall be .0001.

Pyrogens and safety. The tracheal tube shall meet the USP requirements for pyrogenicity and safety.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

Radiopacity. The radiopaque line shall be permantly shown when tested in accordance with Method C of ASTM Designation F640.

Workmanship. The tracheal tubes shall be uniform in quality and free from any defects which affect the appearance or utility of the tube. Tracheal tubes shall be free from foreign matter, embedded particles, burrs, flashing, rough or sharp areas, cracks and other defects. Shall be free from large bubbles which affect the physical integrity of the device and free from areas of excessive bubbles which could affect the transparency of the device.

Unit. Package (PG). One package containing 10 tracheal tubes of one size only constitutes one unit. The tubes shall be individually packaged sterile in a peel-open type container capable of maintaining the sterility of the contents unless the package is damaged or opened. Package shall also be marked with tube I.D. and length.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for tests purposes shall be one tracheal tube. Sampling shall be Inspection level S-3 with an AQL of 2.5 (percent defective).

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities, labels and markings shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

NOTE: The following National Stock Numbers, Sizes, Item Identifications and their respective Types are applicable to this document:

NATIONAL STOCK NUMBER	SIZE	ITEM IDENTIFICATION	TYPE
6515-00-105-0653	3.0	TUBE, TRACHEAL, Radiopaque, Murphy, Disposable, 3.0 mm I.D., 10s.	I
6515-01-154-2150	2.5	TUBE, TRACHEAL, Radiopaque, Murphy, Disposable, 2.5 mm, I.D., 10s.	I
6515-01-153-1683	3.5	TUBE, TRACHEAL, Radiopaque, Murphy, Disposable, 3.5 mm, I.D., 10s	I
6515-00-105-0664	4.0	TUBE, TRACHEAL, Radiopaque, Murphy, Disposable, 4.0 mm I.D., 10s.	I
6515-01-176-0625	4.5	TUBE, TRACHEAL, Radiopaque, Murphy, Disposable, 4.5 mm, I.D., 10s.	I
6515-00-105-0707	5.0	TUBE, TRACHEAL, Radiopaque, Murphy, Disposable, 5.0 mm I.D., 10s.	I

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

NATIONAL STOCK NUMBER	SIZE	ITEM IDENTIFICATION	TYPE
6515-00-105-0720	6.0	TUBE, TRACHEAL, Radiopaque, Murphy, Disposable, 6.0 mm I.D., 10s.	I
6515-01-250-3519	6.5	TUBE, TRACHEAL, Radiopaque, Murphy, Disposable, 6.5 mm, I.D., 10s	I
6515-00-105-0744	7.0	TUBE, TRACHEAL, Radiopaque, Murphy, Disposable, 7.0 mm I.D., 10s.	I
6515-01-036-9034	7.5	TUBE, TRACHEAL, Radiopaque, Murphy, Disposable, 7.5 mm I.D., 10s.	I
6515-00-105-0759	8.0	TUBE, TRACHEAL, Radiopaque, Murphy, Disposable, 8.0 mm I.D., 10s.	I
6515-01-036-9035	8.5	TUBE, TRACHEAL, Radiopaque, Murphy, Disposable, 8.5 mm I.D., 10s.	I
6515-00-105-0808	9.0	TUBE, TRACHEAL, Radiopaque, Murphy, Disposable, 9.0 mm I.D., 10s.	I
6515-00-105-0815	10.0	TUBE, TRACHEAL, Radiopaque, Murphy, Disposable, 10.0 mm I.D., 10s.	I
6515-01-209-0674	2.5	TUBE, TRACHEAL, PEDIATRIC, Radiopaque, Murphy, 2.5 mm, I.D., 14 cm Long, Disposable, 10s	II
6515-01-209-0672	3.5	TUBE, TRACHEAL, PEDIATRIC, Radiopaque, Murphy, 3.5 mm, I.D., 18 cm Long, Disposable, 10s	II
6515-01-252-2483	4.5	TUBE, TRACHEAL, PEDIATRIC, Radiopaque, Murphy, 4.5 mm, I.D., 22 cm Long, Disposable, 10s	II

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

NATIONAL STOCK NUMBER	SIZE	ITEM IDENTIFICATION	TYPE
6515-01-252-0038	5.5	TUBE, TRACHEAL, PEDIATRIC, Radiopaque, Murphy, 5.5 mm, I.D., 26 cm Long, Disposable, 10s	II

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6515-5331
Location: c:\word\doc\nish\tube

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-51509B	12 February 1993

3. DOCUMENT TITLE
TUBE, TRACHEAL, Radiopaque (Murphy)

4. NATURE OF CHANGE *(Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)*

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
Defense Personnel Support Center ATTN: DPSC-MSE	(215)7372870	444-2870
c. ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 *Previous editions are obsolete.* 198/290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.