A-A-51708
17 October 1986

COMMERCIAL ITEM DESCRIPTION

MALLET, BONE SURGERY, STAINLESS STEEL

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a stainless steel bone mallet for use in general orthopedic surgical procedures.

Salient characteristics:

General. Shall be a bone mallet for use during orthopedic surgery.

Material. The entire mallet shall be fabricated from class 3 austenitic stainless steel or class 6 ferritic stainless steel in accordance with ASTM F899. As an alternate, the mallet shall be fabricated from ferritic corrosion-resisting steel having the following composition: Carbon 0.10 - 0.17%, Chromium 15.5 - 17.5%, Manganese 1.5% maximum, Phosphorous 0.045% maximum, Sulfur 0.15 - 0.35%, Silicone 1% maximum and Molybdenum 0.2 - 0.3%. Should the mallet be fabricated from class 3 material, the head shall have a Rockwell hardness of B-70 to C-30. Should the mallet be fabricated from class 6 ferritic stainless steel, the head shall have a Rockwell hardness of B-80 to B-100. The mallet head shall be tested for Rockwell hardness in accordance with ASTM E18.

Size. The one pound mallet shall have a head that weighs approximately one pound. The two pound mallet shall have a head that weighs approximately two pounds.

Characteristics. The mallet shall be in accordance with the dimensional requirements listed in Table I. The handle shall provide a non-slip grip and may be fluted or knurled.

TABLE I

	Striking surface O.D. (inches)	Mallet head overall length (inches)	Overall length mallet (inches)
1 LB mallet	1-1/4 ± 1/8	2-7/8 ± 3/8	9 ± 1-1/2
2 LB mallet	1-1/2 + 7/16 - 1/4	2-7/8 ± 3/8	9-1/4 ± 1-1/2

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

Passivation. Materials used in the fabrication of the mallet shall be passivated, if required.

Finish. The mallet shall be satin or mirror in accordance with Federal Specification GG-I-526.

Boil test. The mallet shall be tested in accordance with Federal Specification GG-I-526.

Copper sulfate test. Should the mallet be fabricated from class 6 ferritic stainless steel, the mallet shall be tested in accordance with Federal Specification GG-I-526.

Marking. Mallets shall bear the name or registered trademark of the manufacturer and the work "stainless". In addition, instruments of foreign manufacture shall bear the name of the country of manufacture.

Workmanship. The mallet shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (Ea). One mallet of the same size, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:00Z
Check the source to verify that this is the currentversion before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for Tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. AQL shall be 1.0 percent defective. Inspection level shall be S-2. The unit of product for test purposes shall be one mallet.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

COMMERCIAL ITEM DESCRIPTION A-A-51708

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-233-8850	Mallet, Bone, Surgical, Stainless Steel, 1 lb.
6515-01-234-0253	Mallet, Bone, Surgical, Stainless Steel, 2 lb.

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC SF022

Project No. 6515-3786

☆U S GOVERNMENT PRINTING OFFICE 1986 — 704-036/45864

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:00Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-ATTS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:00Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-51708	MALLET, BONE SURGERY, STAINLESS STEEL

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:00Z
Check the source to verify that this is the current version before use.