INCH POUND
MIL-T-37637A
1 October 1990
SUPERSEDING
MIL-T-37637
18 July 1978

MILITARY SPECIFICATION

# TABLE, OPERATING, FIELD

This specification is approved for use by all Departments and Agencies of the Department of Defense.

## 1. SCOPE

1.1 Scope. This specification covers an adjustable field operating table with carrying case.

## 2. APPLICABLE DOCUMENTS

2.1 Government documents.

2.1.1 Specifications, standards, and handbooks. The following specifications, standards, and handbooks form a part of this document to the extent specified herein. Unless otherwise specified, the issues of these documents shall be those listed in the issue of the Department of Defense Index of Specifications and Standards (DODISS) and supplement thereto, cited in the solicitation (see 6.2).

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Headquarters, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, Pa 19101, by using the self-addressed Standardization Document Improvement Proposal, (DD Form 1426) appearing at the end of this document or by letter.

AMSC/NA FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release: distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:10Z
Check the source to verify that this is the current version before use.

SPECIFICATIONS

FEDERAL

| | | |
|---|---|---|
| V-T-276 | - | Thread, Cotton |
| V-T-285 | - | Thread, Polyester |
| QQ-A-200/8 | - | Aluminum Alloy 6061 Bar, Rod, Shapes, Tube and Wire, Extruded |
| QQ-A-250/1 | - | Aluminum 1100, Plate and Sheet |
| QQ-A-250/2 | - | Aluminum Alloy 3003, Plate and Sheet |
| QQ-A-250/8 | - | Aluminum Alloy 5052, Plate and Sheet |
| QQ-A-596 | - | Aluminum Alloy, Permanent and Semi-permanent Mold Castings |
| QQ-A-601 | - | Aluminum Alloy Sand Castings |
| CCC-C-419 | - | Cloth, Duck, Cotton, Unbleached, Plied-Yarns, Army and Numbered |
| CCC-D-950 | - | Dyeing and Aftertreating Processes for Cotton Cloths |
| PPP-B-601 | - | Boxes, Wood, Cleated-Plywood |
| PPP-B-621 | - | Boxes, Wood, Nailed and Lock-Corner |
| PPP-B-636 | - | Boxes, Shipping, Fiberboard |

MILITARY

| | | |
|---|---|---|
| MIL-P-116 | - | Preservation, Methods of |
| MIL-C-496 | - | Clips, End, Strap |
| MIL-W-530 | - | Webbing, Textile, Cotton, General Purpose, Natural or in Colors |
| MIL-B-543 | - | Buckle, Tongueless and Web Strap |
| MIL-T-3530 | - | Thread and Twine, Mildew Resistant or Water Repellent Treated |
| MIL-M-7298 | - | Manuals, Commercial Off-The-Shelf |
| MIL-A-8625 | - | Anodic Coatings, for Aluminum and Aluminum Alloys |
| MIL-L-10547 | - | Liner, Case, and Sheet, Overwrap, Water-Vaporproof or Waterproof, Flexible |
| MIL-F-10884 | - | Fasteners, Snap |
| MIL-T-43566 | - | Tape, Textile Cotton or Polyester, General Purpose, Natural or in Colors |

STANDARDS

FEDERAL

| | | |
|---|---|---|
| FED-STD-151 | - | Metal: Test Methods |
| FED-STD-191 | - | Textile Test Methods |
| FED-STD-751 | - | Stitches, Seams, and Stitchings |



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:10Z
Check the source to verify that this is the current version before use.

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes
MIL-STD-129 - Marking for Shipment and Storage
MIL-STD-147 - Palletized Unit Loads
MIL-STD-794 - Parts and Equipment, Procedures For Packaging and Packing of

(Unless otherwise indicated, copies of federal and military specifications, standards, and handbooks are available from the Standardization Documents Order Desk, Building 4D, 700 Robbins Avenue, Philadelphia, PA 19111-5094.)

2.1.2 Other Government documents, drawings, and publications. The following other Government documents, drawings, and publications form a part of this document to the extent specified herein. Unless otherwise specified, the issues are those cited in the solicitation.

DRAWING

Defense Personnel Support Center (DPSC)

DWG. 1796 - Table, Operating, Field

Note: Sheet 3 of 4 and sheet 4 of 4 of DPSC Dwg. 1796 are for reference only. Construction as indicated in this specification takes precedence.

(Copies of the drawing may be obtained from: The Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, PA 19101, ATTN: DPSC-RST).

2.2 Non-Government publications. The following documents form a part of this document to the extent specified herein. Unless otherwise specified, the issues of the documents which are DOD adopted are those listed in the issue of the DODISS cited in the solicitation. Unless otherwise specified, the issues of documents not listed in the DODISS are the issues of the documents cited in solicitation.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:10Z
Check the source to verify that this is the current version before use.

AMERICAN IRON AND STEEL INSTITUTE (AISI)

Steel Products Manual, Stainless and Heat Resisting Steel

(Application for copies should be addressed to the American Iron and Steel Institute, 150 East 42nd Street, New York, NY 10017).

COLOR ASSOCIATION OF THE UNITED STATES, INC.

Unites States Army Standard Color Card for the Official Standardization Shades of Sewing Threads

(Application for copies should be addressed to the Color Association of the United States, Inc., 409 West 44th Street, New York, NY 10036).

(Nongovernment standards and other publications are normally available from the organizations which prepare or which distribute the documents. These documents also may be available in or through libraries or other informational services.)

2.3 Order of precedence. In the event of a conflict between the text of this document and references cited herein, (except for related associated detail specifications, specification sheets, or MS standards), the text of this document takes precedence. Nothing in this document, however, supersedes applicable laws and regulations unless a specific exemption has been obtained.

3. REQUIREMENTS.

3.1 Materials.

3.1.1 Aluminum Alloy.

3.1.1.1 Sheet. Aluminum alloy sheet shall be alloy 5052-H32 in accordance with QQ-A-250/8, alloy 1100-H14, or alloy 1100-H16, in accordance with QQ-A-250/1, or alloy 3003-H14, or alloy 3003-H16, in accordance with QQ-A-250/2, as specified.

3.1.1.2 Bar, rod, shapes, and tubing. Aluminum alloy bar, rod, shapes, and tubing, shall be fabricated from alloy 6061-T6 in accordance with QQ-A-200/8.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:10Z
Check the source to verify that this is the current version before use.

3.1.1.3 Castings. Aluminum alloy castings shall be alloy 355-T6, alloy 356-T6, or alloy A214, either permanent-mold castings in accordance with QQ-A-596, or sand castings in accordance with QQ-A-601.

3.1.2 Corrosion-resisting steel.

3.1.2.1 Bar, rod, and shapes. Corrosion-resisting steel bar, rod, and shapes, shall be AISI type 410 or 416.

3.1.2.2 Sheet. Corrosion-resisting steel sheet shall be AISI type 302, type 304, or type 430.

3.1.2.3 Wire. Corrosion-resisting steel wire shall be AISI type 302. The wire shall be treated for spring temper condition.

3.1.3 Plywood. Plywood shall be 1/4 inch thick, Douglas-fir, exterior grade.

3.1.4 Cotton duck. The cotton duck shall be type I, No. 8 hard texture, as listed in CCC-C-419. The duck shall be vat dyed olive drab No.7, water repellent and mildew resistant treated, in accordance with CCC-D-950, type 1, class B except that the mildew resistant inhibitors shall be limited to a, e, f, and g.

3.1.5 Cotton webbing. The cotton webbing shall be type III, class 4, 1 inch wide, vat dyed olive drab No. 7, or type IIa, class 1, 1 inch wide, or type III, class 1, 1-1/2 inches wide, in accordance with MIL-W-530.

3.1.6 Cotton tape. The tape shall be type I, class 4, 3/4 inch wide, vat dyed olive drab, in accordance with MIL-T-43566.

3.1.7 Thread.

3.1.7.1 Cotton. The cotton thread shall be type IIIa or type IIIb, 3 ply, ticket No. 10, in accordance with V-T-276. The thread shall be vat dyed olive drab, shade S-1, in accordance with the United States Army Color Code Card. Mildew resistance shall be type K, class 1, in accordance with MIL-T-3530.

3.1.7.2 Polyester. The polyester thread shall be type I, class 1, in accordance with V-T-285.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:10Z
Check the source to verify that this is the current version before use.

3.2 Design and construction. The operating table shall consist of a tabletop supported by an adjustable leg assembly. A rack and pinion type of adjusting mechanism shall provide angular adjustment of the entire tabletop. The tabletop shall be constructed in four sections:

1. Removable headrest
2. Back section
3. Seat section
4. Leg section

The seat section shall be fixed to the tabletop frame, and shall be provided with a "U" shaped cutout for urological work. The back section and leg section frames shall be hinged to the main frame so that they can be angularly adjusted relative to the seat section. In addition, the hinge between the seat section and leg section shall be provided with a slot-pin arrangement to permit the leg section to be placed out of position for nonuse or for storage. Side rails shall run the full length of each side of the back, seat and leg sections of the tabletop to support removable and adjustable accessory sockets. The support for the tabletop shall be a triangular leg assembly hinged to the main frame. The table height shall be adjusted by varying the base length of the triangular leg assembly. The table adjustments plus the required accessories shall permit the table to be placed into the following operative positions:

1. Trendelenberg
2. Reverse Trendelenberg
3. Genito-Urinary
4. Straight Chair

When tested as specified in 4.4.3, all adjusting mechanisms shall operate without binding, play, misalignment, or chatter, and the table shall be stable in all possible positions. The table shall pass the load test when tested as specified in 4.4.4.1. The tabletop shall pass the height test as specified in 4.4.3.1.

3.2.1 Tabletop. Each tabletop section shall have the overall dimensions shown below:

Seat section - 14 inches plus or minus 1/4 inch long, by 18-1/4 inches plus or minus 1/8 inch wide.

Back section - 22 inches plus or minus 1/4 inch long, by 17 inches plus or minus 1/8 inch wide.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:10Z
Check the source to verify that this is the current version before use.

Leg section - 25 inches plus or minus 1/4 inch long, by 18-1/4 inches plus or minus 1/8 inch wide.

Removable Headrest - 10-1/4 inches plus or minus 1/4 inch by 16 inches plus or minus 1/4 inch.

The entire table, when in the horizontal position with the headrest unextended, shall have an overall length of 73 inches plus or minus 1/4 inch.

3.2.1.1 Seat section. The seat section of the tabletop shall be formed from alloy 5052-H32 aluminum alloy sheet with a thickness of 0.064 inch. The aluminum sheet shall be shaped to form a snug fit over the outside of the main frame of the seat section (see 3.2.2.1). The seat section shall have a perineal cut-out at the foot end. The cut-out shall have a radius of 5-1/4 inches, shall be centrally located, and shall extend approximately 2 inches into the seat section. If a cast aluminum alloy frame is used, the seat section shall be cast integrally with the frame.

3.2.1.2 Back section. The back section of the tabletop shall be formed from alloy 5052-H32 aluminum alloy sheet with a thickness of 0.064 inch. The aluminum sheet shall be shaped to form a snug fit over the outside of the main frame of the back section (see 3.2.2.2).

3.2.1.3 Leg section. The leg section of the tabletop shall be formed from alloy 5052-H32 aluminum alloy sheet with a thickness of 0.064 inch. The aluminum sheet shall be shaped to form a snug fit over the outside of the main frame of the leg section (see 3.2.2.3).

3.2.1.4 Removable headrest. The removable headrest shall either be cast from aluminum alloy or fabricated with an aluminum alloy frame and a 0.064 inch thick aluminum alloy sheet. The headrest shall be provided with 1/2 inch diameter steel rods which shall fit into the tubular guides of the back section frame, and shall permit a 6 inch extension of the headrest from the back section. The headrest shall be capable of being elevated or lowered by means of a ratchet device with fingertip control. The ratchet device shall be constructed of corrosion-resisting steel, and shall allow the headrest to be placed at a minimum of two different angles, in addition to the horizontal positions.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:10Z
Check the source to verify that this is the current version before use.

3.2.2 Frames.

3.2.2.1 Main frame. The main frame shall be common to the back and seat sections (see 3.2.1.2 and 3.2.1.1 ). The frame shall be constructed from 1-1/2 inch by 1-1/2 inch by 1/4 inch thick aluminum alloy angle, with horizontal legs facing inward, except that the head end shall be a 2 inch wide by 3/8 inch thick aluminum alloy bar stock with the 2 inch dimension horizontal. The seat section end angle, which may be cast aluminum alloy, shall be centrally curved to provide for the seat section "U" shaped cutout. The main frame shall be reinforced transversely by a length of 1 inch outside diameter, by 0.095 inch wall thickness aluminum alloy tubing welded at each side to the vertical legs of the main frame angles. Tubing shall also serve as a support for the seat section tabletop. Alternately, the main frame shall be a cast aluminum alloy frame, with integral seat section (see 3.2.1.1) and, if desired, integral leg hinges, and back section pawl bosses, may be employed. The casting shall be properly ribbed for high strength and all sections of the casting shall have dimensions equivalent to or greater than those specified for the fabricated frame. When the frame is cast, or when aluminum sheet at least 0.125 inch thick is used, the reinforcing tube is not required. The main frame shall be provided with side rails running the entire length (see 3.2.3).

3.2.2.2 Back section frame. The side members of the back section frame shall be formed from 1 inch by 1 inch by 1/8 inch aluminum alloy angle to which shall be welded at the head end a 1-1/4 inch by 1-1/4 inch by 1/8 inch thick aluminum alloy channel, and at the foot end shall be welded a 1 inch outside diameter aluminum alloy tubing with a wall thickness of 0.095 inch. The head end shall be provided with entering holes in the flange of the back section, and with two tubular guides on 11-1/2 plus or minus 1/4 inch centers to receive and support the rods of the removable headrest (see 3.2.1.4). Each guide tube shall be 7 plus or minus 1/4 inch long, 3/4 inch outside diameter aluminum alloy tubing having a wall thickness of 0.065 inch. Guide tubes shall pierce and be welded to the frame channel at one end, and shall be similarly joined to a similar channel welded to the frame side angles at the other end. The guide tubes shall each be reinforced and tapped to receive a thumbscrew to act as a setscrew for the headrest rods. The back section shall be hinged to the main frame at its foot end by means of a 1/2 inch diameter corrosion-resisting steel rod inserted into the back section frame tube and passing into holes in the vertical legs of the main frame angles. The rod shall be prevented from moving



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:10Z
Check the source to verify that this is the current version before use.

axially, and thus dropping out of its bearing surfaces by means of setscrews, or keepers. Aluminum alloy tubular spacers shall be inserted between the back section frame and the mainframe (see 3.2.2.1) at the hinged points. A ratchet and pawl shall permit the back section to be placed in the following angles with respect to the the main frame (in addition to the horizontal position):

* 10 degrees plus or minus 2 degrees
* 25 degrees plus or minus 2 degrees
* 45 degrees plus or minus 2 degrees
* 65 degrees plus or minus 2 degrees

When tested as specified in 4.4.3.4, the ratchet and pawl arrangement shall permit the back section to be placed in the above angles. The back section shall pass the load test when tested as specified in 4.4.4.2.

The ratchet assembly shall consist of two 3/8 inch thick by 1 inch wide aluminum alloy ratchet bar, properly notched and joined, by means of welding within 2 inches of their midpoints and the hinged ends, by 14 inch lengths of 1/2 inch outside diameter aluminum alloy tubing having a wall thickness of 0.058 inch. The ratchet shall be hinged to the backrest by means of a 3/8 inch diameter corrosion-resisting steel rod passing through the ratchet bars and ratchet tubing into holes in the back section frame. The ratchets shall be kept in position by means of aluminum alloy tubular spacers. The outer 2 spacers shall be approximately 7/8 inches long. The inner spacer shall be approximately 17-7/8 inches long. Spacer clearances shall be close to prevent sideways movement of the back section. The pawls shall each be a 1-3/4 plus or minus 1/8 inch length of 1/2 inch diameter corrosion-resisting steel rod, joined to the underside of the horizontal leg of each main frame side angle by means of a cast aluminum alloy fitting securely bolted to the main frame with two bolts.

3.2.2.3 Leg section frame. The leg section frame shall consist of 1-1/4 inch by 1-1/4 inch by 1/8 inch thick aluminum alloy side channel positioned with the flanges downward. At the head end shall be welded a 1 inch outside diameter aluminum alloy tubing having a wall thickness of 0.095 inch. The leg section shall be hinged to the seat section main frame by means of a pair



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:10Z
Check the source to verify that this is the current version before use.

of sturdy hinges, each fitted with hinge pins of a 1/2 inch diameter corrosion-resisting steel rod, welded into a corrosion-resisting steel hinge guide bolted to the leg section channel. The cast aluminum alloy hinge shall be 6 inches long, and shall be slotted so that when the leg section is set perpendicular to the seat section, it can be lowered out of position The hinge slot shall be approximately 5 inches long. A spring-loaded trigger, pinned to each hinge inside a vertical slot in the hinge, shall hold the leg of the leg section out of position when desired. The trigger shall be formed from 0.093 inch thick corrosion-resisting steel sheet. The trigger spring shall be formed from corrosion-resisting steel wire, shall consist of 25 to 30 turns, shall be 1-3/4 inches plus or minus 22-1/2 inches long, and shall have an outside diameter of 3/16 inch plus or minus 1/64 inch. The ends of the spring shall be closed and ground. The trigger spring shall be located inside the lower section of the hinge and shall be held in place by a retaining screw and lock washer. Hinge guides shall be 1-1/8 inches wide by 3/16 inch thick corrosion-resisting steel, 4-1/2 plus or minus 1/8 inch long to act as guides for the cast aluminum hinge when the leg section is being lowered. The leg section shall be adjustable by means of a ratchet and pawl arrangement at the following angles below the horizontal with respect to the seat section (as well as horizontal):

* 15 degrees plus or minus 2 degrees
* 30 degrees plus or minus 5 degrees
* 45 degrees plus or minus 5 degrees

When tested as specified in 4.4.3.5, the ratchet and pawl arrangement shall permit the leg section to be placed in the above angles.

The ratchet shall consist of a 2 inch by 2 inch by 1/4 inch thick aluminum alloy angle in which the vertical leg shall be slotted. The ratchet angle shall be welded to the end angle of the frame and similar transverse reinforcing angle. The pawl or ratchet bar shall consist of 1/2 inch diameter corrosion-resisting steel rod ratchet rod hinge assembly bolted to each side of the main frame. The brace of the hinge assembly shall be locked in the vertical position in normal use by means of a spring-operated bullet catch, but shall be capable of being unlocked and placed in the horizontal position when it is desired to lower the leg section out of position. The ratchet bar shall be activated by a



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:10Z
Check the source to verify that this is the current version before use.

ratchet bar release lever permitting fingertip adjustment of the leg section. The lever shall be constructed of 1/2 inch wide by 3/16 inch thick corrosion-resisting steel, and shall be held in place by a suitable corrosion-resisting steel keeper. The leg section ratchet mechanism shall be so constructed that it shall be necessary to elevate the leg section slightly before it can be lowered into desired position to prevent inadvertent dropping of the leg section while in use. The leg section shall be provided with siderails running the entire length (see 3.2.3).

3.2.3 Siderails. The siderails shall be 3/8 inch thick, 1-1/8 inch wide, made of aluminum alloy. They shall be joined to the main and leg section frames by means of 5/16 inch long, 5/8 inch diameter, aluminum alloy tubular spacers, held in place with 5/16 inch diameter countersunk machine screws, and jamnuts. Tapped screw holes may be used where jamnuts are impractical. The spacers shall be on 5-1/4 inch to 7-3/4 inch centers. The siderails on the main section shall be a minimum of 35-3/4 inches long with a minimum of 6 spacers. The siderails on the leg section shall be a minimum of 23-3/4 inches long with a minimum of 4 spacers.

3.2.4 Leg assembly. The leg assembly shall have the following construction: The legs shall be hinged to the main frame by means of a bolt, washer, and cotter pin. Tubing for the leg assembly and cross braces shall be 1-1/4 inches outside diameter aluminum alloy tubing, having a wall thickness of 0.120 inch. Each leg shall have an aluminum alloy plug in the end. The plug at the top end of the leg shall be extended to form part of the hinge arrangement. There shall be 4 cross braces which shall be continuously welded to the legs. The lower brace shall be welded 4-3/8 inches from the bottom of the leg. The upper leg brace shall be welded 19-3/4 inches from the bottom of the leg. Each leg shall be provided with a cast aluminum alloy floorplate of 3 inch minimum diameter, hinged to each leg by means of a bolt and locknut. The height of the table shall be controlled by the spread of the legs. The height of the table shall be 38 inches plus or minus 1/4 inch in the highest position, and 33-1/2 inches plus or minus 1/4 inch in the lowest position. The spread of the legs shall be accomplished by means of manually telescoping a 3/4 inch diameter aluminum alloy rod within a 1 inch outside diameter aluminum alloy tubing. The rod shall be locked into place by means of a spring-loaded pin which shall engage into one of a series of six holes in the rod, each hole approximately 3 inches apart. The tube end of the telescoping assembly shall be bolted to a clevis continuously welded to the lower cross brace. The



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:10Z
Check the source to verify that this is the current version before use.

rod end shall be bolted to a clevis, welded to the other lower cross brace, by means of a positive spring loaded lock, so that the telescoping assembly can be disengaged at this end and swung out of the way when the table is folded for placing into the carrying case.

3.2.5 Adjusting mechanism. In all height positions except the lowest, the table shall be capable of being tilted to 50 degrees minimum on each side of the horizontal (50 degrees trendelenberg and 50 degrees reverse trendelenberg) by means of a hand operated rack and pinion, worm-gear arrangement. The table shall achieve these trendelenberg and reverse trendelenberg positions when tested as specified in 4.4.3.2 and 4.4.3.3, respectively. The worms and gears shall be made of material as specified in 3.3 and shall have a center distance of approximately 1-1/4 inches. They shall be enclosed in two cast aluminum alloy gear boxes supported on a leg assembly cross brace. The gear boxes shall have cast aluminum alloy or aluminum alloy sheet covers. Each gear box shall be bolted between two ears continuously welded to the upper support brace between the legs on the table head end. The worm shaft shall be 1/2 inch diameter corrosion-resisting steel rod and shall have a 3/16 inch diameter corrosion-resisting steel pin on each side to engage with the hand crank. The pins shall be 3/4 inch long and shall be located 1 inch from each end of the worm shaft. The two racks shall be joined to the tabletop by means of a 1/2 inch diameter corrosion-resisting steel rod, held in place by aluminum alloy castings bolted to each side of the main frame. Racks shall be securely welded to a tubular sleeve which fits smoothly over the corrosion-resisting steel rod. As an alternate,the racks shall be joined to the tabletop by means of a one inch side "U" shaped bracket, with legs of the "U" bracket joined to the frame by means of 1/2 inch diameter corrosion-resisting steel rods, securely inserted and locked to the frame. The racks shall be securely bolted to the "U" bracket.

3.2.5.1 Handcrank. The handcrank shall consist of an arm and a handle. The arm may be of one or two piece construction. The arm socket shall be formed from 3/4 inch diameter aluminum alloy rod, drilled to receive a 1/2 inch steel rod and provided with two 7/32 inch slots, each slot being a minimum of 1/2 inch long, to engage the pins on the worm shaft. The arm socket piece shall be 3-1/4 inches plus or minus 1/8 inch long. The arm shaft shall be formed of aluminum alloy, either flat measuring 3/4 inch wide by 3/8 inch thick, or round with a 1/2 inch diameter. The arm shaft shall be 6-1/2 inches plus or minus 1/8 inch long. The



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:10Z
Check the source to verify that this is the current version before use.

handle shall be black plastic or smoothly finished hardwood , enameled black, 2-1/2 inches minimum in length. The handle pin shall be made from not less than 3/8 inch diameter corrosion-resisting steel wire. The handle shall be free-turning on the pin. The handle pin shall be joined to the arm shaft in such a way that crank handle is retained within 10 degrees of the horizontal position for cranking, and it can be folded down when it is not in use.

3.2.6 Accessories. Each table shall be provided with the following removable accessories:

* One pair of adjustable shoulder supports (see 3.2.6.1).

* One angulating anesthetist's screen (see 3.2.6.2).

* One pair of lithotomy leg holders (see 3.2.6.3).

* One adjustable footrest (see 3.2.6.4).

* Four litter supports (see 3.2.6.5).

* One pair of wristlets with adjustable holders (see 3.2.6.6).

* Six accessory support sockets (see 3.2.6.7).

3.2.6.1 Shoulder support. Each shoulder support shall consist of a suitably curved cast aluminum alloy shoulder rest covered with a 7/8 inch thick sponge-rubber pad, and a shoulder rest mounting bracket formed from a 12-1/2 inch plus or minus 1/8 inch long, by 1 inch wide, by 1/4 inch thick aluminum alloy bar. The mounting bracket shall be formed by bending the aluminum alloy bar at a right angle 5-1/2 inches plus or minus 1/8 inch from one end. The 5-1/2 inch leg of the mounting bracket shall fit the vertical opening in the dual function socket (see 3.2.6.7), and the shoulder rest shall slide onto the other mounting bracket leg. The shoulder rest shall be cable of being secured at any point along the length of the mounting bracket leg.

3.2.6.2 Anesthetist's screen. The angulating anesthetist's screen shall be formed from 1/2 inch diameter aluminum alloy rod and shall fit into the opening in the angulating section of the dual function sockets (see 3.2.6.7). The top height of the screen rod, when in place in the dual function sockets, shall be



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:10Z
Check the source to verify that this is the current version before use.

17-3/8 inches plus or minus 1/8 inch above the tabletop. The transverse width of the rod verticals shall be 23 inches plus or minus 1/4 inch. The 90 degree bend end sections of the screen rod, which fit into the dual function sockets, shall be a minimum of 5-1/2 inches long.

3.2.6.3 Leg support (stirrup). Each lithotomy leg support shall be fabricated from 5/8 inch diameter aluminum alloy rod. The support rod shall be "L" shaped, 33 inches plus or minus 1/4 inch long, shall have a 5 1/2 inches plus or minus 1/2 inch long leg at one end. The long leg shall fit the opening in the leg holder socket (see 3.2.6.7). The short leg shall be flattened and drilled near the end to receive a ring (see 3.6). This ring shall also be attached to one end of a swivel eye bolt (see 3.7). The other end of the swivel eye bolt shall support a D-ring (see 3.6). Each "D" ring shall be furnished with a leg support strap. The strap shall be type IIa, class I, cotton webbing formed into a loop measuring 14 inches plus or minus 1/4 inches in length when measured flat, and shall be securely sewn together using cotton or polyester thread (see 3.1.7). Both leg supports shall fit into the cotton duck pouches (see 3.2.8). The stirrup shall pass the load test when tested as specified in 4.4.4.3.

3.2.6.4 Footrest. The adjustable footrest shall have a frame formed from 1 inch wide by 1/4 inch thick aluminum alloy bar, over which shall be fitted a 0.064 inch thick sheet of aluminum alloy. The aluminum sheet shall cover the sides of and be bent under the frame. The frame shall extend past the aluminum sheet on two sides in order to slide the footrest into the footrest support sockets (see 3.2.6.7). The outside dimensions of the finished footrest shall be 20-1/2 inches long by 8 inches wide with a tolerance of plus or minus 1/8 inch. The frame extension shall be 4-1/2 inches long with suitable stops placed on the aluminum alloy bar 2-1/2 inches from the free end as a means of locating the footrest in the footrest support sockets.

3.2.6.5 Litter support. Each litter support shall have a cast aluminum alloy support cup with a vertical leg. The cup shall have inside dimensions of 3-1/4 inches long by 1-3/4 inches wide by 3/8 inch deep. The leg under the cup shall be 1 inch wide by 1/4 inch thick by a minimum of 2-1/4 inches long. The cup shall support the stirrup of a standard litter and the leg shall slide into the litter support socket (see 3.2.6.7).



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:10Z
Check the source to verify that this is the current version before use.

3.2.6.6 Wristlet with adjustable holder. Each wristlet shall consist of a cast aluminum alloy combination side rail clamp and wristlet support, and a cotton webbing belt. The top of the wristlet support shall sit approximately 3-1/2 inches above the center line of the side rail. The siderail clamp shall slide onto the main frame siderail (see 3.2.3) and be capable of being secured with a locking device at any point along the length of the siderail. The wristlet support shall allow the cotton webbing belt to pass through an opening in it in order to secure the belt to the support. The belt shall be class 1, 1-1/2 inch wide cotton webbing (see 3.1.5). The belt buckle shall be type III, style 1, 1-1/2 inch in accordance with MIL-B-543. The end clip shall be type I, class 1, 1-1/2 inch in accordance with MIL-C-496. The finished length of the belt shall be 22 inches plus or minus 1/4 inch. As an alternate to the brass specified for the belt buckle and end clip, steel, treated to resist corrosion, may be utilized. Wristlets shall be supplied in pairs.

3.2.6.7 Support sockets. Each table shall be supplied with six cast aluminum alloy accessory support sockets. All sockets shall slide along the siderails (see 3.2.3) and shall be capable of being secured at any point along the siderail length. All sockets shall mate freely with, and securely hold, their respective accessories. Two each of the following sockets shall be supplied:

3.2.6.7.1 Dual function socket. Each dual function socket shall have a positive locking, toothed, angulating section with an opening to accommodate the anesthetist's screen (see 3.2.6.2) and a non-angulating vertical section with an opening to accommodate either a shoulder support (see 3.3.6.1), or a litter support (see 3.3.6.5). The socket shall have three locking devices; one to secure the socket to the siderail, one to secure the angulating section to the non-angulating section, and one to secure the anesthetist's screen in the angulating section.

3.2.6.7.2 Leg support socket. Each leg support socket shall have a non-angulating vertical opening to accommodate a leg support (see 3.3.6.3). The socket shall have a locking device.

3.2.6.7.3 Footrest and litter support socket. Each footrest and litter support socket shall have a non-angulating vertical opening to accommodate either one leg of the footrest (see 3.3.6.4), or a litter support (see 3.3.6.5). The socket shall have a locking device.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:10Z
Check the source to verify that this is the current version before use.

3.2.7 Carrying case. The carrying case shall be constructed as specified on DPSC Drawing 1796, sheet 1 of 4. The panels shall be plywood (see 3.1.3), covered on each side with an aluminum alloy sheet. The aluminum alloy sheet shall be 0.025 inch thick and shall be bonded to the plywood with waterproof glue. Aluminum shapes, bars, and strips for the carrying case shall be fabricated from alloy 6061-T61 (see 3.1.1.2). Each securing strap shall be type III, class 4, 1 inch wide cotton webbing (see 3.1.5). The buckle shall be type II, style 1 or 4, class 1, 1 inch wide, in accordance with MIL-B-543. The strap end clip shall be type I, class I, 1 inch wide in accordance with MIL-C-496. Rivets shall be zinc plated steel (see 3.4). Other hardware shall be AISI series 300 corrosion-resisting steel (see 3.5). The carrying case shall serve as a storage and corrosion-resisting carrying case when the operating table is not in use, and as a rigid instrument and supply table when the operating table is in use.

3.2.8 Pouches. The pouch for the leg support brackets and the pouch for the accessories shall be constructed as specified on DPSC Drawing 1796, sheet 2 of 4. The pouches shall be fabricated from cotton duck (see 3.1.4). Binding for the pouches shall be cotton tape (see 3.1.6). The thread shall be either cotton or polyester (see 3.1.7). All stitching shall be machine stitching, type 301, in accordance with FED-STD-751. There shall be a minimum of seven stitches to the inch. When machine stitching is not a continuous thread, it shall be backstitched not less than 1 inch at each break. All stitching shall be backstitched at the ends to prevent raveling, except where the ends are turned under in a hem, or held down by other stitching. Each pouch shall be furnished with a brass snap fastener for securing the cover in the closed position. The fastener shall be style 2, finish 2, in accordance with MIL-F-10884. All fasteners shall be securely clinched to the pouch material.

3.3 Gears. The gears for the adjusting mechanism (see 3.2.5) shall be fabricated using the following materials:

3.3.1 Rack material. Rack material shall be corrosion-resisting steel (see 3.1.2.1).

3.3.2 Worm material. Worms shall be formed from heat-treated corrosion-resisting steel, ground and polished (see 3.1.2.1).



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:10Z
Check the source to verify that this is the current version before use.

3.3.3 Worm gears. Worm gears shall be either steel, or bronze, with machine cut teeth.

3.4 Rivets. All rivets used in the fabrication of the carrying case shall be zinc plated steel. All rivets used in the fabrication of the operating table shall be corrosion-resisting steel.

3.5 Hardware. Hardware such as bolts, nuts, handcrank pin, and thumbscrews, used in the assembly of the operating table, shall be fabricated from AISI series 300 corrosion-resisting steel. Hardware used in the assembly of the carrying case, such as handles, hinges, braces, and locking mechanisms, shall be fabricated from heavily zinc plated carbon steel.

3.6 Rings. The rings for use with the leg supports (see 3.2.6.3) shall have a 1-1/4 inches plus or minus 1/8 inch inside diameter, and shall be formed from either 3/16 inch diameter brass or steel wire. The wire ends shall be brazed together and ground smooth. The rings shall be plated with a minimum 0.0004 inch thickness of nickel, zinc, or chrome. A 1-1/4 inches plus or minus 1/8 inch inside diameter chrome, zinc, or nickel plated steel split key ring shall also be acceptable. D-rings shall be a commercial 1 inch (across flat) size, and shall be formed from either 3/16 inch diameter brass or steel wire, or cast from bronze. The ends of the wire shall be joined by brazing and ground smooth. All D-rings shall be plated with a minimum 0.0004 inch thickness of chrome, zinc, or nickel.

3.7 Swivel eye bolt. The swivel eye bolt shall be fabricated from brass and shall have a 4 inches plus or minus 1/2 inch overall length, with an eye of 3/4 inch inside diameter. The finished eye bolt shall be plated with a minimum 0.0004 inch thickness of chrome, zinc, or nickel.

3.8 Finish. All aluminum alloy parts shall be clear anodized by the sulfuric acid (type II) process, hot water rinsed in accordance with Mil-A-8625.

3.9 Identification marking. Each operating table shall be either impression stamped, or provided with a metal nameplate, with the following information:

* Contract or purchase order number
* Name or registered trademark of the manufacturer
* National Stock Number of the table



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:10Z
Check the source to verify that this is the current version before use.

Each carrying case shall be marked in black, using a silk screen. All identification marking shall be permanent and legible.

3.10 Weight. The maximum weight of the table, including all accessories, pouches, and carrying case, shall not exceed 155 pounds.

3.11 Service data. The contractor shall furnish, with each unit, two copies each of a technical manual which contains complete instructions for installation, operation, and maintenance, and lists of component parts. The manual shall comply with the requirements of MIL-M-7298. As soon as practicable after award of the contract or purchase order, the contractor shall furnish two preliminary copies of his proposed technical manual to the contracting officer for review and comment. The preliminary manuals shall be submitted sufficiently early to permit adequate review by the contracting officer (based on a maximum period of 30 days for such review), and to allow incorporation by the contractor of any required revisions and corrections in the final manual, without delaying delivery under the contract or purchase order. In addition, the contractor shall furnish 15 copies of the approved technical manual to the contracting officer for distribution. Should the suppler have furnished acceptable manuals for the identical item within two years preceding the date of the contract, and propose to furnish manuals identical with those previously accepted, then the requirements for reviews of drafts and submission of 15 additional copies may be waived upon request to the contracting officer. Waiver of submission of drafts and submission of 15 additional copies shall be at the option of the Government.

3.12 Workmanship. Operating tables and carrying cases shall be free from defects which detract from their appearance or impair their serviceability. All burrs and sharp edges shall be removed.

3.12.1 Welding. Welding shall be durable, sound in every detail, and in accordance with the latest practices recommended by the American Welding Society. All welds shall be produced by qualified welding operators with satisfactory welding equipment, using satisfactory filler metal, or electrode where applicable, to provide the best possible weld. Surfaces to be welded shall be free from grease, oil, oxides, or any other foreign matter. Any cleaning method which will remove these substances and leave the surface clean, is satisfactory. All weld joints shall blend into the adjacent metal in gradual, smooth curves. Beads shall



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:10Z
Check the source to verify that this is the current version before use.

be smooth and free of slag, excessive undercut, or excessive splatter. Sufficient metal shall be added to provide a suitable fillet or reinforcement, unless otherwise specified. Welds shall not be oxidized. All welds shall be continuous, with no visible voids, pin holes, or cracks.

## 4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements (examinations and tests) as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in this specification, where such inspections are deemed necessary, to ensure supplies and services conform to prescribed requirements.

4.1.1 Responsibility for compliance. All items shall meet all requirements of sections 3 and 5. The inspection set forth in this specification shall become a part of the contractor's overall inspection system or quality program. The absence of any inspection requirements in the specification shall not relieve the contractor of the responsibility of ensuring that all products or supplies submitted to the Government for acceptance comply with all requirements of the contract. Sampling inspection, as part of manufacturing operations, is an acceptable practice to ascertain conformance to requirements, however, this does not authorize submission of known defective material, either indicated or actual, nor does it commit the Government to accept defective material.

4.1.2 Records. Records of examinations and tests performed by, or for, the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract, and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.3 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:10Z
Check the source to verify that this is the current version before use.

4.1.4 Certificates of quality. Certificates of quality supplied by the manufacturer of the component or material may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the National Stock Number (NSN), the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and table I. The unit of product for examination purposes shall be one complete operating table and carrying case.

TABLE I. Sampling for examination.

| | Inspection level | AQL (percent defective) |
|---|---|---|
| For visual examination: | | |
| Major defects | II | 1.0 |
| Minor defects | II | 4.0 |
| For dimensional examination | S-1 | 2.5 |

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 for an acceptance number of zero. The inspection level for testing the operating table shall be 100 percent. The inspection level for component material testing shall be S-1 with not less than three samples tested regardless of lot size.

4.3 Examination. The operating tables shall be examined for defects including, but not limited to, those listed in table II.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:10Z
Check the source to verify that this is the current version before use.

4.3.1 Classification of defects. Examination shall be conducted in accordance with the classification of defects listed in table II below. Examination shall not be restricted to the classified possible defects.

TABLE II. Classification of defects.

| Category | Defects |
|---|---|
| Major | |
| 101 | Any component missing |
| 102 | Components not free-moving |
| 103 | Table and accessories not capable of being placed in operative position |
| 104 | Metal components not free of burrs and rough edges |
| Minor | |
| 201 | Finished unit not clean and not free of foreign matter |
| 202 | Identification marking not complete, correct, legible, and permanent |
| 203 | Swivel eyebolt not free-moving |

4.3.2 Dimensional examination. The tables shall be examined for defects in dimensions. Any dimension not within the tolerances specified herein shall be classified as a defect.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Chemical analysis of the metals. Four ounces of each lot of each metal used in the construction of the tables shall be analyzed in conformance with FED-STD-151.

4.4.2 Material Testing.

4.4.2.1 Cotton duck. The finished cotton duck shall be tested in accordance with the test procedures of CCC-D-950, except that the performance of the tests for mildew-inhibitor



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:10Z
Check the source to verify that this is the current version before use.

content and mildew-resistance shall be conducted under the conditions specified in 4.4.2.5.

4.4.2.2 Cotton webbing. The cotton webbing shall be tested in accordance with the test procedures of MIL-W-530, except the the performance of the tests for mildew-inhibitor content and mildew-resistance shall be conducted under the conditions specified in 4.4.2.5.

4.4.2.3 Cotton tape. The cotton tape shall be tested in accordance with the test procedures of MIL-T-43566, except that the performance of the tests for mildew inhibitor content and mildew resistance shall be conducted under the conditions specified in 4.4.2.5.

4.4.2.4 Cotton thread. The finished cotton thread shall be tested in accordance with the test procedures of MIL-T-3530, except that the performance of the tests for mildew content and mildew resistance shall be conducted under the conditions specified in 4.4.2.5.

4.4.2.5 Mildew content and mildew resistance.

a. If the material conforms to the requirements for mildew inhibitor content, the mildew resistance tests shall not be performed, and the material shall be considered as complying with the mildew resistance requirements.

b. If the mildew inhibitor content exceeds the maximum requirement, the material shall be rejected.

c. If the mildew inhibitor content is found to be below the minimum requirement, the material may be accepted with respect to mildew resistance provided it passes the mildew resistance test, method 5762 of FED-STD-191. When the mildew resistance test is performed, the leaching process shall be omitted.

4.4.3 Table operation testing.

4.4.3.1 Tabletop height test. The tabletop shall be placed in the horizontal position in the highest height position. The height of the tabletop shall then be measured. A tabletop height of less than 37-3/4 inches, or more than 38-1/4 inches, shall constitute a failure. The table shall then be placed in the horizontal position in the lowest height position. A tabletop



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:10Z
Check the source to verify that this is the current version before use.

height of less than 33-1/4 inches, or more than 33-3/4 inches, shall constitute a failure.

4.4.3.2 Trendelenberg test. After putting the tabletop in the horizontal position in the highest height position,, the tabletop shall be tilted with the head section 50 degrees below the horizontal. After returning the tabletop to the horizontal position, the table shall be then placed in the next to lowest height position. The table shall then be tilted, using the adjusting mechanism, with the head section 50 degrees below the horizontal. Inability for the table to achieve 50 degrees trendelenberg in either of the above heights shall constitute a failure.

4.4.3.3 Reverse Trendelenberg test. After placing the tabletop in the horizontal position at the highest height position, the table shall be tilted with the foot section 50 degrees below the horizontal. After returning the tabletop to the horizontal position, the table shall then be placed in the next to lowest height position. The table shall then be tilted with the head section 50 degrees below the horizontal. Inability for the table to achieve 50 degrees reverse trendelenberg in either of the above heights shall constitute a failure.

4.4.3.4 Back section adjustment test. The back section shall be placed in the horizontal position with respect to the main frame. The back section shall then be placed in the following angular positions with respect to the main frame using the ratchet and pawl arrangement:

* 10 degrees
* 25 degrees
* 45 degrees
* 65 degrees

The inability to achieve any of the above angular positions in addition to the horizontal position, within plus or minus 2 degrees, shall constitute a failure.

4.4.3.5 Leg section adjustment test. The leg section shall be placed in the horizontal position with respect to the main frame. The leg section shall then be placed in the following angular positions with respect to the main frame using the



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:10Z
Check the source to verify that this is the current version before use.

ratchet and pawl arrangement:

* 15 degrees

* 30 degrees

* 45 degrees

The inability to achieve any of the above angular positions as well as the horizontal position, within plus or minus 2 degrees, shall constitute a failure.

4.4.4 Table load testing.

4.4.4.1 Load test (table). With the tabletop in its normal horizontal position, sand filled bags totaling 300 pounds in weight shall be equally distributed and secured to the top. The table shall be adjusted to all possible positions a minimum of five times. At the end of the adjustments, the table shall be examined for permanent distortion. Inabililty for the table to be adjusted to all possible positions with the weight in place, or any signs of permanent distortion, shall constitute a failure.

4.4.4.2 Load test (back section). The back section shall not collapse or move in all raised positions when a load of 125 pounds is applied perpendicular to the plane of the section and 12 inches from the head end.

4.4.4.3 Load test (stirrup). Each stirrup shall be extended to its full length and height and locked in position. The stirrups shall be capable of withstanding a 50 pound force applied perpendicular to the support section of the stirrup. After removal of the weight, the stirrups and supporting system shall then be examined for permanent distortion. Any sign of distortion shall constitute a failure.

4.5 Inspection of packaging. The sampling and inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:10Z
Check the source to verify that this is the current version before use.

5. PACKAGING

5.1 Preservation. Preservation shall be level A or C, as specified (see 6.2).

5.1.1 Level A.

5.1.1.1 Preservative applications. Preservatives shall not be used.

5.1.1.2 Unit package. Each table shall be packaged in a solid fiberboard or double-faced corrugated fiberboard carton, of appropriate size and design constructed in accordance with PPP-B-636, Type CF, class domestic. Each carton shall be adequately secured to prevent accidental opening. Closure shall be as specified in the applicable box specification.

5.1.2 Level C.

5.1.2.1 Unit package. Preservation for the table shall conform to the requirements of MIL-STD-794 for level C.

5.2 Packing. Packing shall be level A, B, or C, as specified (see 6.2).

5.2.1 Level A. Each table packaged as specified in 5.1.1.2 shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, overseas type, PPP-B-621, class 2, or PPP-B-636, class weather-resistant. Closure and strapping shall be as specified in the applicable box specification except that steel strapping shall not be acceptable for fiberboard boxes.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall be as specified in the liner specification. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed as specified in PPP-B-636.

5.2.2 Level B. Each table packaged as specified in 5.1.1.2 shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, domestic, PPP-B-621, class 1, or PPP-B-636, class domestic. Closure shall be as specified in the applicable box specification.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:10Z
Check the source to verify that this is the current version before use.

5.2.3 Level C. Each table shall be packed in accordance with the requirements of MIL-STD-794 for level C.

5.2.4 Packaging variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.2.5 Unitized loads. Unitized loads, commensurate with level of packing specified in the contract or order, shall be used whenever total quantities for shipment to one destination exceed 250 pounds or 20 cubic feet. Loads shall be unitized on pallets conforming to MIL-STD-147. Quantities less than 20 cubic feet need not be unitized. Pallet loads including the pallet shall not exceed 43 inches in length, 52 inches in width, and 54 inches in height. Pallet shall conform to Type IV or Type V.

5.2.5.1 Levels A and B. The table packed as specified in 5.2.1 and 5.2.2 shall be unitized on pallets as specified in 5.2.5.

5.2.5.2 Level C. The table packed as specified in 5.2.3 shall be unitized as specified in MIL-STD-794.

5.3 Marking.

5.3.1 Levels A, B, and C. Each unit pack, intermediate package, exterior container and unitized load shall be marked as specified in MIL-STD-129.

5.4 General.

5.4.1 Exterior container. Exterior container (see 5.2.1, 5.2.2, and 5.2.3) shall be of minimum tare and cube consistent with the protection required and shall contain equal quantities of identical stock numbered items to the greatest extent practicable.

5.4.2 Packaging inspection. The inspection of these packaging requirements shall be in accordance with paragraph 4.5.

6. NOTES

(This section contains information of a general or explanatory



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:10Z
Check the source to verify that this is the current version before use.

nature that may be helpful, but is not mandatory.)

6.1 Intended use. The operating table is intended for use by the surgical service for field operations.

6.2 Acquisition requirements.

a. Title, number, and date of the specification.

b. Issue of DODISS to be cited in the solicitation and, if required, the specific issue of individual documents referenced (see 2.1).

c. Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

d. National Stock Number.

6.3 Subject term (key word) listing.

| | |
|---|---|
| Clinic | Reverse Trendelenberg |
| Hospital | Surgery |
| M.U.S.T. | Surgical table |
| Operating room | Trendelenberg |
| Operation | |

6.4 National Stock Number (NSN) coverage. This specification covers the following item shown in the Federal Supply Catalog.

| National Stock Number | Item Identification |
|---|---|
| 6530-00-709-8175 | TABLE, OPERATING, FIELD |

6.5 This specification does not cover all types, classes, grades,or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the type which ls normally procured to meet requirements.

6.6 Changes from previous issue. Marginal notations are not used in this revision to identify modifications, additions, corrections, and deletions from the previous issue due to the extensiveness of the changes.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:10Z
Check the source to verify that this is the current version before use.

Custodians:

Army - MD
Navy - MS
Air Force - 03

Civil Agency Coordinating Activities:

FDA-MPQAS
USPHS

Review Activities:

Army - MD
Navy - MS
Air Force - 03

Preparing activity:

DOD-MB

Agent:

DLA-DM

Project Number: 6530-2149

Location: TBLMIL13.A1



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:10Z
Check the source to verify that this is the current version before use.

# STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

*(See Instructions – Reverse Side)*

| 1. DOCUMENT NUMBER | 2. DOCUMENT TITLE |
|---|---|
| MIL-T-37637A | TABLE, OPERATING, FIELD |

| 3a. NAME OF SUBMITTING ORGANIZATION | 4. TYPE OF ORGANIZATION *(Mark one)* |
|---|---|
| | ☐ VENDOR |
| | ☐ USER |
| b. ADDRESS *(Street, City, State, ZIP Code)* | ☐ MANUFACTURER |
| | ☐ OTHER *(Specify)*: |

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation.

6. REMARKS

| 7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional | b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional |
|---|---|
| c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional | 8. DATE OF SUBMISSION *(YYMMDD)* |

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:10Z
Check this source to verify that this is the current version before use.