[INCH-POUND]
A-A-59198
March 31, 1998
SUPERSEDING
GG-P-725B
July 14, 1978

COMMERCIAL ITEM DESCRIPTION

PSYCHROMETERS

The General Services Administration has authorized the use of this commercial item description for all federal agencies.

1. SCOPE. This commercial item description (CID) covers the requirements for psychrometers used for determining relative humidity.

2. CLASSIFICATION. This CID covers psychrometers of the following types. The type to be furnished shall be as specified (see 7.2).

2.1 Type: Type I - Wall mounted
Type II - Sling
Type III - Pocket Sling
Type IV - Fan-ventilated, battery-operated
Type V - Hand-held, battery-operated, digital

3. SALIENT CHARACTERISTICS

3.1 General requirements. Unless otherwise specified (see 7.2), the thermometers shall indicate degrees Fahrenheit (°F) with a minimum range of 20 °F to 120 °F to an accuracy of ±1°. Unless otherwise specified (see 7.2), the indicating liquid of the thermometers shall be organic based. A psychometric chart shall be provided with each psychrometer, except Type V. When required (see 7.2), a National Institute of Standards and Technology Certification shall be provided. Different temperature ranges, indicating scales and accuracy shall be provided as specified (see 7.2).

Beneficial comments, recommendations, additions, deletions, clarifications, etc., and any data which may improve this document, should be sent to: Defense Supply Center Richmond, ATTN: DSCR-VBD, 8000 Jefferson Davis Highway, Richmond, VA 23297-5610.

AMSC N/A FSC 6685
DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:05Z
Check the source to verify that this is the current version before use.

3.2 Type I. The indicating scale of the thermometers shall be divided into increments of 2°.

3.3 Type II. The indicating scale of the thermometers shall be divided into increments of 1°.

3.4 Type III. The indicating scale of the thermometers shall be divided into increments of 1°. A carrying case for the psychrometer shall be provided.

3.5 Type IV. The indicating scale of the thermometers shall be divided into increments of 1°.

3.6 Type V. The unit shall display the temperature and humidity digitally with a resolution of 0.1 or greater. The psychrometer shall be accurate to ±1°F and ±3% relative humidity.

4. REGULATORY REQUIREMENTS

4.1 Recovered materials. The offeror/contractor is encouraged to use recovered materials to the maximum extent practicable, in accordance with paragraph 23.403 of the Federal Acquisition Regulation (FAR).

4.2 Environmental protection. The item shall meet all applicable Environmental Protection Agency restrictions in effect on the date of the contract. These regulations apply to the emission of materials hazardous to the environment or the user's health and shall be met during the manufacturing, service, transportation, storage, and operation/use of the item.

5. QUALITY ASSURANCE PROVISIONS

5.1 Product conformance. The products provided shall meet the salient characteristics of this commercial item description, conform to the producer's own drawings, specifications, standards, and quality assurance practices, and be the same product offered for sale in the commercial market. The Government reserves the right to require proof of such conformance.

6. PACKAGING. Preservation, packing, and marking shall be as specified (see 7.2).

7. NOTES

7.1 Source of documents.

7.1.1 Government documents. Copies of Federal documents may be obtained from the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20402.

7.2 Ordering data. Acquisition documents must specify the following:

a. Title, number, and date of this document.
b. Type of psychrometer required (see 2)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:05Z
Check the source to verify that this is the current version before use.

c. Thermometers in degrees Celsius, if required (see 3.1).
d. Non-organic based indicating liquid, if required (see 3.1).
e. National Institute of Standards and Technology Certification, if required (see 3.1).
f. Temperature range, indicating scale and accuracy of psychrometer, if different (see 3.1).
g. Preservation, packing, and marking (see 6).

7.3 Part identification number (PIN). The following part identification numbering procedure is for government purposes and does not constitute a requirement for the contractor.

A-A-59198 -1

Type: 1 - Type I, Wall Mounted
2 - Type II, Sling
3 - Type III, Pocket Sling
4 - Type IV, Fan-ventilated, Battery-operated
5 - Type V, Digital, Hand-held, Battery-operated

CID number

MILITARY INTERESTS:

Custodians
Army - CR
Navy - SH

PREPARING ACTIVITY:
DLA - GS

(Project 6685-0943)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:05Z
Check the source to verify that this is the current version before use.