A-A-50753
2 March 1987

COMMERCIAL ITEM DESCRIPTION

GENERATOR, VARIABLE FREQUENCY

The General Service Administration has authorized the use of this commercial item description.

Scope. Generator designed for field, shop or laboratory testing and calibration of electro mechanical and solid state instantaneous and time delay under and over frequency relays. It is a portable test set used on transducers, recorders and synchronizers.

Applicable documents. The following specification forms a part of this commercial item description to the extent specified herein. Unless otherwise specified, the issue of this document shall be that listed in the issue of the Department of Defense Index of Specifications and Standards (DODISS) and Supplement thereto cited in the solicitation.

SPECIFICATION

MIL-T-28800 Test Equipment for use with Electrical and Electronic Equipment General Specification for.

(Copies of specification required by contractors in connection with specific acquisition functions should be obtained from the contracting activity or as directed by the contracting activity).

Salient characteristics.

1. Input voltage. 115, single phase.

2. Input frequency: 60 hertz.

3 Output frequency. Two precision potentiometers with counter dials are incorporated to provide precise adjustment of output frequency. One is used to set the "normal" frequency and the other to set the "fault" frequency.

   a. Range: 50-70 hertz on 60 hertz input units.

   b. Stability. (Over a 15-minute span) within 0.1 hertz after 5 minutes operation and within 0.05 hertz after 15 minutes warmup.

4. Output potential: 0.125 volts, single phase.

5. Capacity: 35 VA at 125 volts.

AMSC N/A FSC 662

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:46Z
Check the source to verify that this is the current version before use.

6. Voltage regulation: Not more than ±5 volts at any combination of output load and output frequency.

7. Distortion: 1% maximum departure from sine wave at any frequency or load.

8. Instrumentation: A solid digital frequency meter designed specifically for power system applications is incorporated to precisely measure the output frequency. It may also be switched to measure input line frequency.

a. Display: Non-flickering 0.55" (1.4 cm) numerals.

b. Range: 50.00 - 70.00 hz on 60 hz input units.

c. Resolution: 0.01 hertz.

d. Accuracy. Better than 0.02 hertz.

9. Protection: Overload and short circuit protection is provided on input and output circuits.

10. Enclosure. The unit is in a rugged, heavy-duty formica case with removable hinged cover and carrying handles on both top and sides. A compartment is provided for storage of test leads.

Certification. The contractor shall certify that the product offered meets the salient characteristics of this description, and that the product conforms to the producer's own drawings, specifications, standards, and quality assurance practices, and is the same product offered for sale in the commercial marketplace. Small businesses that cannot offer products with established commercial market acceptability, may compete by submission of bid samples. (The Government reserves the rights to require proof of such conformance before first delivery and thereafter as may be otherwise provided for under the provisions of the contract).

Packaging, packing, and marking. The items shall be packaged in accordance with normal commercial practice and packed to assure acceptance by common carrier and provide product protection against loss and damage during multiple shipments, handling, and storage. The shipping container shall be in compliance with national motor freight classification and uniform freight classification. Shipping container shall be marked as specified in the contract or order.

System safety: Shall be in accordance with MIL-T-28800, paragraph titled Safety.

Design: MIL-T-28800, type II, class 5, style E, color R.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:46Z
Check the source to verify that this is the current version before use.

Known acceptable product.

Manufacturer	Model	NSN
Multi-Amp Corp.	FG-50-DM	6625-01-059-2443

Custodian:
Air Force - 99
Navy - EC
DLA - ES

Preparing activity:
Air Force - 82

Reviewing activity:
Army - CR

(Project 6625-F846)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:46Z
Check the source to verify that this is the current version before use.