A-A-50756
1 OCT 1986

COMMERCIAL ITEM DESCRIPTION

SWEEP GENERATOR

The General Services Administration has authorized the use of this commercial item description.

SCOPE Application is AN/EPS 85.

APPLICABLE DOCUMENT. MIL-T-28800 Test Equipment for use with Electrical and Electronic Equipment, General Specification for.

SALIENT CHARACTERISTICS.

1. Frequency range. 330 - 1010 MHZ.
2. Center frequency range. 375 - 1000 MHZ.
3. Output.
   a. Sweep and modulation (MOD) sweep.
      (a) Attenuator Out 14 VRMS (4 watts)
      (b) Attenuator In 10 VRMS (2 watts).
   b. CW and MOD CW.
      (a) Attenuator Out 10 VRMS (2 watts)
      (b) Attenuator In 7 VRMS (1 watt).
4. Impedance. 50 OHMS.
5. Flatness. ±0.4 DB over maximum sweep width
   ±0.5 DB over full frequency range.
6. Attenuator (Turret, 3-62DB in 1 DB steps includes a 3 DB pad in coaxial switch assembly).
7. Display linearity ±2% nominal
8. Sweep rate Line frequency (50/60-2
9. Zero baseline. Oscillator off during sweep retrace.
10. Scope horizontal output = 14V peak-to-peak.
11. Source VSWR. 1.2 1

FSC 6625

DISTRIBUTION STATEMENT A. Approved for public release; distribution unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:46Z
Check the source to verify that this is the current version before use.

12. Modulation. Internal 1 KHZ oscillator (may be used to modulate the sweep and CW).

13. Input power. 115/230 VAC, 50/60 HZ (200 watts)

14 Level adjust. RF level control used to set unattenuated output at specified level.

15 Meter = Indicates the actual RF output

16 Sweep width = Adjusts the sweep width by varying the voltage supplied to the sweep

17. Marker size. Adjusts the amplitude of the plug-in and external markers

18 Fixed markers Switches B+ to the plug-in markers

19 External marker An external signal source can be inserted at this connector to produce a birdy marker on the RF output

20 Marker adder

a In = The demodulated signal from the device being tested is fed into this connector so that the marker signal can be superimposed.

b Out = This connector should be connected to the vertical input terminals of the scope for viewing the demodulated signal and markers

21 Scope horizontal = Oscilloscope horizontal drive signal is preset at this connector which should be connected to the scope horizontal input terminals.

22. Function switch = selects the internal control circuitry for the following functions

a AC power on-off

b Sweep

c CW.

d Mod sweep

e Stand-by

f Mod CW

23. Design MIL-T-28800, Type II, Class 3, Style F, Color R

24. Construction Bench, system safety per MIL-T-28800, para 3.2.3.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:46Z
Check the source to verify that this is the current version before use.

CERTIFICATION. The contractor shall certify that the product offered meets e salient characteristics of this description. A bid sample marked "Not for Stock" will be tested at San Antonio Air Logistics Center for first article acceptance of commercial modified off-the-shelf hardware.

PACKAGING, PACKING AND MARKING. The items shall be packaged in accordance with normal commercial practice and packed to assure acceptance by common carrier and provide product protection against loss and damage during multiple shipments, handling, and storage. The shipping container shall be in compliance with National Motor Freight Classification and Uniform Freight Classification. Shipping containers shall be marked as specified in the contract or order.

KNOWN ACCEPTANCE PRODUCT(S). The following product(s) are known to meet the requirements of this CID.

Manufacturer	Model	NSN
Telonic (no longer made)	PD-8	6625-00-988-2774

UL Standard 1236 is available from Underwriters' Laboratories (UL), Inc., 1285 Walt Whitman Road, Melville, Long Island, New York 11747.

Custodians:
Air Force - 0C
Navy - EC
Army - EF

Preparing activity
Air Force - 82

(Project 6625-0649)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:46Z
Check the source to verify that this is the current version before use.