COMMERCIAL ITEM DESCRIPTION

PULSE GENERATOR

The General Services Administration has authorized the use of this commerical item description.

Scope. This is a double pulse generator. The first pulse is fixed at approximately zero delay. The second pulse is delayable from 0.5 to 10,000 microseconds with respect to the leading edge of the first pulse. Frequency range is 2 Hz to 2 MHz. It is used on military end items.

Applicable documents. MIL-T-28800

Salient characteristics.

Oscillator Repetition Rate. The internal oscillator is continuously variable from 2 cycles per second to 2 megacycles per second in six ranges.

2 cps	-	20 cps
20 cps	-	200 cps
200 cps	-	2 kc
2 kc	-	20 kc
20 kc	-	200 kc
200 kc	-	2 mc

External Triggering. In addition to the internal oscillator, it may also be triggered from an external source by a trigger exhibiting the following characteristics.

1.	Amplitude	20 V., minimum
2.	Repetition Rate	2 mc, maximum
3.	Rise Time	1.0 μsec, or less
4.	Polarity	Positive or negative

5. Note: At lower repetition rates and/or faster rise times, the instrument triggers with a much lower amplitude pulse. The fast rise time quality of the sync. pulse out allows it to be triggered from the sync. pulse at all repetition rates.

Pulse Delay. Main pulse delay is continuously variable, with respect to the sync. pulse out, from 0.0 microseconds to 10,000 microseconds in five ranges.

FSC 6625

DISTRIBUTION STATEMENT A. Approved for public release; distribution unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

0.0 μsec.	-	1.0 μsec.
1.0 μsec.	-	10 μsec.
10 μsec.	-	100 μsec.
100 μsec.	-	1,000 μsec.
1,000 μsec.	-	10,000 μsec.

Note: In the double pulse mode, the first pulse is fixed at approximately zero delay. The second pulse is delayable from 0.5 to 10,000 microseconds with respect to the leading edge of the first pulse.

Pulse Width.

1. Main pulse width is continuously variable from 0.05 microseconds to 10,000 microseconds in five ranges.

0.05 μsec.	-	1.0 μsec.
1.0 μsec.	-	10 μsec.
10 μsec.	-	100 μsec.
100 μsec.	-	1,000 μsec.
1.000 μsec.	-	10,000 μsec.

2. Note: In the double pulse mode, both output pulses are controlled simultaneously in width.

Main Pulse. The following are the specifications of the main output pulse.

1. Amplitude 50V., minimum, into a 50Ω load

2. Polarity Positive or negative, but not simultaneously.

3. Rise Time 0.015 microseconds

4. Fall Time 0.015 microseconds

5. Note: The rise and fall time of the output pulse may be independently degraded to approximately 1 microsecond. In the double pulse mode, rise times degrade simultaneously and fall times degrade simultaneously.

6. Overshoot and Undershoot 5% maximum, except when pulse is attenuated more than 30db. with the 10db. step attenuator, 30% maximum, at 50db. down.

7. Droop:

0-100 μsec. wide pulse	Negligible
100-1,000 μsec. wide pulse	15%, maximum
1,000-10,000 μsec. wide pulse	25%, maximum

8. Duty Factor. 30%, maximum, at full output amplitude. Automatic limiting circuit reduces the output amplitude of duty factors greater than 10%. The unit may be safely operated to approximately a 70% duty factor. Output amplitude is approximately, 35 volts at 50% duty factor and approximately 20 volts at 70%. In the double pulse mode, duty factor applies to the sum of both pulse widths.

9. Attenuation 0 to 60db, continuously variable.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

10. Output Impendance - 500.

Sync Pulse Out. The following sync. pulse is available at a front panel connector.

1.	Amplitude	10V., minimum.
2.	Width	0.05 μsec., minumum
3.	Rise Time	0.03 μSec., maximum
4.	Polarity	Positive

Miscellaneous Specifications. The following are miscellaneous specifications of the instrument.

POWER REQUIRED 115V., 4 amps., 50-60 cps

Certification: Bid samples from the bidders will be tested at San Antonio ALC for compliance with the salient characteristics of this Commercial Item Description.

Packaging, Packing, and Marking: The preservation, packaging, packing, labeling, and marking shall be as specified in the contract or order.

Intended Use. General Purpose Air Force applications.

Known Acceptable Products.

Manufacturer	Model	NSN
Dynasciences Copr. Instrument Systems Div, Chatsworth, CA (No longer made) MFC 06692	B7F5	6625 00 908 9119

Custodians:

Air Force - 99
Navy - EC

Reviewers: Army - CR

Preparing Activity:

Air Force - 82

(Project: 6625-0654)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.