COMMERCIAL ITEM DESCRIPTION

WORK STATION, LAMINAR FLOW

The General Services Administration
has authorized the use of this
commercial item description

Scope: This commercial item description covers a Console Type, Horizontal Laminar AirFlow, Class 100 Cleanliness Level Clean Work Station.

Applicable documents. The following documents, of the issues in effect on date of invitation for bids or request for proposal, form a part of this commercial item description to the extent specified herein:

FED-STD-209	Clean Room and Work Station Requirements, Controlled Environment
T.O. 00-25-203	Contamination Control of Aerospace Facilities, U.S. Air Force

NATIONAL BOARD OF FIRE UNDERWRITERS

Pamphlet 70	National Electrical code (Electrical wiring and Apparatus)

(Application for copies should be addressed to the National Fire Protection Association, 60 Battery March Street, Boston MA 02110)

NATIONAL ELECTRICAL MANUFACTURER'S ASSOCIATION STANDARD

MG- 1	Standards for Motors and Generators

(Application for copies should be addressed to the National Electrical Manufacturer's Association, Engineering Department, 155 East 44th Street, New York, NY 10017).

Salient characteristics:

1.	Construction	- 16 gauge steel, perforated anodized aluminum filter shields.
2.	Finish	Baked acrylic or epoxy type enamel, color white.
3.	End panels	- Clear acrylic plastic

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:47Z
Check the source to verify that this is the current version before use.

4.	Work surface	- Melamine resin equal to Formica or Micarta (registered trade marks). Color blue, light gray, frost green or white.
5.	Lighting	- 150 foot candles min. at work surface, color whites fluorescent type, self-contained Underwriter Laboratories (U.L.) approved.
6.	Prefilters	- Throwaway replacement type, removable from front.
7.	Final filters	- HEPA (High Efficiency Particulate Air) 99.99% or more effective on particles 0.3 micron and larger 100% leak probed. Aluminum. separators removable from bear.
8.	Filter mounting	- Spring tension compression seal.
9.	Blower	- Direct drive type. To maintain minimum average 90 cubic feet per minute. Variably controlled.
10.	Blower control	- Solid state. Mounted in blower housing.
11.	Electrical	- 115v/60 Hz, single phase, shall meet NFPA 70. All components shall be U.L. approved.
12.	Power cord	- Six foot three wire power cord of adequate size to operate fan motors and lights shall exit from back of cabinet. Termination shall be standard, three prong plug.
13.	Motor	- Shall meet NEMA Standard MG-1.
14.	Work area	- Minimum 46.5 inches wide by 22 inches deep by 22 inches high. Work surface 30 to 32 inches from floor.
15.	Performance	- Dust count: Less than 100 particles/cu ft 0.3 micron and larger shall meet class 100 "Clean Work Station" requirements of Federal Standard 209 and Air Force T.O. 00-25-203.
16.	Overall size and weight	- 60 inches high X 34 inches deep X 51 inches wide - 450 lbs, (for reference only).

Workmanship. Clean Work Stations shall be clean and free from faulty construction or any defect which affects appearance, serviceability, and durability.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

Certification: The contractor shall certify that the product offered meets the salient characteristics of this description and that the product conforms to the producers own drawings, specifications, standards, and quantity assurance practices and is the same product offered for sale in the commercial marketplace. The Government reserves the right to require proof of such " conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Packaging, packing and marking. The items shall be packaged in accordance with normal commercial practice and packed to assure acceptance by common carrier and provide product protection against loss and damage during multiple shipments, handling, and storage. The shipping container shall be in compliance with National Motor Freight Classification and Uniform Freight Classification. Shipping containers shall be marked and specified in the contract or order.

Known acceptable products. The following products are known to meet the requirements of this CID.

Manufacturer"	Model	NSN
Dexon (16085)	H54E229	3694-00-766-5486
Fluid Conditioning Products (94592)	20775	3694-00-766-5486
Liberty Industries (12746)	4-410M	3694-00-766-5486

Custodians:
Air Force - 99
Navy - AS

Preparing Activity
Air Force - 84

(Project: 3694-0030)

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:47Z
Check the source to verify that this is the current version before use.