FEDERAL SPECIFICATION

BANDAGE, ABSORBENT, ADHESIVE

This specification was approved by the Commissioner, Federal Supply Service, General Services Administration, for the use of all Federal agencies.

1. SCOPE AND CLASSIFICATION

1.1 Scope.—This specification covers complete, sterile, protective bandages for use on minor cuts, abrasions, and puncture wounds.

1.2 Classification.

1.2.1 *Sizes.*—The bandages shall be of the following sizes, as specified (see 6.2):

Size 1—⅝ by 1½ inches.
Size 2—¾ by 3 inches.
Size 3—1 by 3 inches.
Size 4—1½ by 1½ inches.
Size 5—⅞ inch round.

2. APPLICABLE SPECIFICATIONS, STANDARDS, AND OTHER PUBLICATIONS

2.1 Specifications and standards.—The following specifications and standards, of the issues in effect on date of invitation for bids, form a part of this specification:

Federal Specifications:

UU-P-31 — Paper; General Specifications and Methods of Testing.
CCC-G-101—Gauze; Plain.
PPP-B-566 — Boxes, Folding, Paperboard.

(Activities outside the Federal Government may obtain copies of Federal Specifications and Standards as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U. S. Government Printing Office, Washington 25, D. C.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Atlanta, Chicago, Kansas City, Mo., Dallas, Denver, San Francisco, Los Angeles, Seattle, and Washington, D. C.

(Federal Government activities may obtain copies of Federal Specifications and Standards and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

Military Specifications:

JAN-P-108—Boxes, Fiberboard, Corrugated and Solid (Military Type).
MIL-B-16152—Box, Folding (First Aid Packages).

Military Standards:

MIL-STD-105 — Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 — Marking for Shipment and Storage.

(Copies of Military Specifications and Standards required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. — The following document forms a part of this specification. Unless otherwise indicated, the issue in effect on date of invitation for bids, forms a part of this specification:

U. S. Pharmacopoeial Convention Publication:

Pharmacopoeia of the United States.

(Application for copies should be addressed to the Mack Publishing Company, Easton, Pa.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:31Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 Material.

3.1.1 *Gauze.*—The gauze shall be type I or type II bleached gauze conforming to Federal Specification CCC–G–101.

3.1.2 *Adhesive plaster backing.*—The backing material shall be a flesh-colored, continuous, uniformly formed, vinyl film plasticized with a nontoxic compound, and shall have a breaking strength of 6 pounds minimum (per inch width).

3.1.3 *Adhesive mass.*—The adhesive mass shall be a mixture having pressure-sensitive adhesive properties, the base of which shall be new rubber, synthetic rubber, or rubber substitute. Reclaimed material shall not be used, wholly or in part.

3.1.4 *Facing material.*—The facing material shall be smooth, continuous, organic or metallic film, laminated or plain. The facing material shall be capable of being removed without breaking or cracking, and shall be chemically inert to the adhesive mass.

3.1.5 *Paper for envelope.*—The paper shall be 100 percent sulfite or glassine type, having a basis weight of not less than 25 pounds for 500 sheets (24 by 36 inches), and having a bursting strength of not less than 15 points when tested in accordance with method 112 of Federal Specification UU–P–31.

3.2 Construction.

3.2.1 *Adhesive plaster.* — The adhesive plaster shall be made of an adhesive plaster backing as specified in 3.1.2, with a suitable adhesive mass as specified in 3.1.3, which is either sprayed, spread, or calendered on one side. If the adhesive mass is sprayed, the solvent used shall not remain within the plaster mass. Except for size 1, the backing for the bandages shall be perforated where the compress is to be attached, with holes or slots to permit adequate ventilation.

3.2.2 *Compress for sizes 2, 3, and 4.*—The compress for sizes 2, 3, and 4, shall be made in accordance with either of the following constructions:

Compress shall consist of not less than four layers of gauze. The top layer (the layer furthest from the plastic backing) shall be type I gauze, and the under layers, either type I or type II gauze.

Compress shall be composed of a clean, absorbent, cellulosic material covered with one layer of type I gauze, and the weight shall be not less than that of a compress of the same area composed of four layers of type I gauze. The absorbent capacity of the pad shall be equal to or greater than that of a pad of equal area composed of four layers of type I gauze. The pad shall be composed of materials which shall not disintegrate when standing in water for 24 hours.

3.2.2.1 *Compress for sizes 1 and 5.*—The compress for sizes 1 and 5 shall be made in accordance with either of the following constructions:

Compress shall consist of not less than four layers of type I gauze.

Compress shall be composed of one layer of bleached, absorbent, unnapped cotton flannel weighing not less than 6 ounces per square yard.

3.2.3 *Assembly.* — The compress shall be firmly attached to the approximate center of the strip of adhesive plaster in such a manner that there will be a minimum of ravelings or loose threads from the compress. Two strips of the facing material shall be applied to the exposed portion of the adhesive plaster in such a manner as to overlap the compress and allow removal without contamination of the compress.

3.2.4 *Dimensions.*—Dimensions of bandage and compress shall be as shown in table I.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:31Z
Check the source to verify that this is the current version before use.

TABLE I.

Size	Overall dimension of bandage		Overall dimensions of compress	
	Width or diameter	Length	Width or diameter	Length
	Inches	*Inches*	*Inches*	*Inches*
1	⅜ ± 1/32	1½ ± 3/32	⅜ ± 1/32	9/16 to 5/8
2	⅝ ± 1/16	2 27/32 minimum	⅝ ± 1/16	1 ± ⅛
3	⅞ ± 1/16	2 27/32 minimum	⅞ ± 1/16	1 ± ⅛
4	1½ ± 1/16	1½ ± 1/16	⅝ ± ⅛	⅝ ± ⅛
5	⅞ ± 1/16	----------	⅞ ± 1/16	----------

3.2.4.1 The width of the compress for sizes 1, 2, and 3, shall be equal to or less than the width of the adhesive plaster backing.

3.2.4.2 The compress for sizes 4 and 5 shall be approximately centered on the adhesive backing.

3.3 Adhesive quality.—The adhesive quality of the plaster shall meet the test specified in 4.3.2.

3.4 Wrapping. — Each completed bandage shall be enclosed in a paper envelope made from paper as specified in 3.1.5 and sealed. The envelope shall be so constructed and sealed that sterility is maintained until the envelope is opened. The envelope shall be marked so as to permit ready identification and shall bear name or registered trade mark of the manufacturer. The envelope shall be marked with the legend "STERILE" and such other information required by law.

3.5 Sterilization and certification of sterility.

3.5.1 *Sterilization.* — Sterilization shall be accomplished by high-pressure steam or other suitable means so as to render the bandage sterile in the final package. The bandage shall be thoroughly dried after sterilization.

3.5.2 *Certification of sterility.*—A certificate of sterility, listing all lot or control numbers, stating that samples from each sterilizer charge have been tested and found to be sterile, shall be submitted by the contractor to the procuring activity.

3.6 Workmanship.—Workmanship shall be first class throughout. Compress and adhesive plaster bandages shall be free from defects which detract from their appearance or which may impair their serviceability.

4. SAMPLING, INSPECTION, AND TEST PROCEDURES

4.1 Sampling.

4.1.2 *Method of drawing samples.*

4.1.2.1 *Sampling for visual inspection.*— The sample size for visual inspection shall be determined according to the lot size in accordance with Military Standard MIL–STD–105.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:31Z
Check the source to verify that this is the current version before use.

4.1.2.2 *Sampling for dimensional inspection and tests.*—Unless otherwise specified, sampling shall be in accordance with Military Standard MIL–STD–105.

4.2 Inspection.—Samples shall be inspected to determine compliance with specification requirements (see 6.2).

4.2.1 *Location of inspection.*—Inspection of the finished bandages may be made either at point of production or at point of delivery, as specified (see 6.1). In case of factory inspection, every facility shall be afforded inspectors, by the manufacturer, for the prosecution of their work.

4.3 Tests.

4.3.1 *Sterilization.*—The purchasing agency may satisfy itself that sterilization has been properly accomplished by testing for sterility. Sterility shall be determined in accordance with the current edition of the Pharmacopoeia of the United States for solids.

4.3.2 *Adhesive quality.*—Random samples of adhesive bandages shall be placed in an oxygen bomb. Oxygen pressure of 300 pounds per square inch shall be maintained for two days at 70° centigrade. At the end of the two days, the adhesive bandages shall be removed and the adhesive mass shall appear essentially unchanged. Adhesive bandages on which the adhesive mass strings out in long threads when the thumbnail is impressed and withdrawn, or adhesive bandages applied to the skin upon removal, leave quantities of adhesive mass on the skin, are considered rejectable.

4.3.3 *Classification of defects.*—Inspection of the bandage, absorbent, adhesive, shall be conducted as follows:

Categories of defects

Major AQL 0.65 percent

101. Adhesive does not stick.

102. Facing material not removable without contaminating the compress.

103. Adhesive plaster not perforated, where compress is attached, with holes or slots, except for size 1.

104. Paper envelopes not properly sealed.

105. Thread count of compress low.

Minor AQL 2.5 percent

201. Raveling of the edges of the compress.

202. Compress not centered properly.

203. Unit package not properly sealed.

5. PREPARATION FOR DELIVERY

(For definitions and levels see 6.1.)

5.1 Packaging.

5.1.1 *Unit of issue.*—Eighteen or one hundred bandages, of the stock number specified, constitute one unit.

5.1.2 *Quantities.*—Packaging and packing quantities shall be as specified in table II.

TABLE II.—*Packaging and packing quantities*

Stock number	Number of bandages in unit carton	Number of units in intermediate carton	Number of units in shipping container
6510–597–7468	18	48	576
6510–597–7469	100	36	144

5.1.2.1 *Packing variation permitted.*—If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:31Z
Check the source to verify that this is the current version before use.

5.1.3 *Level A.*

5.1.3.1 *Unit carton.*

5.1.3.1.1 *Stock No. 6510–597–7468.*—The number of bandages specified in table II shall be packaged in a size 1 carton conforming to Military Specification MIL–B–16152.

5.1.3.1.2 *Stock No. 6510–597–7469.*—The number of bandages specified in table II shall be packaged in a carton of appropriate size constructed in accordance with Federal Specification PPP–B–566, style I. Closure shall comply with appendix thereto.

5.1.3.2 *Intermediate package.*—Intermediate package shall be a box of appropriate size constructed in accordance with Military Specification JAN–P–108, compliance symbol W5c or W5s. Box shall be sealed by means of a water-resistant tape.

5.1.4 *Level C.*—Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the end product for its intended purpose within the continental limits of the United States of America.

5.2 **Packing.**

5.2.1 *Level A.*

5.2.1.1 *Exterior container.*—Exterior container shall be a box of appropriate size constructed in accordance with Military Specification JAN–P–108, compliance symbol V3c or V3s. Closure shall comply with appendix thereto. Strapping will not be required.

5.2.2 *Level C.* — The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 **Marking.**

5.3.1 *Unit carbon.*

5.3.1.1 *Stock No. 6510–597–7468.* — Each unit carton shall be marked as specified in 3.4 of Military Specification MIL–B–16152. Front face of carton shall bear the word "STERILE" in addition to the other marking required. Back face of carton shall bear the following instructions:

"DIRECTIONS

Cleanse wound. Apply bandage placing pad over wound. Do not apply too tightly. For best results be sure skin is clean and dry before applying bandage. Good medical practice suggests changing bandage if pad becomes wet."

5.3.1.2 *Stock No. 6510–597–7469.* — Each unit carton shall be marked in accordance with section 5 of Military Standard MIL–STD–129, and shall bear the word "STERILE". Lot (control) number will not be required.

5.3.2 *Intermediate cartons.* — Each intermediate carton shall be marked in accordance with section 5 of Military Standard MIL–STD–129, and shall bear the word "STERILE". Lot (control) number will not be required.

5.3.3 *Shipping containers.*—Each shipping container shall be marked in accordance with Military Standard MIL–STD–129, and shall bear the word "STERILE". Service color marking and symbol are not required on containers for level A packs when addressed to domestic establishments for the Medical Services of the Armed Forces. Lot (control) number shall be shown on shipping container.

6. NOTES

6.1 Federal Standard 102 should be referred to for definitions and applications of the various levels of packaging protection for supplies and equipment.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:31Z
Check the source to verify that this is the current version before use.

6.2 **Ordering data.** — Procurement documents should specify the following:

(*a*) Title, number, and date of this specification.
(*b*) Size required (see 1.2.1).
(*c*) Sampling required (see 4.1).
(*d*) Inspection required (see 4.2).
(*e*) Location of inspection (see 4.2).
(*f*) Selection of applicable levels of packaging and packing (see 5.1, and 5.2).
(*g*) Marking (see 5.3).

6.3 This specification covers the following items appearing in the Armed Services Medical Stock List:

6510–597–7468 Bandage, Absorbent, Adhesive, ¾ by 3 inches, 18s.
6510–597–7469 Bandage, Absorbent, Adhesive, ¾ by 3 inches, 100s.

6.4 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of the specification, or which are commercially available, but is intended to cover the types, etc., which are normally procured to meet Federal Government requirements.

6.5 **Transportation description.**—Transportation description applicable to this item is:

Bandages, surgeons,
Carload minimum weight 24,000 pounds, subject to Rule 34, Uniform Freight Classification.
Truckload minimum weight 30,000 pounds.

Patent notice.—When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility, nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data, is not to be regarded by implication or otherwise, as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

MILITARY INTERESTS:
Army—M
Navy—Md
Air Force.

Source: https://assist.dla.mil -- Downloaded: 2026t0-26T12:31Z
Check the source to verify that this is the currentversion before use.