DDD-B-40a
DECEMBER 23, 1958
SUPERSEDING
Fed. Spec. DDD-B-40
August 22, 1949

FEDERAL SPECIFICATION

BANDAGE, CANTON FLANNEL

This specification was approved by the Commissioner, Federal Supply Service, General Services Administration, for the use of all Federal agencies.

1. SCOPE AND CLASSIFICATION

1.1 Scope.—This specification covers canton flannel bandages.

1.2 Classification.

1.2.1 *Sizes.*—Canton flannel bandages shall be of the following sizes, as specified (see 6.1).

Size 1.—3 inches by 5 yards.

Size 2.—5 inches by 5 yards.

2. APPLICABLE SPECIFICATIONS AND STANDARDS

2.1 The following specifications and standards, of the issues in effect on date of invitation for bids, form a part of this specification:

Federal Specifications:

CCC-T-191—Textile Test Methods.
PPP-B-566 — Boxes, Folding, Paperboard.
PPP-B-585—Boxes; Wood, Wirebound.
PPP-B-601—Boxes, Wood, Cleated-Plywood.
PPP-B-621—Boxes, Wood, Nailed and Lock-Corner.
PPP-B-676 — Boxes, Set-Up, Paperboard.

Federal Standards:

Fed. Std. No. 102—Preservation, Packaging, and Packing Levels.
Fed. Std. No. 123—Marking for Domestic Shipment (Civilian Agencies).

(Activities outside the Federal Government may obtain copies of Federal Specifications, Standards, and Handbooks as outlined under General Information in the Index of Federal Specifications, Standards, and Handbooks and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U. S. Government Printing Office, Washington 25, D. C.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Atlanta, Chicago, Kansas City, Mo., Dallas, Denver, San Francisco, Los Angeles, Seattle, and Washington, D. C.

(Federal Government activities may obtain copies of Federal Specifications, Standards, and Handbooks and the Index of Federal Specifications, Standards, and Handbooks from established distribution points in their agencies.)

Military Specifications:

MIL-B-10377—Box, Wood, Cleated, Veneer, Paper Overlaid.
MIL-L-10547 — Liners, Case, Waterproof.

Military Standards:

MIL-STD-105 — Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 — Marking for Shipment and Storage.

(Copies of Military Specifications and Standards required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

481464/58/168

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:31Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 Material.—The material shall be made of thoroughly cleaned, unbleached, and undyed cotton, free from waste.

3.1.1 *Over-all length.*—The average over-all length of the bandages shall be 5 yards ± 2 inches, when tested as specified in 4.3.1.

3.1.2 *Width.*—The over-all width of the bandage shall be 3 inches for size 1 and 5 inches for size 2, ± ⅛ inch, when tested as specified in 4.3.2.

3.1.3 *Weight per square yard.*—The weight per square yard shall be not less than 5.25 ounces nor more than 5.55 ounces when tested in accordance with 4.3.3.

3.1.4 *Thread count.*—Thread count shall be 66 in the warp and 42 in the filling. A minus tolerance of three threads and any plus tolerance will be permitted in each direction, when tested in accordance with 4.3.4.

3.1.5 *Breaking strength.*—The breaking strength shall be not less than 45 pounds in the warp, when tested in accordance with 4.3.5.

3.1.6 *Nap.*—The material shall be napped on the filling side.

3.1.7 *Weave.*—The weave shall be 3 or 4 harness twill.

3.2 Construction.—The bandages shall be of the roller type and of one continuous piece of material, tightly and evenly rolled. The edges of the bandages shall be pinked to prevent raveling.

4. SAMPLING, INSPECTION, AND TEST PROCEDURES

4.1 Sampling.

4.1.1 *For visual examination.*—Sampling procedures for visual examination shall be conducted in accordance with the procedures set forth in Military Standard MIL–STD–105. Sample size for visual examination shall be as indicated for inspection level L4, in Appendix to Military Standard MIL–STD–105. The unit of product for inspection purposes shall be in the individual bandage, except where otherwise specified. Samples used for visual examination may also be used for tests.

4.1.2 *For tests.*

4.1.2.1 *End item.*—Sample size for tests shall be as indicated for inspection level L1, in Appendix to Military Standard MIL–STD–105, except that the acceptance number shall be zero for all sample sizes, and a minimum of three samples shall be taken.

4.2 Inspection.—The inspector shall ascertain, where possible, that the canton flannel used in the fabrication of the bandages has been inspected and accepted in accordance with the requirements of this specification.

4.2.1 *Classification of defects.*

Categories and defects[1]

Major A — *AQL 1.0 percent defective*

101. Bandage not clean, not free from dirt or other foreign matter.
102. Bandage not one continuous piece.
103. Bandage not pinked.

Minor A — *AQL 2.5 percent defective*

201. Bandage not evenly rolled.
202. Bandage not tightly rolled.
203. Length and width not within tolerances specified.
204. Napping not even.
205. Bandage not free of holes, tears, or cuts.

Minor B — *AQL 2.5 percent defective*
(Unit of product — package)

251. Labeling or printing on package not legible and free from defacing marks, scratches or tears.

[1] Inspection is not restricted to the classified possible defects listed above.

4.3 Tests.

4.3.1 *Over-all length.*—The length of the bandages shall be determined according to

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:31Z
Check the source to verify that this is the current version before use.

method 5010, procedure 4.2, of Federal Specification CCC–T–191. The report shall be the average of the samples measured. No individual value shall vary more than 2 percent of 5 yards.

4.3.2 *Width.*—The width of the bandage, exclusive of the pinked edge, shall be determined according to method 5020 of Federal Specification CCC–T–191.

4.3.3 *Weight per square yard.*—Test for weight per square yard shall be conducted according to method 5041 of Federal Specification CCC–T–191.

4.3.4 *Thread count.*—Test for thread count shall be conducted according to method 5050 of Federal Specification CCC–T–191. Three specimens each, in the warp, and filling directions, shall be tested from each unit or product.

4.3.5 *Breaking strength.*—Test for breaking strength shall be conducted according to method 5100 of Federal Specification CCC–T–191, except that when tests are performed on the end item, the size of the specimens for the size 1 bandage shall be 3 inches by 6 inches for the warp direction. Tests on the filling shall not be performed. Only one specimen shall be tested from each unit of product.

4.4 **Supplier responsibilities for inspection for military procurement.**—The supplier shall perform examinations and tests set forth in this specification to substantiate conformance of supplies to specification requirements. The examinations and tests shall be at the expense of the supplier. Suppliers shall have adequate test facilities or shall arrange for test facilities acceptable to the Government to be used for executing specific prescribed tests. Records of examinations and tests performed by the supplier shall be maintained by the supplier and made available to the Goverment upon the Government's request, at any time, or from time to time for a period of 2 years after the delivery of the supplies to which these records relate. The supplier shall be responsible for compliance with all requirements of this specification and all other technical requirements of the contract.

5. PREPARATION FOR DELIVERY

(Nonmilitary agencies.—Levels of packaging and packing protection used are defined in Federal Standard No. 102.)

5.1 For nonmilitary agencies.

5.1.1 *Packaging and packing.*—The subject commodity shall be packaged and packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.1.2 *Marking.*

5.1.2.1 *Interior packages.*—Each package shall be marked to identify contents.

5.1.2.2 *Shipping containers.* — Shipping containers shall be marked with stock number, name of material, quantity, name of consignee, name of consignor, and contract number as shown on contract or purchase order. In addition to any special markings specified in the contract or order, shipping containers shall be marked in accordance with Federal Standard No. 123.

5.2 For military agencies.

5.2.1 *Packaging.*

5.2.1.1 *Unit of issue.*—Twelve (12) bandages, of the stock number specified, constitute one unit.

5.2.1.2 *Packaging quantities.*—The number of units of issue indicated in table I shall be packaged in each unit and exterior container,

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:31Z
Check the source to verify that this is the current version before use.

as applicable for the required level of protection specified in procurement document.

TABLE I.—*Packaging quantities*

Federal stock number	Unit package	Intermediate package	Exterior container
6510–200–1000	1 unit	Not required	24 units
6510–200–2000	1 unit	Not required	12 units

5.2.1.2.1 *Packing variation permitted.*—If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.2.1.3 *Level A.*

5.2.1.3.1 *Unit package.* — Each bandage shall be securely wrapped and sealed in a good quality, heavy white paper. Twelve bandages (1 unit) shall be packaged in a carton of appropriate size constructed in accordance with Federal Specification PPP–B–566 or PPP–B–676. Closure shall be adequate to prevent spillage of contents under normal handling.

5.2.1.4 *Level C.*—Unit shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose within the continental limits of the United States of America.

5.2.2 *Packing.*

5.2.2.1 *Level A.*

5.2.2.1.1 *Exterior container.* — Exterior container shall be designed for type 2 load, and constructed in accordance with Federal Specification PPP–B–585, style 3 for class 3 use; PPP–B–601, table II, using type III, class 1 plywood; PPP–B–621, table III; or Military Specification MIL–B–10377, table II.

5.2.2.1.2 *Case liner.*—Each level A box shall be lined with a waterproof case liner, conforming to Military Specification MIL–L–10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto.

5.2.2.1.3 *Strapping.*—Strapping, when required, shall be in accordance with applicable box specification.

Note.—Strapping will not be required for shipment forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2.2 *Level B.*

5.2.2.2.1 *Exterior container.* — Exterior container shall conform to Federal Specification PPP–B–601, table I, using type III, class 1 plywood, or shall be designed for type 2 load and constructed in accordance with Federal Specification PPP–B–585, style 3, for class 2 use; PPP–B–621, table I; or Military Specification MIL–B–10377, table I.

5.2.2.3 *Level C.*—The subject commodity shall be packed in substantial commercial containers of the type, size and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.2.3 *Marking.*

5.2.3.1 *Individual bandage.*—Each wrapper shall be printed with name of commodity, size, name or registered trade-mark of manufacturer and such other information as may be required by statute.

5.2.3.2 *Unit package.*—Each unit package shall be marked in accordance with section 5 of Military Standard MIL–STD–129.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:31Z
Check the source to verify that this is the current version before use.

5.2.3.3 *Exterior container.*—Exterior container shall be marked in accordance with Military Standard MIL–STD–129. Service color and service symbol markings are required only on direct shipments from the contractor's plant to overseas activities.

6. NOTES

6.1 Ordering data.—Purchasers should exercise any desired options offered herein, and procurement documents should specify the following:

(*a*) Title, symbol, and date of this specification.

(*b*) Size required (see 1.2).

(*c*) Selection of applicable levels of packaging and packing (see 5.2.1 and 5.2.2).

6.2 This specification covers the following items appearing in the Armed Services Medical Stock List:

6515–200–1000 BANDAGE, CANTON FLANNEL, 3 inches by 5 yards, 12s.
6515–200–2000 BANDAGE, CANTON FLANNEL, 5 inches by 5 yards, 12s.

6.3 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of this specification or which are commercially available, but is intended to cover the types, etc., which are normally procured to meet Federal Government requirements.

6.4 Transportation description.—Transportation description applicable to this item is:

Bandage, Surgeons'
Carload minimum weight 24,000 pounds subject to Rule 34, Uniform Freight Classification.
Motor volume minimum weight 24,000 pounds.

Notice. — When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data, is not to be regarded by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

MILITARY INTERESTS:
Army—M
Navy—Md
Air Force.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:31Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:31Z
Check the source to verify that this is the current version before use.