FEDERAL SPECIFICATION

BANDAGE, GAUZE, TUBULAR

This specification was approved by the Commissioner, Federal Supply Service, General Services Administration, for the use of all Federal agencies.

1. SCOPE AND CLASSIFICATION

1.1 **Scope.**—This specification covers tubular gauze bandages designed for various portions of the body.

1.2 **Classification.**

1.2.1 *Sizes.*—Tubular gauze bandages shall be of the following sizes, as specified (see 6.1).

Size 1.—5/8 inch by 5 yards.
Size 2.—5/8 inch by 50 yards.
Size 3.—7/8 inch by 5 yards.
Size 4.—7/8 inch by 50 yards.
Size 5.—1½ inches by 5 yards.
Size 6.—1½ inches by 50 yards, close weave.
Size 7.—1½ inches by 50 yards, loose weave.
Size 8.—2 inches by 50 yards.

2. APPLICABLE SPECIFICATIONS AND STANDARDS

2.1 The following specifications and standards, of the issues in effect on date of invitation for bids, form a part of this specification:

Federal Specifications:

PPP-B-566 — Boxes, Folding, Paperboard.
PPP-B-636 — Boxes, Fiber.
PPP-B-676 — Boxes, Set-Up, Paperboard.
PPP-T-76 — Tape, Pressure-Sensitive Adhesive, Paper, Water Resistant.

Federal Standards:

Fed. Std. No. 102—Preservation, Packaging, and Packing Levels.
Fed. Std. No. 123—Marking for Shipment and Storage (Civil Agencies).

(Activities outside the Federal Government may obtain copies of Federal Specifications, Standards, and Handbooks as outlined under General Information in the Index of Federal Specifications, Standards, and Handbooks and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U. S. Government Printing Office, Washington 25, D. C.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Atlanta, Chicago, Kansas City, Mo., Dallas, Denver, San Francisco, Los Angeles, Seattle, and Washington, D. C.

(Federal Government activities may obtain copies of Federal Specifications, Standards, and Handbooks and the Index of Federal Specifications, Standards, and Handbooks from established distribution points in their agencies.)

Military Standards:

MIL-STD-105 — Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 — Marking for Shipment and Storage.

481466/59/221

FSC 6510

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:32Z
Check the source to verify that this is the current version before use.

(Copies of Military Specifications and Standards required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material.

3.1.1 *Gauze.*—Gauze shall be made from double-carded peeler, knitting twist, soft, non-paraffined bleached yarn.

3.1.2 *Applicator.*—Applicator shall be fabricated from aluminum, corrosion-resisting steel, or cellulose acetate.

3.2 **Fabric construction.** — Bandages shall be made from plain, seamless, tubular knit gauze, as specified in 3.1.1, and shall conform to the requirements shown in table I for the fabric construction. Bandages shall be so constructed and shall have such form and flexibility as to make application and removal easy, and shall conform to all contours, without binding. Bandages shall be in one continuous piece.

3.2.1 *Stretching width of fabric.*—A length of tubular bandage, as indicated in table II for the appropriate size, shall be capable of being stretched, without tearing, to the stretched width specified, when tested as described in 4.4.1.

3.3 **Color.**—Color of tubular bandages shall be white.

3.4 **Applicators.**—Applicators shall be fabricated of material specified in 3.1.2, and shall be designed to facilitate application of tubular bandages to the fingers and toes. Applicators shall be for use with sizes 1, 2, 3, and 4 bandages, and shall be supplied with those sizes.

3.5 **Instructions.** — Complete instructions for use and application of the bandages shall be furnished with each unit package by the contractor. Bidder shall submit with bid the text or illustration of instructions for use and application that he proposes to furnish.

3.6 **Workmanship.**—Workmanship shall be first class throughout. The bandages and applicators shall be free from defects which detract from their appearance or may impair their serviceability.

4. SAMPLING, INSPECTION, AND TEST PROCEDURES

4.1 The provisions contained in this section are applicable to inspection performed by and

TABLE I.—*Fabric construction*

Size	Width of bandage on core ± 10%	Weight per roll ± 5%	Cylinder needles	Stitches or courses per linear inch ± 5%	Yarn count	Length of bandage ± 2%
	Inches	*Ounces*				*Yards*
1	⅝	0.575	40	18	20/1	5
2	⅝	5.75	40	18	20/1	50
3	⅞	0.740	50	18	20/1	5
4	⅞	7.40	50	18	20/1	50
5	1½	1.1	80/81	8¾	20/1	5
6	1½	11.0	80/81	11	20/1	50
7	1½	10.50	80/81	8¾	20/1	50
8	2	11.50	110	9	24/1	50

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:32Z
Check the source to verify that this is the current version before use.

TABLE II.—*Stretching width of fabric*

Size	Length to be tested	Stretched width ± 5%
	Inches	*Inches, minimum*
1	2	2½
2	2	2½
3	2	2¾
4	2	2¾
5	8	10½
6	8	8
7	8	10¾
8	8	13½

for the Government. Sampling, inspection, and test procedures for components or materials described shall be in accordance with this specification and with specifications and standards referenced herein to the extent applicable except that this specification shall govern in the event of conflict.

4.1.1 *Supplier responsibilities for inspection for military procurement.*—Such examinations and tests as are set forth in this specification, or as shall otherwise be appropriate or necessary to insure that supplies conform to specification requirements, shall be performed by and at the expense of the supplier. Suppliers who do not have facilities adequate for such tests shall arrange for the use of test facilities acceptable to the Government. Records of examinations and tests performed by the supplier shall be maintained by the supplier and made available to the Government, upon the Government's request, at any time, or from time to time, for a period of 2 years after delivery of the supplies to which such records relate.

4.1.2 Inspection, as used in this specification, is defined as both examination (such as visual, auditory, etc. investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and the test results obtained and indicating that material is as specified, may be furnished.

4.2 **Sampling.** — Sampling shall be conducted in accordance with Military Standard MIL–STD–105 and appendix thereto.

4.2.1 Where possible, the same sample shall be used for the determination of two or more tests.

	Inspection level	AQL percent defective
For visual examination (Unit for inspection shall be packaged roll)		
Major A	L4	2.5
Total		6.5
Major B	L4	2.5
Total		6.5
For dimensional examination (Unit for inspection shall be packaged roll)	L2	6.5

For end item testing. Requirement is applicable to individual unit only.

Characteristic	Test procedure	Rqmt. indiv. unit	Sample unit	Insp. level	AQL	Minimum sample size
Stitching width	4.4.1	x	packaged roll	L4	*	3

* Acceptance number shall be zero (0) for all sample sizes.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:32Z
Check the source to verify that this is the current version before use.

4.3 **Inspection.**—The tubular gauze bandage shall be inspected to determine compliance with all requirements contained in sections 3 and 5 of this specification. Nonconformance with these requirements will be permitted to the extent indicated in 4.2.

4.3.1 *Classification of defects.*—Inspection shall be conducted in accordance with the following classifications of defects:

*Categories and defects****

Major A

101. Applicator missing or broken on sizes 1, 2, 3 and 4.
102. Weight per roll not as specified.
103. Bandage not free of holes, tear, or cuts.
104. Bandage not one continuous length.
105. Fifty-yard rolls not free-wheeling.

Major B

151. Labeling not present or not correct.
152. Instructions not present.

Minor A

201. Bandage not evenly and uniformly rolled.
202. Bandage not clean and free from foreign material.
203. Cylinder needles and courses not as specified.

Minor B

251. Labeling or printing not legible and free from defacing marks.

** Inspection is not restricted to the classified possible defects listed above.

4.3.2 *Dimensional examination.*—The tubular gauze bandage shall be inspected for defects in dimensions. Any dimensions not within the tolerance specified herein shall be classified as a defect.

4.4 **Tests.**

4.4.1 *Stretching width.*—The prescribed length of bandage shall be cut from the roll, and two polished steel pins measuring approximately 9 inches long by $\frac{5}{32}$ inch in diameter shall be inserted into the tube formed by the cut length of bandage. A slow, steady, manual pull shall be exerted on one pin; the other pin being fixed. The pin shall be maintained parallel during the test. The distance between the two pins shall be measured with the stretching force applied, and the stretching width shall be taken as the distance between the farthest points on the bandage in the direction of pull.

5. PREPARATION FOR DELIVERY

For civil agency procurement the levels of preservation, packaging, and packing protection used are defined in Federal Standard No. 102.

5.1 **For nonmilitary agencies.**

5.1.1 *Packaging and packing.*—The subject commodity shall be packaged and packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.1.2 *Marking.*

5.1.2.1 *Interior packages.*—Each package shall be marked to identify contents.

5.1.2.2 *Shipping containers.* — Shipping containers shall be marked with stock number, name of material, quantity, name of consignee, name of consignor, and contract number as shown on contract or purchase order. In addition to any special marking, shipping containers shall be marked in accordance with Federal Standard No. 123.

5.2 **For military agencies.**

5.2.1 *Packaging.*

5.2.1.1 *Unit of issue.*—One roll of bandage, of the stock number specified, constitutes one unit.

5.2.1.2 *Packaging quantities.*—The number of units of issue indicated in table III shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in procurement document.

Source: https://assist.dla.mil -- Downloaded: 2026l0-26T12:32Z
Check the source to verify that this is the currentversion before use.

5.2.1.2.1 *Packing variation permitted.*—If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.2.1.3 *Level A.*

5.2.1.3.1 *Unit package.*—The number of yards specified in table III shall be neatly and evenly wound on a core to form a roll of suitable size and design. Fifty-yard rolls shall be fitted with a free-wheeling type core. Each unit shall be packaged in a carton of appropriate size constructed in accordance with Federal Specification PPP–B–566 or PPP–B–676. Carton shall be of such design as to permit dispensing of the bandage without removal of the roll from carton. Carton shall be adequately secured to prevent accidental opening. Each carton shall be completely wrapped and heat sealed in 300 MST cellophane, or equal.

5.2.1.3.2 *Intermediate package.* — Intermediate package shall be a weather-resistant fiberboard box constructed in accordance with Federal Specification PPP–B–636, class 2, compliance symbol W5c or W5s, style RSC, except that the manufacturer's joint shall be effected by the application of water-resistant tape or by gluing integral lap to outer face of end or side panel using water-resistant adhesive in lieu of metal stitches. Closure shall be effected by means of water-resistant gummed or pressure-sensitive tape. The tape shall be not less than 2 inches wide and shall be applied to both top and bottom so that the center seams are covered their full length. Tape shall extend at least 2 inches onto each of the end panels.

5.2.1.4 *Level C.*—Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose within the continental limits of the United States of America.

5.2.2 *Packing.*

5.2.2.1 *Level A.*

5.2.2.1.1 *Exterior container.* — Exterior container shall be a box of appropriate size constructed in accordance with Federal Specification PPP–B–636, class 2, compliance symbol V3c or V3s, style RSC. Closure shall be as specified in 30.2.2.1 thereof and, in addi-

TABLE III.—*Packaging and packing quantities*

Stock No.	Width	Number of yards constituting one unit	Number of unit cartons to be included in each intermediate carton	No. of intermediate cartons to be included in each exterior container	Total number of units to be included in each exterior container
	Inches				
6510–200–7007	⅝	5	72	4	288
6510–200–7010	⅝	50	36	2	72
6510–200–7013	⅞	5	72	4	288
6510–200–7015	⅞	50	36	2	72
6510–200–7017	1½	5	72	4	288
6510–200–7020	1½	50	24	2	48
6510–200–7025	1½	50	24	2	48
6510–200–7030	2	50	24	2	48

Source: https://assist.dla.mil -- Downloaded: 20261O-26T12:32Z
Check the source to verify that this is the current version before use.

tion, the top and bottom center seams shall be closed by means of tape conforming to the requirements of Federal Specification PPP–T–76. The tape shall be not less than 3 inches wide and shall extend at least 2 inches onto each of the end panels. Strapping will not be required.

5.2.2.2 *Level C.*—The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 *Unit package.* — Each unit package shall be marked with stock number, item identification, unit, quantity, contract number, and name of contractor. In those instances where cellulose acetate applicators are supplied, the unit package shall also bear the following legend which may be applied by means of a label:

"Applicator may be sterilized, if so desired, by use of alcohol.

WARNING: Sterilization by use of phenol solutions, autoclaving or boiling may destroy the applicator."

5.3.2 *Intermediate package.*—Each intermediate package shall be marked in accordance with section 5 of Military Standard MIL–STD–129.

5.3.3 *Exterior container.*—Exterior container shall be marked in accordance with Military Standard MIL–STD–129, except that 6.2(c), 6.2(d), and 6.5 shall not be applicable. Service color and service symbol markings are required only on direct shipments from the contractor's plant to overseas activities.

6. NOTES

6.1 Ordering data.—Purchasers should exercise any desired options offered herein, and procurement documents should specify the following:

(*a*) Title, number, and date of this specification.
(*b*) Size required (see 1.2.1).
(*c*) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 Intended use.—Tubular gauze bandages are of the various sizes designed for the purpose of bandaging parts of the body, as follows:

Sizes 1 and 2—Small fingers and toes.
Sizes 3 and 4—Large fingers and toes.
Sizes 5 and 7—Hands, feet, forearms, legs, knees, and elbows.
Size 6 —Foot dressings, wrists, elbows, wet dressings.
Size 8 —Heads and shoulders.

6.3 This specification covers the following items appearing in the Armed Services Medical Stock List:

6510–200–7007 BANDAGE, GAUZE, TUBULAR, 5/8 inch by 5 yards.
6510–200–7010 BANDAGE, GAUZE, TUBULAR, 5/8 inch by 50 yards.
6510–200–7013 BANDAGE, GAUZE, TUBULAR, 7/8 inch by 5 yards.
6510–200–7015 BANDAGE, GAUZE, TUBULAR, 7/8 inch by 50 yards.
6510–200–7017 BANDAGE, GAUZE, TUBULAR, 1½ inches by 5 yards.
6510–200–7020 BANDAGE, GAUZE, TUBULAR, 11 stitches per inch, 1½ inches by 50 yards.
6510–200–7025 BANDAGE, GAUZE, TUBULAR, 8¾ stitches per inch, 1½ inches by 50 yards.
6510–200–7030 BANDAGE, GAUZE, TUBULAR, 2 inches by 50 yards.

6.4 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of this specification, or which are commercially available, but

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:32Z
Check the source to verify that this is the current version before use.

Acceptance number shall be zero (0) for all sample sizes.

is intended to cover the types, etc., which are normally procured to meet Federal Government requirements.

6.5 **Transportation description.**—Transportation description applicable to this item is:

Bandages, Surgeons'
Carload minimum weight 24,000 pounds subject to Rule 34, Uniform Freight Classification.
Motor volume minimum weight 24,000 pounds.

Notice. — When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data, is not to be regarded by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

MILITARY INTERESTS:
Army—M
Navy—Md
Air Force.

Copies of this specification may be purchased for 5 cents each.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:32Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:32Z
Check the source to verify that this is the current version before use.