INTERIM AMENDMENT

TO

FEDERAL SPECIFICATION

BANDAGE, GAUZE, (ROLLER AND COMPRESSED)

This interim amendment was developed by the Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, PA 19101, based upon currently available technical information. It is recommended that Federal agencies use it in procurement and forward recommendations for changes to the preparing activity at the address shown above.

The General Serivces Administration has authorized the use of this interim amendment as a valid exception to Federal Specification DDD-B-61E, dated September 11, 1972.

PAGE 1

2.1 Add the following:

"Military Specifications:

"MIL-S-36586 Sterilization Test Strip Set, Bacterial Spore."

PAGE 2

3.2 Description. At the end of the paragraph add:
"Type II size e shall be supplied with two medium safety pins conforming to size 2 of FF-P-416. Type II size c shall be supplied with two small safety pins conforming to size 1 of FF-P-416. The two safety pins required for type II sizes c, e and g shall be placed inside a kraft or other suitable type envelope, and the envelope placed inside the polyethylene bag."

Delete paragraphs 3.3 through 3.3.3 and substitute:

"3.3 Sterility. The bandage shall be sterile when tested with a biological indicator as specified in 4.2.4.2.

Source: https://assist.dla.mil -- Downloaded: 202610-26T12:32Z
Check the source to verify that this is the currentversion before use.

"3.3.1 Certificate of sterility. A certificate of sterility, stating that samples from each sterilizer lot have been tested and found to be sterile, shall be submitted to the procuring activity. The certificate of sterility shall include:

a. Item Identification.
b. National stock number.
c. Contract or purchase order number.
d. Sterility lot or control number.
e. Date of sterilization.
f. Quantity of product in sterilizer lot.
g. Type of sterilization (gas, radiation, heat, steam, etc.).
h. Form of biological indicator (i.e., inoculated paper spore strip or carrier, inoculated product).
i. Test organism(s) and population density used.
j. Test medium, incubation temperature, incubation time, and dates of tests.
k. Number of test units.
l. Statement that samples comply with sterility tests.
m. Signature of head of a laboratory or a designated qualified alternate.

When inspection is made at destination, the certificate of sterility shall be submitted by the contractor to the Directorate of Medical Materiel, ATTN: DPSC-ATQ, Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, PA 19101, and a copy forwarded with each shipment to the consignee. When inspection is made at source, certificate of sterility shall be furnished to the cognizant Government Quality Assurance Representative for submission to the procuring activity.

"3.3.2 Sterility lot. A sterility lot is that single quantity of product subjected to the same manufacturing operation and simultaneously sterilized in the same sterilizer chamber. Each unit package shall be clearly marked with a sterility lot (control) number and date of sterilization. Each sterility lot shall be traceable to the pertinent manufacturing lot(s)."

PAGE 3

3.4 Class 2, camouflaged. Delete last sentence and substitute "The gauze shall be dyed brown with a suitable non-toxic dye with a resultant color similar to 0.125 percent pontamine brown B.T. The dye shall be colorfast."

3.5.2 Width. At end of paragraph add: "Width of type II, size g bandage shall average 3 3/4 inches with no bandage less than 3 11/16 inches in width."

3.6 Compression. To list add: "Size g - 4 by 1 1/2 by 1/2 inches."

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:32Z
Check the source to verify that this is the current version before use.

Delete paragraph 4.1 and substitute:

"4.1 Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years fter delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the gauze and pins may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements."

PAGE 4

4.2.2 Examination of end item. At end of paragraph, add: "The AQL shall be 6.5 for Minor B Defects." Fourth line insert "A" between "Minor" and "Defects". In table I column headings insert "A" after "Minor" and add the heading "Minor B". Reclassify the defect "Bandage not clean (not free of loose threads, raveling or lint)" from Major to Minor B. Below table I add "examination shall not be limited to the defects listed."

4.2.4 Tests. Delete fifth sentence and substitute "For sterility tests, the sample size shall be as specified in paragraph 4.2.4.2." Delete the fourth sentence and substitute: "The acceptance number shall be zero for all samples sizes."

Source: https://assist.dla.mil -- Downloaded: 202610-26T12:32Z
Check the source to verify that this is the current version before use.

4.2.4.1 Breaking strength. At end of paragraph add: "For bandages of one inch width, the warp breaking strength shall be performed on the gauze prior to cutting."

4.2.4.2 and 4.2.4.2.1 Delete entirely and substitute:

"4.2.4.2 Sterility test. The bandage shall be tested for sterility in accordance with the method specified in the USP for purified cotton, gauze, surgical dressing and related material using inoculated product, or paper spore strip (carrier). Product containing a biological indicator shall be marked or labeled so that it clearly differentiates from product intended for distribution. The incubation medium and minimum time in each case shall be soybean casein digest for 7 days for inoculated product, inoculated simulated product, or paper spore strip (carrier). The incubation temperature shall be as indicated in 4.2.4.2.1. The minimum number of test units, regardless of lot size, shall be 10, except for inoculated carrier which shall be 20 for type I sterilization and 30 for type II sterilization including inoculated carrier and product test units. The inspection level shall be S-3. The biological indicator shall be located within the product where the material is most dense and most inaccessible to the killing agent.

"4.2.4.2.1 Inoculated product. For each type of sterilization, the organism used shall be as listed below:

Type of Sterilization	Organism	Incubation Temperature
Steam	B. stearothermophilus	55° - 65°C
Dry Heat	B. stearothermophilus, or	55° - 65°C
	B. subtilis var. niger (or globigii)	30° - 35°C
Gas	B. subtilis var. niger (or globigii)	30° - 35°C
Radiation	B. pumilus	30° - 35°C

The number of viable cells shall be sufficient to provide assurance that the product has been sterilized.

"4.2.4.2.3 Inoculated carrier. Inoculated carriers (paper spore strips) if used, shall be in accordance with MIL-S-36586. Other paper carriers shall be inoculated to be of equivalent resistance. The method of incubation and number of test units shall be as indicated in 4.2.4.2 Sterility test and 4.2.4.2.1 Inoculated product. The method of sterilization will control the organism in the spore strip to be tested for as shown in 4.2.4.2.1. NOTE: When an inoculated carrier is chosen, product samples must also be tested as required by the USP.

Source: https://assist.dla.mil -- Downloaded: 202610-26T12:32Z
Check the source to verify that this is the current version before use.

"4.2.4.2.4 Alternate method. Suppliers conducting sterility tests or designating lots in a manner other than that indicated herein shall submit proof that the alternate method provides equal assurance concerning the sterility of the product. However, all sterility tests shall be conducted using biological indicators. Requests for approval of any alternate method of testing for sterility shall be made to the Contracting Officer, and approval obtained before employment of the alternate method on a government contract.

"4.2.4.2.5 Product testing. Testing for sterility of the product after receipt by the government shall be without a biological indicator and in accordance with the USP."

PAGE 5

Table II, Packaging quantities: In column 5, in lines 8 thru 10, delete "576, 432 and 240" and substitute "288"; also in column 6, delete "1152, 864 and 480" and substitute "576".

PAGE 6

5.2.1.1.1 Wrapper. Line 1, following "type I" insert ", class 1"; also in line 3, following "II" insert "and type I, class 2".

5.2.1.1.2 Bag (applicable to type I, class 2 and type II, class 2). In line 4, delete "size g" and substitute: "sizes c, e and g".

5.2.1.3 Protective pack. Line 3, delete "weather-resistant, grades W5C or W5S" and substitute "domestic".

5.3.1 Level A. Line 3, delete "grade V3C or V3S". Delete the last sentence and substitute: "Closure, waterproofing and strapping shall be as specified in the appendix of the box specification."

5.3.2 Level B. Delete the second sentence and substitute: "Closure shall be as specified in the appendix of the box specification."

PAGE 7

5.4.2.1 Applicable to type II, class 2. Lines 2 and 3, following "name of contractor" insert "contract or purchase order number and the requirements of the Federal Food, Drug and Cosmetic Act."

5.4.3 Individual carton (applicable to type I, class 1). Delete and substitute: "5.4.3 Individual box (applicable to type I, class 1). Each individual box shall be marked in accordance with the requirements of the Federal Food, Drug and Cosmetic Act. In addition, marking shall include the item identification, size, lot or control number, name and address of the manufacturer and the legends: "STERILE" and "STERILITY GUARANTEED UNLESS PACKAGE IS DAMAGED OR OPENED" or similar statement.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:32Z
Check the source to verify that this is the current version before use.

5.4.4.1 Applicable to type I, class 1 and 2. Delete and substitute:

"5.4.4.1 Applicable to type I class 1 and 2. Each unit box shall be marked in accordance with MIL-STD-129 and the requirements of the Federal Food, Drug and Cosmetic Act. Sterility lot (control) number, contract or purchase order and name of contractor shall be shown. The method of sterilization as specified in 5.4.2.1 shall also be shown."

5.4.5 Intermediate package. Delete and substitute:

"5.4.5 Intermediate package. Each intermediate package shall be marked as specified in MIL-STD-129. The lot (control) number and method of sterilization as specified in 5.4.2.1 shall be shown."

PAGE 8

Add the following:

"6.3 This specification covers the following items appearing in the Federal Supply Catalog:

National Stock No.	Item Identification	Type	Class	Size
6510-00-200-3180	BANDAGE, GAUZE, Compressed, Camouflaged, 2 inches by 6 yards	II	2	C
6510-00-200-3185	BANDAGE, GAUZE, Compressed, Camouflaged, 3 inches by 6 yards	II	2	e
6510-00-200-3190	BANDAGE, GAUZE, Compressed, Camouflaged, 4 inches by 6 yards	II	2	g
6510-00-200-3985	BANDAGE, GAUZE, Roller, 1 inch by 6 yards, 12s	I	1	a
6510-00-200-4000	BANDAGE, GAUZE, Roller, 2 inches by 6 yards, 12s	I	1	c
6510-00-200-5000	BANDAGE, GAUZE, Roller, 3 inches by 10 yards, 12s	I	1	f
6510-00-200-6000	BANDAGE, GAUZE, Roller, 4 inches by 10 yards, 12s	I	1	h

Source: https://assist.dla.mil -- Downloaded: 20261O-26T12:32Z
Check the source to verify that this is the currentversion before use.

National Stock No.	Item Identification	Type	Class	Size
6510-00-200-6204	BANDAGE, GAUZE, Roller, Camouflaged, 2 inches by 6 yards, 12s	I	2	c
6510-00-200-6206	BANDAGE, GAUZE, Roller, Camouflaged, 3 inches by 10 yards, 12s"	I	2	f

Preparing activity:
DLA - DM

Project No. 6510-0629

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:32Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:32Z
Check the source to verify that this is the current version before use.