DDD-P-0053c(GSA-FSS)
October 14, 1963
Interim Revision of
Fed. Spec. DDD-P-53b
December 9, 1957

INTERIM FEDERAL SPECIFICATION

PAD, ABDOMINAL (DRAINAGE DRESSING, COMBINATION)

This Interim Federal Specification was developed by General Services Administration, Federal Supply Service, Washington 25, D. C., based upon currently available technical information. It is recommended that Federal agencies use it in procurement and forward recommendations for changes to the preparing activity at the address shown above. The General Services Administration has authorized Federal agencies to use this Interim Federal Specification as a valid exception to Federal Specification DDD-P-53b.

1. SCOPE AND CLASSIFICATION

1.1 Scope. This specification covers sterile and non-sterile abdominal pad combinations used as dressings to absorb drainage during operations and treatment of patients.

1.2 Classification.

1.2.2 Types, classes and sizes. Abdominal pads in accordance with this specification, shall be of the following types, classes and sizes.

Type I - Non-absorbent
Type II - All absorbent

Classes
Class A - Steril (individually pre-wrapped)
Class B - Non-sterile (bulk)

Sizes
5 x 9 inches
7½ x 8 inches
8 x 10 inches
10 x 12 inches
12 x 16 inches

2. APPLICABLE SPECIFICATIONS AND STANDARDS

2.1 The following specifications and standards, of the issues in effect on date of invitation for bids, form a part of this specification:

Federal Specifications:
L-C-166 Cellulose, Absorbent, Surgical.
CCC-G-101 Gauze, Absorbent.
JJJ-C-561 Cotton; Sterile and Non-sterile.
PPP-B-636 Boxes, Fiber.

Federal Standards:
Fed. Std. No. 102 Preservation, Packaging and Packing Levels.
Fed. Std. No. 123 Marking for domestic shipment (civilian agencies)

(FOR INSTRUCTIONS ON OBTAINING COPIES OF FEDERAL SPECIFICATIONS AND STANDARDS, SEE LAST PAGE)

FSC 6510

Source: https://assist.dla.mil -- Downloaded: 202610-26T12:31Z
Check the source to verify that this is the current version before use.

Military Standards:
MIL-STD-105 Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 Marking for Shipment and Storage.

(Copies of Military Specifications and Standards required by the contractor in connection with specific procurement functions, should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following publications, of the issue in effect on date of invitation for bids, form a part of this specification.

U. S. Pharmacopeial Convention Publication:
The Pharmacopeia of the United States (latest revision).

(Copies are available from Mack Publishing Company, Easton, Pennsylvania).

3. REQUIREMENTS

3.1 Material.

3.1.1 Gauze. The gauze shall be 14/10 commercial designation, bleached gauze and shall conform to the requirements of Federal Specification CCC-G-101; except in regard to thread count. A variation in the average count of three threads to the inch in the warp and two threads in the filling, but not more than four threads to the square inch shall be allowed. Gauze shall weigh 11.5 grams per square yard with an allowable variation of ± 12 percent.

3.1.2 Non-Woven Fabric. The non-woven fabric shall consist of all cotton, all rayon or a mixture of cotton and rayon plus a binder. The binder shall be non-toxic. The fabric shall have either a regular pattern of openings or a non-woven construction of adequate porosity to permit penetration of fluid. It shall not adhere to itself when steam sterilized. The weight of the fabric shall be not less than 150 grains per square yard. When tested as specified in the U. S. P., the water-soluble substances of the fabric shall not exceed 0.75% and the ignited residue of the water extract shall not exceed 0.20%. The fabric shall be free of acid, alkali, starch or dyes. The fatty matter of the fabric shall not exceed 0.7%. The ash of the fabric shall not exceed 3.5%.

3.1.3 Bleached Absorbent Fiber. Fiber shall be fully bleached, made from purified new or new reworked cotton (e.g., comber or cardstrip); or rayon; or fully bleached cotton linters; or cotton or rayon reworked from clean yarn threads or fabric; or a mixture of these types. The absorbent fiber shall be white, substantially free from lumps, oil spots, dirt and other foreign material. When tested as specified in 4.4.2, the bleached fiber shall be completely submerged in 15 seconds and shall absorb not less than 16 times its weight of water. The fiber shall be free of acid, alkali or dyes. The water-soluble substances of the fiber shall not exceed 1.00%. The ash shall not exceed 2.0%.

3.1.4 Unbleached Nonabsorbent Cotton. Nonabsorbent cotton shall be unbleached, made from new or reworked cotton or cotton fibers that has been reworked from clean unbleached yarn, threads or fabric or blend of these types. The mixture, however, shall contain not more than 50% cotton linters.

3.1.5 Absorbent cellulose wadding. Shall be a clean grade of absorbent cellulose wadding conforming to the requirements of L-C-166.

3.2 Description.

3.2.1 Type I. Type I non-absorbent abdominal pads shall be constructed of a layer of bleached absorbent fibers, or a layer of bleached absorbent fibers backed by absorbent cellulose wadding, (see 3.2.3) backed by a layer of un-bleached non-absorbent cotton and covered with gauze or non-woven fabric.

3.2.2 Type II. Type II absorbent abdominal pads shall be constructed of bleached absorbent fiber covered with gauze or non-woven fabric. A layer of absorbent cellulose wadding may be placed between layers of the bleached absorbent fiber (see 3.2.3).

3.2.3 Absorbent layers. The absorbent sections used in the construction of type I or type II abdominal pads shall be not less than 25% by weight bleached absorbent fibers.

Source: https://assist.dla.mil -- Downloaded: 20261O-26T12:31Z
Check the source to verify that this is the current version before use.

3.2.4 Covering of Gauze or Non-Woven Fabric. The covering shall overlap at the backing side at least 1-1/2 inches, or shall be secured by a bland moisture resistant flexible adhesive which will withstand normal sterilization and handling without breakdown of the seal. The covering shall be the same length as the pad and shall cover all surfaces except the cut ends.

3.3 When specified, a printed set of instructions for identifying the nonabsorbent side of the pads shall be enclosed in each shipping case. (see 6.3).

3.4 Nonabsorbent backing. The weight of the nonabsorbent backing shall be as specified in Table I, when tested in accordance with 4.4.4.

TABLE I. Weight of nonabsorbent backing

Size	Weight
	Grams
5 x 9	1.1 ± .5
7½ x 8	1.9 ± .7
8 x 10	3.3 ± 1.3
10 x 12	4.8 ± 1.6
12 x 16	8.0 ± 2.6

3.5 Dimensions and weight. The dimensions and weight of the completed pads shall be as specified in table II, when tested in accordance with 4.4.5.

TABLE II. Dimensions and weight of completed pads.

Size Nominal	Length minimum	Width minimum	Weight, minimum
	Inches	Inches	Grams
5 x 9	4-3/4	8-1/2	8
7½ x 8	6-3/4	7-3/4	13
8 x 10	7-3/4	9-1/2	17
10 x 12	9-3/4	11-1/2	26
12 x 16	11-3/4	15-1/2	43

3.6 Sterility. Each class A pad shall be pre-packaged, and sterilized by any commercially acceptable procedure. The packaging shall be adequate to maintain sterility until opened. Each class A pad shall be sterile when tested as specified in 4.4.6. (see 4.1)

3.6.1 Certificate of Sterility. When specified by the procuring authority, a certificate of sterility, listing all lot or control numbers and stating that samples from each sterilizer charge have been tested in accordance with U. S. P. methods and found to be sterile, shall be forwarded by the contractor to the procuring activity.

3.7 Labeling. Commercial Markings are acceptable on paper container if name of item, size, contents, name of manufacturer and a suitable inscription denoting contents are sterile, are included.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:31Z
Check the source to verify that this is the current version before use.

3.8 Workmanship. Shall be first class throughout. The combination pads shall be free from defects which may detract from their appearance or which may impair their serviceability.

4. SAMPLING, INSPECTION, AND TEST PROCEDURES

4.1 Responsibility for inspection. Unless otherwise specified in the contract of purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection, as used in this specification, is defined as both examination (such as visual, auditory, etc., investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be in accordance with instructions furnished by the procuring activity.

4.2.2 For tests. Sample size for each test shall be as indicated for inspection level L4 in Appendix to MIL-STD-105, except that the acceptance number shall be zero (0) for all sample sizes, and a minimum of two (2) samples shall be taken.

4.3 Tests. Tests shall be conducted, as necessary to determine compliance with specification requirements.

4.4 Tests.

4.4.1 Gauze, and Non-Woven Fabric. Shall be performed as described in CCC-G-101.

4.4.2 Absorbent Fiber. Shall be performed as described in JJJ-C-561. Applicable tests are for absorbency, acid and alkali, dyes, water-soluble substances, and ash.

4.4.3 Absorbent cellulose wadding. Shall be performed as described in L-C-166.

4.4.4 Nonabsorbent backing. The weight of the backing shall be determined by weighing only the nonabsorbent backing. The number of samples tested shall be in accordance with 4.2.2.

4.4.5 Dimensions and weight. The dimensions (in inches) and weight (in grams) of the completed pad shall be determined by measuring and weighing a number of pads in accordance with 4.2.2.

4.4.6 Sterility. Class A pads shall be tested for sterility as described in the Pharmacopeia of the United States.

4.5 Examination. The pads shall be examined to determine compliance with all requirements contained in this specification

5. PREPARATION FOR DELIVERY
(For civil agency procurement the definitions and applications of levels of packaging and packing shall be in accordance with Fed. Std. No. 102).

5.1 Packaging. Packaging shall be level A, B, or C as specified (see 6.1).

5.1.1 Level A. The individual pre-wrapped sterile pads (see 3.6) shall be unit packaged, in quantities shown in Table III, in a box conforming to PPP-B-636, class 2, W6c, style RSC, or to PPP-B-665, style optional. Each box shall be closed and sealed in accordance with the applicable specification or appendix thereto and waterproofed in accordance with MIL-P-116, method IC-2. The non-sterile pads shall require no unit packaging.

5.1.2 Level B. The pads shall be packaged as specified in 5.1.1 except that waterproofing of the boxes in accordance with MIL-P-116 shall not be required.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:31Z
Check the source to verify that this is the current version before use.

5.1.3 Level C. The pads shall be packaged in accordance with the suppliers commercial practice.

5.1.4 Intermediate packaging. Unless otherwise specified (see 6.1) the individual pre-wrapped sterile pads shall be intermediate packaged in boxes conforming to PPP-B-636, class 2, W6c, style RSC. Each box shall be closed and waterproof sealed with tape in accordance with the appendix to the box specification. Quantities shall be as shown in Table III. The non-sterile pads shall require no intermediate packaging.

5.2 Packing. Packing shall be level A, B, or C as specified (see 6.1).

5.2.1 Level A. The individual pre-wrapped sterile pads and non-sterile pads shall be separately packed in a box conforming to PPP-B-636, class 2, V3c, style RSC. Each box shall be closed, waterproof sealed with tape, and reinforced with bonds of filament tape in accordance with the appendix to the box specification. The quantities shall be as shown in Table III.

5.2.2 Level B. The pads shall be packed as specified in 5.2.1 except that waterproof sealing with tape shall not be required.

5.2.3 Level C. The pads shall be packed to insure carrier acceptance and safe delivery to destination in containers complying with the rules and regulations applicable to the mode of transportation.

5.3 Marking.

5.3.1 Civil agencies. In addition to markings required by the contract or order, the interior packages and shipping containers shall be marked in accordance with Fed. Std. No. 123.

5.3.2 Military activities. In addition to markings required by the contract or order, the interior packages and shipping containers shall be marked in accordance with MIL-STD-129.

TABLE III

Type	Class	Size	No. of Pads in Each Shipping Container
I (non-absorbent)	A (sterile)	5 x 9	400
(non-absorbent)	(sterile)	7½ x 8	500
(non-absorbent)	(sterile)	8 x 10	400
(non-absorbent)	(sterile)	10 x 12	400
(non-absorbent)	(sterile)	12 x 16	400
(non-absorbent)	B (non-sterile)	5 x 9	500
(non-absorbent)	(non-sterile)	7½ x 8	500
(non-absorbent)	(non-sterile)	8 x 10	500*
(non-absorbent)	(non-sterile)	10 x 12	288
(non-absorbent)	(non-sterile)	12 x 16	200**
(non-absorbent)	(non-sterile)	12 x 16	72
II (all absorbent)	A (sterile)	5 x 9	400
(all absorbent)	(sterile)	7½ x 8	500
(all absorbent)	(sterile)	8 x 10	400
(all absorbent)	(sterile)	10 x 12	400

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:31Z
Check the source to verify that this is the current version before use.

TABLE III - Continued

Type	Class	Size	No. of Pads in Each Shipping Container
II (all absorbent)	A (sterile)	12 x 16	400
(all absorbent)	B (non-sterile)	5 x 9	576
(all absorbent)	(non-sterile)	7½ x 8	576
(all absorbent)	(non-sterile)	8 x 10	400
(all absorbent)	(non-sterile)	10 x 12	300
(all absorbent)	(non-sterile)	12 x 16	200

* 50 pads in a unit package
5 unit packages in an intermediate package
2 intermediate packages in a shipping container

** 20 pads in a unit package
5 unit packages in an intermediate package
2 intermediate packages in a shipping container

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Title, number, and date of this specification.
(b) Size required (see 1.2)
(c) Instructions required (see 3.3).
(d) Inspection required (see 4.2).
(e) Place of final inspection (see 4.2).
(f) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 It is believed that this specification adequately describes the characteristics necessary to secure the desired material, and that normally no samples will be necessary prior to award to determine compliance with this specification. If, for any particular purpose, samples with bids are necessary, they should be specifically asked for in the invitation for bids, and the particular purpose to be served by the bid sample should be definitely stated, the specification to apply in all other respects.

6.3 This specification does not provide for printing or other identification on the surface of the pad itself, and purchasers should specify if a set of instructions for identifying the non-absorbent side of the pad is desired (see 3.3).

6.4 This specification does not include all types, classes, grades, etc., of the commodity indicated by the title of the specification, or which are commercially available, but is intended to cover the type, etc., normally procured to meet Federal Government requirements.

6.5 Transportation description. Transportation description applicable to this item is:

Dressings, Surgeons'.
Carload minimum weight 24,000 pounds,
subject to Rule 34, Uniform Freight Classification.
Motor volume minimum weight 30,000 pounds.

Notice: See last page.

MILITARY INTEREST
Defense Supply Agency - DMSC.
Army - Med.
Navy - BuMED.
Air Force - AFCSG.

GSA WASH DC 64-5948

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:31Z
Check the source to verify that this is the current version before use.

ADDENDUM

INSTRUCTIONS ON OBTAINING COPIES OF FEDERAL SPECIFICATIONS AND STANDARDS

(Activities outside the Federal Government may obtain copies of Federal Specifications, Standards, and Handbooks as outlined under General Information in the Index of Federal Specifications, Standards and Handbooks and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U. S. Government Printing Office, Washington 25, D. C.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Washington, D. C., Atlanta, Chicago, Kansas City, Mo., Dallas, Denver, San Francisco, and Auburn, Wash.

(Federal Government activities may obtain copies of Federal Specifications, Standards, and Handbooks and the Index of Federal Specifications, Standards and Handbooks from established distribution points in their agencies.)

NOTICE

When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data is not to be regarded by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:31Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 119-R004

INSTRUCTIONS

This sheet is to be filled out by personnel either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity *(as indicated on reverse hereof)*.

SPECIFICATION

ORGANIZATION *(Of submitter)*	CITY AND STATE

CONTRACT NO.	QUANTITY OF ITEMS PROCURED	DOLLAR AMOUNT
		$

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES.

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO IF "YES", IN WHAT WAY?

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity)*	DATE

DD, FORM 1 APR 63 1426 REPLACES NAVSHIPS FORM 4863, WHICH IS OBSOLETE C-8878

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:31Z
Check the source to verify that this is the current version before use.

FOLD

DEPARTMENT OF THE NAVY

POSTAGE AND FEES PAID
NAVY DEPARTMENT

OFFICIAL BUSINESS

GENERAL SERVICES ADMINISTRATION
FEDERAL SUPPLY SERVICE
PROGRAM DEVELOPMENT BRANCH
ROOM 4921, REGIONAL OFFICE
7TH STREET (S.W.)
WASHINGTON, D. C.

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:31Z
Check the source to verify that this is the current version before use.