DDD-P-53b

December 9, 1957

SUPERSEDING
Int. Fed. Spec. DDD–P–0053a (GSA–FSS)
August 23, 1956 and
Fed. Spec. DDD–P–53
November 1, 1946

FEDERAL SPECIFICATION

PAD, ABDOMINAL (DRAINAGE DRESSING COMBINES)

This specification was approved by the Commissioner, Federal Supply Service, General Services Administration, for the use of all Federal agencies.

1. SCOPE AND CLASSIFICATION

1.1 **Scope.**—This specification covers abdominal pad combines used as dressings to absorb drainage during operations and treatment of patients.

1.2 **Classification.**

1.2.1 *Type and sizes.*—Combination pads in accordance with this specification shall be of one type and of the following sizes, as specified (see 6.2).

Size 1.— 8 by 10 inches.

Size 2.—10 by 12 inches.

Size 3.—12 by 16 inches.

2. APPLICABLE SPECIFICATIONS AND STANDARDS

2.1 The following specifications and standards, of the issues in effect on date of invitation for bids, form a part of this specification:

Federal Specifications:

L–C–166—Cellulose, Absorbent, Surgical.

CCC–G–101—Gauze, Absorbent.

JJJ–C–561—Cotton; Sterile and Nonsterile.

PPP–B–636—Boxes, Fiber.

Federal Standards:

Fed. Std. No. 102—Preservation, Packaging and Packing Levels.

(Activities outside the Federal Government may obtain copies of Federal Specifications and Standards as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U. S. Government Printing Office, Washington 25, D. C.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Atlanta, Chicago, Kansas City, Mo., Dallas, Denver, San Francisco, Los Angeles, Seattle, and Washington, D. C.

(Federal Government activities may obtain copies of Federal Specifications and Standards and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

Military Standards:

MIL–STD–105 — Sampling Procedures and Tables For Inspection by Attributes.

MIL–STD–129 — Marking for Shipment and Storage.

(Copies of Military Specifications and Standards required by the contractor in connection with specific procurement functions, should be obtained from the procuring activity or as directed by the contracting officer.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:31Z
Check the source to verify that this is the current version before use.

3 REQUIREMENTS

3.1 Material.

3.1.1 *Gauze.*—The gauze shall be $\frac{14}{10}$ commercial designation bleached gauze and shall conform to the requirements of Federal Specification CCC–G–101, except in regard to thread count.

3.1.2 *Bleached cotton.* — Cotton shall be fully bleached, made from purified new or new reworked cotton (e.g., comber or card-strip); white, fully bleached cotton linters; or, cotton which has been reworked from clean yarn, threads, or fabric; or, a mixture of these types. No more than 25 percent of the aggregate shall be reworked yarn, threads, or fabric. The bleached cotton shall be substantially free from lumps, oilspots, dirt, and all other foreign material. When tested as specified in 4.3.2, the bleached cotton shall be completely submerged in 15 seconds and shall absorb not less than 16 times its weight of water; the cotton shall be free of acid, alkali, or dyes; ash shall be not greater than 0.5 percent; water-soluble substances shall be not greater than 0.75 percent of the weight of the cotton.

3.1.3 *Unbleached nonabsorbent cotton.* — Nonabsorbent cotton shall be unbleached, made from new or reworked cotton, and shall contain not more than 50 percent cotton linters. The cotton shall be substantially free from lumps, oil spots, streaks of dirt, or similar foreign matter.

3.1.4 *Absorbent cellulose wadding.*—Shall be a clean grade of absorbent cellulose wadding conforming to the requirements of Federal Specification L–C–166.

3.2 **Description.** — The combination pads shall consist of a layer of bleached absorbent cotton with a backing layer of unbleached nonabsorbent cotton, with a covering of bleached gauze. Not more than 75 percent of the bleached absorbent cotton may be replaced by absorbent cellulose, in which case it shall be placed between the layers of bleached absorbent and unbleached nonabsorbent cotton.

3.2.1 *Gauze covering.*—The gauze covering shall overlap on the unbleached nonabsorbent backing side at least 1½ inches, or shall be secured by a bland moisture-resistant flexible adhesive which will withstand normal sterilization and handling without breakdown of the seal. The gauze shall be the same length as the pad and shall cover all surfaces except the cut ends.

3.3 When specified, a printed set of instructions for identifying the nonabsorbent side of the pads shall be enclosed in each shipping case. (See 6.4.)

3.4 **Nonabsorbent backing.**—The weight of the nonabsorbent backing shall be as specified in table I, when tested in accordance with 4.3.4.

TABLE I. — *Weight of nonabsorbent backing*

Size	Weight
	Grams
1	3.3 ± 1.3
2	4.8 ± 1.6
3	8.0 ± 2.6

3.5 **Dimensions and weight.**—The dimensions and weight of the completed pads shall be as specified in table II, when tested in accordance with 4.3.5.

TABLE II. — *Dimensions and weight of completed pads.*

Size	Length, minimum	Width, minimum	Weight, minimum
	Inches	*Inches*	*Grams*
1	9½	7¾	17
2	11½	9¾	26
3	15½	11¾	43

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:31Z
Check the source to verify that this is the current version before use.

3.6 **Workmanship.** — Shall be first class throughout. The combination pads shall be free from defects which may detract from their appearance or which may impair their serviceability.

4. SAMPLING, INSPECTION, AND TEST PROCEDURES.

4.1 **Sampling.**—Sampling procedures shall be conducted in accordance with the procedures set forth in Military Standard MIL–STD –105. Destructive testing shall be in accordance with Appendix to Military Standard MIL–STD–105. In addition, samples may be requested by the procuring activity for testing at Government laboratory.

4.2 **Inspection.** — Inspection may be made throughout the entire process of manufacture. The passing as satisfactory of any detail of construction or materials shall not relieve the contractor of responsibility for faulty workmanship or materials which may be discovered at any time prior to final acceptance. Final inspection of the finished article shall be made either at point of production or at point of delivery, as specified (see 6.2). In case of factory inspection, every facility shall be afforded inspectors, by the manufacturer, for the prosecution of their work.

4.3 **Tests.**

4.3.1 *Gauze.* — Shall be performed as described in Federal Specification CCC–G–101.

4.3.2 *Absorbent cotton.* — Shall be performed as described in Federal Specification JJJ–C–561. Applicable tests are for absorbency, acid and alkali, dyes, water-soluble substances, and ash.

4.3.3 *Absorbent cellulose wadding.*—Shall be performed as described in Federal Specification L–C–166.

4.3.4 *Nonabsorbent backing.*—The weight of the backing shall be determined by weighing only the nonabsorbent backing. The number of samples tested shall be in accordance with 4.1.

4.3.5 *Dimensions and weight.*—The dimensions (in inches) and weight (in grams) of the completed pad shall be determined by measuring and weighing a number of pads in accordance with 4.1.

5. PREPARATION FOR DELIVERY

5.1 **Packaging.**

5.1.1 *Unit of issue.*—The number of pads specified in table III constitute one unit.

5.1.2 *Packaging quantities.* — The number of units of issue indicated in table III shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in procurement document.

TABLE III.—*Packaging quantities*

Federal stock number	Size	Number of pads per unit	Number of unit packages in each intermediate package	Number of intermediate packages in shipping containers	Total number of units in shipping container
6510–203–2250 ----	1	50	5	2	10
6510–203–2270 ----	3	20	5	2	10

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:31Z
Check the source to verify that this is the current version before use.

5.1.2.1 *Packing variation permitted.* — If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.1.3 *Level A.*

5.1.3.1 *Unit package.*—Each unit shall be packaged in a bag fabricated from virgin polyethylene seamless tubing having an average wall thickness of 0.004-inch with no wall thickness at any point below 0.00325 inch. Excess air shall be expelled or evacuated, and closure effected by heat sealing.

5.1.3.2 *Intermediate package.*—Intermediate package shall be a fiberboard box constructed in accordance with Federal Specification PPP–B–636, compliance symbol W5c or W5s, except that the manufacturer's joint shall be effected by the application of a water-resistant tape, or by gluing integral lap to outer face of end or side panel using water-resistant adhesive, in lieu of metal stitches. Closure shall comply with appendix thereto.

5.1.4 *Level C.*—Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose within the continental limits of the United States of America.

5.2 Packing.

5.2.1 *Level A.*

5.2.1.1 *Exterior container.*—Exterior container shall be a fiberboard box constructed in accordance with Federal Specification PPP–B–636, compliance symbol V3c or V3s. Closure shall comply with appendix thereto, except that strapping will not be required.

5.2.2 *Level C.* — The subject commodity shall be packed in substantial commercial containers of the type, size, and kind, commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 *Unit package.* — Each unit package shall be marked in accordance with section 5 of Military Standards MIL–STD–129. Preservation and date marking are not required. The required information shall be printed directly on the bag or on a label inserted within the bag in such manner as to permit ready identification.

5.3.2 *Intermediate package.* — Each intermediate package shall be marked in accordance with section 5 of Military Standard MIL–STD–129.

5.3.3 *Shipping container.* — Shipping container shall be marked in accordance with Military Standard MIL–STD–129. Service color and service symbol markings are required only on direct shipments from the contractor's plant to overseas activities.

6. NOTES

6.1 Federal Standard No. 102 should be referred to for definitions and applications of the various levels of packaging protection for supplies and equipment.

6.2 **Ordering data.** — Procurement documents should specify the following:

(a) Title, number, and date of this specification.
(b) Size required (see 1.2).
(c) Instructions required (see 3.3).
(d) Inspection required (see 4.2).
(e) Place of final inspection (see 4.2).
(f) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:31Z
Check the source to verify that this is the current version before use.

6.3 It is believed that this specification adequately describes the characteristics necessary to secure the desired material, and that normally no samples will be necessary prior to award to determine compliance with this specification. If, for any particular purpose, samples with bids are necessary, they should be specifically asked for in the invitation for bids, and the particular purpose to be served by the bid sample should be definitely stated, the specification to apply in all other respects.

6.4 This specification does not provide for printing or other identification on the surface of the pad itself, and purchasers should specify if a set of instructions for identifying the nonabsorbent side of the pad is desired (see 3.3).

6.5 This specification does not include all types, classes, grades, etc., of the commodity indicated by the title of the specification, or which are commercially available, but is intended to cover the type, etc., normally procured to meet Federal Government requirements.

6.6 **Transportation description.**—Transportation description applicable to this item is:

Dressings, Surgeons'.
Carload minimum weight 24,000 pounds, subject to Rule 34, Uniform Freight Classification.
Motor volume minimum weight 30,000 pounds.

6.7 This specification covers the following items appearing in the Armed Services Medical Stock List:

6510–203–2250—Pad, Abdominal, 8 by 10 inches, 50s.

6510–203–2270—Pad, Abdominal, 12 by 16 inches, 20s.

Notice. — When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility, nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data, is not to be regarded by implication or otherwise, as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

MILITARY INTEREST:
Army—M
Navy—Md
Air Force: USAF.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:31Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:31Z
Check the source to verify that this is the current version before use.