DATA ITEM DESCRIPTION

Form Approved
OMB No. 0704-0188

Public reporting burden for this collection of information is estimated to average 110 hours per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing the collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden to Washington Headquarters Services, Directorate for Information Operations and Reports, 1215 Jefferson Davis Highway, Suite 1204, Arlington, VA 22202-4302, and to the Office of Management and Budget, Paperwork Reduction Project (0704-0188), Washington, D.C. 20503.

1. TITLE	2. IDENTIFICATION NUMBER
Calibration Certificate	DI-QCIC-80798A

3. DESCRIPTION/PURPOSE

3.1 The Calibration Certificate documents the performance of calibration in accordance with contract specifications for acquired equipment.

4. APPROVAL DATE (YYMMDD)	5. OFFICE OF PRIMARY RESPONSIBILITY (OPR)	6a. DTIC APPLICABLE	6b. GIDEP APPLICABLE
073097	F-AFMETCAL/MLEE		

7. APPLICATION/INTERRELATIONSHIP

7.1 This Data Item Description (DID) contains the format and content preparation instructions for the data product generated by the specific and discrete task requirement as delineated in the contract.

7.2 This DID is applicable to contracts for equipment requiring certified calibration prior to delivery.

7.3 This DID supersedes DI-QCIC-80798.

8. APPROVAL LIMITATION	9a. APPLICABLE FORMS	9b. AMSC NUMBER
		F7284

10. PREPARATION INSTRUCTIONS

10.1 Format. Contractor format is acceptable.

10.2 Content. The Calibration Certificate shall include:

a. Title, e.g. "Calibration Certificate" or "Calibration Report".

b. Name and address of laboratory, and location where the calibration was carried out, if different from the address of the laboratory.

c. Unique identification of the certificate or report (such as serial number), each page, and total number of pages.

(Continued on Page 2)

11. DISTRIBUTION STATEMENT

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:40Z
Check the source to verify that this is the current version before use.

Block 10, Preparation Instructions (Continued)

d. Name and address of customer, when appropriate.
e. Description of the item calibrated, including model number and serial number.
f. Condition of the calibration item.
g. Date (s) of performance of calibration, when appropriate.
h. Identification of the calibration procedures used, or description of any non-standard method used.
i. Any deviation from, additions to, or exclusions from the calibration method, and information relevant to a specific calibration, such as environmental conditions.
j. Measurements, examinations, and derived results supported by tables, graphs, sketches, and photographs, as appropriate, and any failures identified.
k. A statement of the estimated uncertainty of the calibration results (See ISO Guide to the Expression of Uncertainty of Measurement).
l. A signature and title, or equivalent identification of the person(s) assuming responsibility for the content of the certification or report, and date of issue.
m. Where relevant, a statement to the effect that the results relate only to the items calibrated.
n. A statement that the certificate or report shall not be reproduced except in full, without the written approval of the laboratory.
o. Special limitations of use.
p. A traceability statement, including a complete description of all standards in the chain of traceability to national standards. NOTE: When the certificate or report contains results of calibrations performed by subcontractors, these results shall be clearly identified.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:40Z
Check the source to verify that this is the current version before use.