FED SUP CLASS
6630

| PART NUMBER | DESCRIPTION | ELECTRODE LENGTH (CM) | TEMP RANGE °C | APPLICATION |
|---|---|---|---|---|
| MS36585-1 | Dual Platinum Plate | 11.4 | -10 to 80 | Chlorine Titration |
| MS36585-2 | Dual Platinum | 13.0 | -5 to 100 | KF Titrations, Sulfur Analysis |
| MS36585-3 | Platinum/AG/ AGCL | 13.0 | -5 to 100 | Oxidation Reduction |
| MS36585-4 | Silver | 12 7 | -5 to 100 | Titration of Halogens |
| MS36585-5 | Silver Billet | 15.8 | -5 to 100 | Determination of Halide Ions |

MATERIAL: Electrodes shall be made with internal elements and external materials to meet the requirements for temperature, and resistance to chemical attack.

DESIGN: Electrodes may be either glass or plastic with side filling holes The leads may be detachable or permanently attached and shall be a minimum of 76 cm in length The electrode plugs shall be designed so that they are readily adaptable for use with most pH meters

PURPOSE: For use in determining oxidation reduction measurements.

NOTE: Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with material requirements as specified herein

USER ACTIVITIES: ARMY - CE
NAVY - NC

REVIEWER ACTIVITIES: ARMY - MD

APPROVED 28 JAN 1982 REVISED

This military standard is approved for use by all Departments and Agencies of the Department of Defense Selection for all new engineering and design applications and for repetitive use shall be made from this document

P A: EA
Other Cust: MS 99
Procurement Specifications

INTERNATIONAL INTEREST

TITLE
ELECTRODE, HYDROGEN METER (METALLIC TYPE)

SUPERSEDES: MS36465

MILITARY STANDARD
MS 36585
SHEET 1 OF 1

DD FORM 1 MAY 73 672 (COORDINATED)

PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

PROJECT NUMBER 6630-0191
STANDARD DATE

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:44Z
Check the source to verify that this is the current version before use.