FED. SUP CLASS
6640



MATERIAL: Cap and grid shall be made of brass.

DESIGN: Cap is designed to allow slower burning natural gas to pass through at lower velocity and to eliminate flame blowout. Top diameter is 40 mm, lower diameter shall fit standard bunsen burner, 13 mm, tube.

NOTE: Items differing from this Miltiary Stand... variations in design or dimensions are considered st... ument, provided such variations do not affect ...onflict with the material requir...

CANCELLED

CANCELLED AFTER 19 AUG 1991

NO SUPERSEDING STANDARD

Ⓐ CANCELLATION

APPROVED 7 JAN 1983 REVISED Ⓐ 19 AUG 1991

USER ACTIVITIES: ARMY - MD, OM
NAVY - MS

REVIEWER ACTIVITIES: ARMY - ME
NAVY - OS
AIR FORCE - 03

AMSC N/A
DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

P.A. EA	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
Other Cust MS 99		CAP AND GRID, BURNER	MS36597
Procurement Specifications NONE	SUPERSEDES:		SHEET 1 OF 1

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE CANCELLATION

Project Number 6640-1719-15

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:44Z
Check the source to verify that this is the current version before use.