FED. SUP CLASS
6640

USER ACTIVITIES: ARMY - EL, ME

PART NUMBER	OUTSIDE DIAMETER ± 0.5 mm	OVERALL LENGTH ± 2.5 mm
MS36598-1	10	75
MS36598-2	13	100
*MS36598-3	16	125

* Nominal dimensions

MATERIAL: Borosilicate glass in accordance with Federal Specification DD-G-541, Type I, Class A or B.

DESIGN: Test tubes shall be cylindrical with a rounded bottom and an open end having a lipless, fire-polished edge.

Thickness of side and bottom walls shall be uniform and shall be not less than 0.50 mm thick.

NOTES:

1. Procurement document for MS36598-1 and MS36598-2 is Military Specification MIL-T-36564.
2. Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements as specified herein.
3. For design feature purposes this standard takes precedence over procurement document referenced herein.
4. Referenced documents shall be of the issue in effect on date of invitations for bid.

APPROVED 20 June 1969 REVISED

This military standard is approved by the Department of Defense and is mandatory on all activities. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

P.A. MU	TITLE	MILITARY STANDARD
Other Cust MS	TEST TUBE disposable	MS36598
Procurement Specification See Note 1	SUPERSEDES:	SHEET 1 OF 1

DD FORM 1 SEP 63 672 (Coordinated) EDITION OF 1 SEP 57 MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:44Z
Check the source to verify that this is the current version before use.