INCH-POUND

A-A-58011
30 JAN 93

COMMERCIAL ITEM DESCRIPTION

CLEANER, ULTRASONIC, SIX GALLON

The General Services Administration has authorized the use of this Commercial Item Description.

Abstract. This Commercial product description covers a six gallon nominal capacity ultrasonic cleaner with a variable temperature heater suitable for branch mounting.

Salient characteristics.

1, Dimensions. The cleaner shall have unobstructed internal dimensions of not less than 12 inches long by 10 Inches wide by 12 inches cleaning medium depth.

2. Heater. The cleaner shall be heated by a variable, thermostatically controlled electric heater.

3. Generator. The generator chassis may be either integral with the tank or a separate unit that is cable connected to the tank. It shall operate at a nominally rated frequency of not greater than 40 kHz. The true cleaning power output of the tank shall be less than 100 Watts per gallon of cleaning medium.

4. Mounting. The cleaner shall be suitable for either placement on a shop bench or mounted in an opening cut in a bench top.

5. Drain. The cleaner shall be fitted with a bottom drain.

6. Cover. The cleaner shall be supplied with its customary lift-off cover.

7. Basket. The cleaner shall be supplied with its customary stainless steel parts basket.

8. Cleaning Medium. The ultrasonic cleaning medium shall be water with readily available appropriate cleaning agents added such as soaps, detergents, or wetting agents.

Beneficial comments, recommendations, additions, deletions, clarifications, etc., and any data which may improve this document should be sent to: The Resources and Logistics Services Division, SA-ALC/TILDM, Kelly AFB, Tx 78241-5000.

AMSC N/A FSC 4940

DISTRIBUTION STATMENT A. Approved for public release, distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:58Z
Check the source to verify that this is the current version before use.

9. Electric power. The cleaner shall be suitable for connection to a source of 120 volts 1 phase 50/60 hertz power.

10. Completeness. The cleaner shall be furnished complete with all items normally supplied to non-military customers. The cleaner shall be ready for use after bench mounting, tank filling, connection to an electric power supply, and tank drain hook–up.

Contractor Certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producer's own drawings, specifications, standards and quality assurance practices. The government reserves the rights to require proof of such conformance prior to first delivery and thereafter as may be provided for under the provisions of the contract.

Metric Products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured to inch/pound units, provided they fall within the tolerances specified in the inch/pound units. A request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements. The offeror/contractor is encouraged to use recovered materials in accordance with Public Law 94–580 to the maximum extent possible.

Preservation, packaging, packing, labeling, and marking. The preservation, packaging, packing, labeling, and marking shall be as specified in the contract or order.

Notes. Purchasers should include the following information in procurement documents.

a. Title, number, and date of the Commercial item description.

CID based part number identification number. he following part identification numbering procedure is for government purposes and does not constitute a requirement for the contractor.






Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:58Z
Check the source to verify that this is the current version before use.

Known acceptable products. The following product(s) are known to have previously met the requirements of this CID and need not resubmit bid samples or certification.

| Manufacturer | Part number | NSN |
|---|---|---|
| Sonic Star International | HT-5.6 | 4940-01-290-4510 |
| Sonicore | TS-1647HC | Not Assigned |
| Krest Ultrasonics | 4HT-1O14-6 | Not Assigned |
| Baron Blakslee | AC-144 | 4940-01-043-3814 |

MILITARY INTEREST

Custodians:
Air Force - 99

PREPARING ACTIVITY:

Air Force - 82

(Project number 4940-F600)



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:58Z
Check the source to verify that this is the current version before use.