[INCH POUND]
A-A-58032
October 17, 1994

COMMERCIAL ITEM DESCRIPTION

TEST SET, ELECTRICAL POWER

The General Services Administration has authorized
the use of this Commercial Item Description (CID) for all federal agencies

1. SCOPE.

This CID, describes a Test Set, Electrical Power This CID is meant as a minimum requirement for the Test Set, Electrical Power in which only those manufacturers that meet or surpass the following requirements are qualified per this CID The Test Set is to be used for making level, frequency, noise, and noise-to-ground measurements on telephone lines

2. CLASSIFICATION.

Equipment covered by this CID may be commercially available equipment modified to the extent necessary to meet the following description The equipment shall be Type III, Class 5, Style E and Color R in accordance with MIL-T-28800

3. SALIENT CHARACTERISTICS.

3 1 Design and Construction. The equipment shall be UL (Underwriters Laboratories) listed and approved and/or shall comply with the type III safety requirements of - MIL-T-28800

3 2 Electrical Power Sources The equipment shall operate from nominal commercial, military and shipboard power sources of 115/230 volts AC (+-10%) at line frequencies of 50 Hz to 60 Hz Transient state conditions shall meet MIL-T-28800 Class 5 requirements The unit shall also operate from an internal, rechargeable battery pack for a minimum of three (3) hours

Beneficial comments, recommendations, additions, deletions, clarifications etc., and any data which may improve this document should be sent to The Resources and Logistics Services Division, SA-ALC/TILDD, Bldg 171 Post C-12 485 Quentin Roosevelt Rd , Kelly AFB TX 78241-6524.

AMSC N/A FSC 6625

DISTRIBUTION STATEMENT A Approved for public release, distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:59Z
Check the source to verify that this is the current version before use.

3.3 Temperature Requirements.

(a) Operating Temperature The equipment shall meet its performance and accuracy requirements in an operating environment of 0°C to +50°C

(b) Non-operating Temperature The equipment shall meet its performance and accuracy requirements in a non-operating environment of -40°C to +60°C

3 4 Mean Time Between Failures (MTBF) The equipment MTBF shall be 2.000 hours The lower test MTBF shall be 2.000 hours and the upper test MTBF shall be 4.000 hours

3.5 Pre-conditioning Requirement Each deliverable equipment shall be capable of being subjected to the environmental screening effects of burn in and thermal cycling, with thermal cycling being accomplished in the sample testing phase

3 6 Calibration and Maintenance Adjustments The design of the system shall provide for readily accessible calibration adjustments and maintenance adjustments These adjustments shall be provided by variable value components which are adjustable by the use of simple means The calibration by substitution of selected components or parts is unacceptable unless specifically approved The calibration adjustments, wherever possible, shall be accessible without removal of the instrument case or modules The calibration interval shall be a period of one year or greater based on an operating time of 2,000 hours

3 7 Performance Characteristics The Test Set shall be capable of performing the following measurements in accordance with IEEE 743-1984 and shall meet or exceed the measurement parameters listed below

(a) Level Measurements

(a) Range	40 0 dBm to +13 dBm
(b) Resolution	0 1 dBm
(c) Accuracy	±0.5 dBm up to 20 kHz

(b) Frequency Measurements

(a) Range	50 Hz to 20 kHz
(b) Resolution	1 Hz from 50 to 4 kHz
	10 Hz from 4K to 20 kHz
(c) Accuracy	±1 Hz from 50 to 4 kHz
	0 Hz from 4 K to 20 kHz

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:59Z
Check the source to verify that this is the current version before use.

(c) Noise Measurements

(a) Range	10 dBm to 99 dBm
(b) Resolution	1 dBm
(c) Accuracy	±2 0 dBm from 0 to 20 dBm
	±1 0 dBm from 20 to 99 dBm

(d) Noise-to-Ground Measurements

(a) Range	50 dBm to 130 dBm
(b) Resolution	1 dBm
(c) Accuracy	±1 5 dBm

3 8 Connectors The equipment shall provide the following front panel connectors WECO type 309 or 310, additionally, binding posts are preferred but not required

4. REGULATORY REQUIREMENTS.

In accordance with section 23 403 of the Federal Acquisition Regulations, the Government's policy is to acquire items composed of the highest percentage of recovered materials practicable, consistent with maintaining a satisfactory level of competition without adversely affecting performance requirements or exposing the supplier's employees to undue hazards from the recovered materials

5. PACKAGING.

The items shall be packed in accordance with normal commercial practice and packed to assure acceptance by common carrier and provide product protection against loss and damage during multiple shipments, handling, and storage The shipping shall be in compliance with national motor freight classification and uniform freight classification Shipping container shall be marked as specified in the contract or order

6. QUALITY ASSURANCE.

Certification The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this CID, and that the product conforms to the producer's own drawings, specifications, standards and quality assurance practices.

The Government reserves the right to require proof of such conformance prior to first delivery, and thereafter, as may be otherwise provided for under the provisions of the contract

The Contracting Officer has the option of accepting or rejecting the product

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:59Z
Check the source to verify that this is the current version before use.

7. NOTES.

Metric Products Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, provided they fall within the specified tolerances using conversion tables contained in the latest revision of Federal Standard No 376 and all other requirements of this CID are met If a product is manufactured to metric dimensions exceeding the tolerances specified in the inch-pound units a request should be made to the contracting officer to determine if the product is acceptable

Federal and Military specifications are available from the Defense Documents Order Desk, Building 4D 700 Robins Avenue Philadelphia, PA 19111-5094

Known Acceptable Products. The following product(s) are known to have previously met the salient characteristics of this CID and need not resubmit bid samples or certification All other manufacturers must meet the requirements of this CID and submit to bid testing and certification as required

Manufacturer	Model	NSN
Hekimian	42-10-2Q	6625-01-067-8835
Ameritec	AM42	6625-01-276-6503
Hewlett-Packard	4935A/E01	6625-01-205-6492
CXR Telecom	701A-02	6625-01-269-7949

MILITARY INTERESTS

PREPARING ACTIVITY

Air Force - 82

Custodian
Air Force - 99

Project Nos 6625-0731
6625-0804

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:59Z
Check the source to verify that this is the current version before use.