NOTICE OF INACTIVATION
FOR NEW DESIGN

NOT MEASUREMENT
SENSITIVE

MIL-STD-45662A
NOTICE 1
20 January 1995

MILITARY STANDARD

CALIBRATION SYSTEMS REQUIREMENTS

MIL-STD-45662A, dated 1 August 1988, is inactive for new design after 26 December 1994. This notice creates a transition period, during which it will not be contractually invoked for new designs unless a contractor proposes, and the contracting office accepts, its use. MIL-STD-45662A may be placed on contract for reprocurement efforts during the transition period.

For new designs during the transition period, refer to International Organization for Standards (ISO) 10012-1, Quality Assurance Requirements for Measuring Equipment, Part 1: Meteorological Confirmation System for Measuring Equipment; American National Standards Institute (ANSI)/National Conference of Standards Laboratories (NCSL) Z540-1, General Requirements for Calibration Laboratories and Measuring and Test Equipment; or comparable standards as alternatives to MIL-STD-45662A.

(DoD activities may obtain copies of ISO 10012-1 and ANSI/NCSL from the Standardization Document Order Desk, 700 Robbins Avenue, Building 4D, Philadelphia, PA 19111-5094. The private sector and other Government agencies may purchase copies from The American National Standards Institute (ANSI), 11 West 42nd Street, New York, NY 10036.)

Preparing activity:
Army - MI
(Project QCIC-0169)

AMSC N/A AREA QCIC

DISTRIBUTION STATEMENT A. Approved for public release; disribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-24T02:43Z
Check the source to verify that this is the current version before use.