INCH-POUND

A-A-54695
29 June 1992

COMMERCIAL ITEM DESCRIPTION

PSYCHROMETER (SLING)

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-P-37748 which is cancelled.

This Commercial Item Description covers one type of sling psychrometer used for determining relative humidity of atmospheric air by means of wet and dry bulb temperature readings.

Salient characteristics:

General. The sling psychrometer shall basically be made of two thermometers attached to a mounting frame, a handle, and a carrying case. Relative humidity reading shall be obtained from a psychrometric table provided.

Each thermometer shall be cylindrical in shape with a nominal length of not less than 9 inches and shall contain mercury as the indicator. The stem shall be made of glass tubing with a strip of opalescent glass, running the entire length of the back of the stem.

Each thermometer shall have a range of 20°F to 120°F with graduations of 1°F. Each incremental 10° graduation, starting at 20°F, shall be numerically marked. Graduations and numerals shall be marked with permanent black pigment. The accuracy of thermometer readings shall be ± 1.2°F when comparing with a calibrated standard certified to a ±0.1°F accuracy in a liquid bath.

AMSC N/A FSC 6685

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

TABLE I. Gage types and characteristics.

The mounting frame shall be made of corrosion-resistant metal having a non-reflective finish and shall be provided with two parallel longitudinal grooves for mounting the thermometers. The groove design shall position the thermometers so that, in psychrometer operation, air passing over the wet-bulb does not pass over the dry bulb. The frame design shall include provisions for the attachment of the handle.

The handle shall be designed to allow the frame to rotate freely about its axis in a 360°circle at maximum angle of 90° during psychrometry operation.

The wick shall be made of muslin or cotton sewn or woven into a tube. The wick diameter shall enable the wick to fit snugly over the thermometer bulb and the length shall allow the wick to extend 3/8 inch beyond the bulb on to the stem, and 1/2 inch below the bulb. The wick shall be securely tied to the wet bulb by means of cotton or nylon thread.

The psychrometric table shall be made of plastic marked with durable printing which shall not wash off with soap and water. The table shall be able to read relative humidity by interpolating wet and dry bulb temperature readings. Based on operating at normal barometric pressure, the psychrometric table shall cover the 0 to 100 percent relative humidity range corresponding to either the wet and dry bulb thermometer readings, or the dry bulb thermometer readings and the difference between the readings.

A protective case with hinged or sliding cover shall be provided to protect the psychrometer when not in use and to accommodate the spare wicks, thread and psychrometric table. The case and cover shall be fabricated from corrosion resisting metal, plastic or wood. The hinged cover shall include a positive closure mechanism.

Spare parts. The sling psychrometer shall be provided with twelve extra wicks and at least eighty feet of thread with each psychrometer.

Identification marking. Each sling psychrometer frame or case shall be permanently and legibly marked with the name or trademark of the manufacturer, model or stock number.

Instruction. Instruction for use of the psychrometer shall be provided on the back of the psychrometric table.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Workmanship. The psychrometer shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One psychrometer, as specified, constitutes one unit.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination and Testing. Examination and testing shall be conducted on the psychrometer in accordance with commercial practice. The sampling for examination and testing shall be in accordance with commercial practice.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:37Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor Certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction , the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug, and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered material. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	Item Identification
6685-00-438-2400	PSYCHROMETER, SLING

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6685-0879

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-54695	29 JUNE 1992

3. DOCUMENT TITLE
PSYCHROMETER (SLING)

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
Defense Personnel Support Center	(215)737-2870	444-2870
c. ADDRESS (Include Zip Code) ATTN: DPSC-MSE 2800 S. 20th Street Philadelphia, PA 19101	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 *Previous editions are obsolete.* 198/2

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:37Z
Check the source to verify that this is the current version before use.